

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050323\
 Data File : BN025234.D
 Acq On : 03 May 2023 10:36
 Operator : CG/JU
 Sample : PB152462BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK462

Quant Time: May 04 04:28:34 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 15:28:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.913	152	10536	0.400	ng/u1	0.00
4) Naphthalene-d8	10.721	136	36809	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.561	164	22922	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.309	188	48172	0.400	ng/u1	0.00
17) Chrysene-d12	21.503	240	39816	0.400	ng/u1	0.00
23) Perylene-d12	23.949	264	38053	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	60423	5.539	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.316	152	16653	0.352	ng/u1	0.00
18) Fluoranthene-d10	19.340	212	38783	0.369	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050323\
Data File : BN025234.D
Acq On : 03 May 2023 10:36
Operator : CG/JU
Sample : PB152462BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK462

Quant Time: May 04 04:28:34 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 15:28:32 2023
Response via : Initial Calibration

