

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050323\
 Data File : BN025245.D
 Acq On : 03 May 2023 18:31
 Operator : CG/JU
 Sample : PB152462BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK462

Quant Time: May 04 04:30:31 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 15:28:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.917	152	8470	0.400	ng/ul	0.00
4) Naphthalene-d8	10.726	136	29347	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.562	164	17886	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.310	188	36865	0.400	ng/ul	0.00
17) Chrysene-d12	21.512	240	29139	0.400	ng/ul	0.00
23) Perylene-d12	23.941	264	27071	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	48366	5.515	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.315	152	13916	0.369	ng/ul	0.00
18) Fluoranthene-d10	19.341	212	31054	0.404	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050323\
Data File : BN025245.D
Acq On : 03 May 2023 18:31
Operator : CG/JU
Sample : PB152462BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK462

Quant Time: May 04 04:30:31 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 15:28:32 2023
Response via : Initial Calibration

