

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050421\
 Data File : BN014526.D
 Acq On : 04 May 2021 14:47
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: May 04 18:47:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 04 16:29:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.620	152	5142	0.40	ng	0.00
7) Naphthalene-d8	10.391	136	17560	0.40	ng	# 0.00
13) Acenaphthene-d10	14.257	164	8662	0.40	ng	0.00
19) Phenanthrene-d10	17.007	188	17152	0.40	ng	# 0.00
29) Chrysene-d12	21.207	240	12770	0.40	ng	# 0.00
35) Perylene-d12	23.400	264	13409	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.220	112	763	0.07	ng	0.00
5) Phenol-d6	6.798	99	611	0.05	ng	0.00
8) Nitrobenzene-d5	8.680	82	173	0.02	ng	-0.09
11) 2-Methylnaphthalene-d10	11.999	152	2389	0.09	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	134	0.05	ng	0.01
15) 2-Fluorobiphenyl	12.887	172	2507	0.08	ng	0.01
27) Fluoranthene-d10	19.049	212	4012	0.09	ng	0.00
31) Terphenyl-d14	19.660	244	2806	0.10	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.166	88	699	0.11	ng	# 82
3) n-Nitrosodimethylamine	3.569	42	48	0.01	ng	# 90
6) bis(2-Chloroethyl)ether	7.056	93	1053	0.09	ng	98
9) Naphthalene	10.441	128	4401	0.10	ng	# 94
10) Hexachlorobutadiene	10.733	225	914	0.11	ng	# 100
12) 2-Methylnaphthalene	12.071	142	2540	0.09	ng	# 95
16) Acenaphthylene	13.978	152	2985	0.09	ng	100
17) Acenaphthene	14.321	154	2267	0.10	ng	97
18) Fluorene	15.310	166	2578	0.09	ng	100
20) 4,6-Dinitro-2-methylph...	15.475	198	5	0.00	ng	# 1
21) 4-Bromophenyl-phenylether	16.216	248	688	0.08	ng	# 91
22) Hexachlorobenzene	16.313	284	1205	0.11	ng	96
23) Atrazine	16.483	200	427	0.08	ng	# 80
24) Pentachlorophenol	16.702	266	80	0.03	ng	# 65
25) Phenanthrene	17.043	178	4134	0.09	ng	99
26) Anthracene	17.140	178	2615	0.07	ng	97
28) Fluoranthene	19.079	202	4214	0.08	ng	99
30) Pyrene	19.441	202	4213	0.10	ng	99
32) Benzo(a)anthracene	21.197	228	2382	0.08	ng	95
33) Chrysene	21.250	228	4055	0.10	ng	98
34) Bis(2-ethylhexyl)phtha...	21.133	149	1603	0.12	ng	100
36) Indeno(1,2,3-cd)pyrene	25.598	276	3538	0.08	ng	96
37) Benzo(b)fluoranthene	22.750	252	3678	0.09	ng	# 60
38) Benzo(k)fluoranthene	22.791	252	4971	0.09	ng	# 74
39) Benzo(a)pyrene	23.311	252	4132	0.10	ng	# 15
40) Dibenzo(a,h)anthracene	25.625	278	2717	0.07	ng	# 73
41) Benzo(g,h,i)perylene	26.262	276	4183	0.09	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050421\
Data File : BN014526.D
Acq On : 04 May 2021 14:47
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: May 04 18:47:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 04 16:29:35 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.620 min Scan# 601
 Delta R.T. -0.000 min
 Lab File: BN014526.D
 Acq: 04 May 2021 14:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 5142
 Ion Ratio Lower Upper
 152 100
 150 152.7 123.8 185.8
 115 58.9 47.0 70.4



#2
 1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.166 min Scan# 48
 Delta R.T. 0.000 min
 Lab File: BN014526.D
 Acq: 04 May 2021 14:47

Tgt Ion: 88 Resp: 699
 Ion Ratio Lower Upper
 88 100
 58 64.5 44.4 66.6
 43 36.1 15.9 23.9#









#3
n-Nitrosodimethylamine
Concen: 0.01 ng
RT: 3.569 min Scan# 98
Delta R.T. 0.024 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 48
Ion Ratio Lower Upper
42 100
74 233.3 175.1 262.7
44 0.0 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.07 ng
RT: 5.220 min Scan# 303
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion: 112 Resp: 763
Ion Ratio Lower Upper
112 100
64 56.2 41.8 62.6
63 30.8 22.0 33.0









#5
Phenol-d6
Concen: 0.05 ng
RT: 6.798 min Scan# 499
Delta R.T. 0.008 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion: 99 Resp: 611
Ion Ratio Lower Upper
99 100
42 23.7 13.2 19.8#
71 33.7 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 0.09 ng
RT: 7.056 min Scan# 531
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion: 93 Resp: 1053
Ion Ratio Lower Upper
93 100
63 73.4 58.0 87.0
95 32.0 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.391 min Scan# 851
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	17560
Ion Ratio	Lower	Upper	
136	100		
137	11.6	8.9	13.3
54	8.5	3.8	5.8#
68	6.7	3.8	5.6#



#8
Nitrobenzene-d5
Concen: 0.02 ng
RT: 8.680 min Scan# 716
Delta R.T. -0.089 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:	82	Resp:	173
Ion Ratio	Lower	Upper	
82	100		
128	41.9	29.5	44.3
54	41.9	36.9	55.3









#9
Naphthalene
Concen: 0.10 ng
RT: 10.441 min Scan# 855
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 4401
Ion Ratio Lower Upper
128 100
129 14.0 9.0 13.4#
127 15.2 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.11 ng
RT: 10.733 min Scan# 878
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:225 Resp: 914
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.09 ng
RT: 11.999 min Scan# 1050
Delta R.T. 0.009 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2389
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.09 ng
RT: 12.071 min Scan# 1066
Delta R.T. 0.005 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:142 Resp: 2540
Ion Ratio Lower Upper
142 100
141 88.5 70.2 105.4
115 42.1 26.7 40.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.257 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 8662
Ion Ratio Lower Upper
164 100
162 100.7 78.6 118.0
160 50.0 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.05 ng
RT: 15.767 min Scan# 1421
Delta R.T. 0.013 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:330 Resp: 134
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.1 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.08 ng
RT: 12.887 min Scan# 1194
Delta R.T. 0.013 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	172	Resp:	2507
Ion	Ratio	Lower	Upper
172	100		
171	36.1	27.1	40.7
170	26.1	18.3	27.5



#16
Acenaphthylene
Concen: 0.09 ng
RT: 13.978 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:	152	Resp:	2985
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.6	23.4
153	12.7	10.5	15.7









#17
Acenaphthene
Concen: 0.10 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:154 Resp: 2267
Ion Ratio Lower Upper
154 100
153 115.6 90.2 135.4
152 60.3 46.2 69.2



#18
Fluorene
Concen: 0.09 ng
RT: 15.310 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:166 Resp: 2578
Ion Ratio Lower Upper
166 100
165 98.0 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.007 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:188 Resp: 17152
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.4 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.051 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:198 Resp: 5
Ion Ratio Lower Upper
198 100
51 416.3 865.3 1297.9#
105 155.1 93.6 140.4#









#21
4-Bromophenyl-phenylether
Concen: 0.08 ng
RT: 16.216 min Scan# 1457
Delta R.T. 0.012 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:248 Resp: 688
Ion Ratio Lower Upper
248 100
250 102.0 80.5 120.7
141 63.4 37.2 55.8#



#22
Hexachlorobenzene
Concen: 0.11 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:284 Resp: 1205
Ion Ratio Lower Upper
284 100
142 37.6 26.6 39.8
249 29.4 23.6 35.4









#23
Atrazine
Concen: 0.08 ng
RT: 16.483 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:200 Resp: 427
Ion Ratio Lower Upper
200 100
173 43.0 23.0 34.6#
215 59.5 38.6 57.8#



#24
Pentachlorophenol
Concen: 0.03 ng
RT: 16.702 min Scan# 1497
Delta R.T. 0.024 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:266 Resp: 80
Ion Ratio Lower Upper
266 100
264 43.8 24.2 36.4#
268 56.3 25.2 37.8#









#25
Phenanthrene
Concen: 0.09 ng
RT: 17.043 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178	Resp:	4134	
Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.2	22.8
179	14.8	12.4	18.6



#26
Anthracene
Concen: 0.07 ng
RT: 17.140 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:178	Resp:	2615	
Ion	Ratio	Lower	Upper
178	100		
176	20.1	14.8	22.2
179	16.6	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.09 ng
RT: 19.049 min Scan# 1813
Delta R.T. 0.005 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 4012
Ion Ratio Lower Upper
212 100
106 13.1 10.1 15.1
104 7.2 5.4 8.2



#28
Fluoranthene
Concen: 0.08 ng
RT: 19.079 min Scan# 1819
Delta R.T. 0.005 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:202 Resp: 4214
Ion Ratio Lower Upper
202 100
101 10.9 8.4 12.6
203 17.1 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.207 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 12770
Ion Ratio Lower Upper
240 100
120 12.9 5.8 8.8#
236 27.1 20.6 30.8



#30
Pyrene
Concen: 0.10 ng
RT: 19.441 min Scan# 1892
Delta R.T. 0.005 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47


Tgt Ion:202 Resp: 4213
Ion Ratio Lower Upper
202 100
200 20.0 16.3 24.5
203 18.3 14.0 21.0








#31
Terphenyl-d14
Concen: 0.10 ng
RT: 19.660 min Scan# 1936
Delta R.T. 0.010 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:244 Resp: 2806
Ion Ratio Lower Upper
244 100
212 11.6 6.0 9.0#
122 15.6 9.3 13.9#



#32
Benzo(a)anthracene
Concen: 0.08 ng
RT: 21.197 min Scan# 2118
Delta R.T. 0.011 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:228 Resp: 2382
Ion Ratio Lower Upper
228 100
226 28.8 21.7 32.5
229 22.9 15.7 23.5









#33
Chrysene
Concen: 0.10 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.011 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:228 Resp: 4055
Ion Ratio Lower Upper
228 100
226 30.1 23.8 35.6
229 21.2 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.12 ng
RT: 21.133 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Tgt Ion:149 Resp: 1603
Ion Ratio Lower Upper
149 100
167 25.7 20.5 30.7
279 3.9 2.9 4.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.400 min Scan# 2522

Delta R.T. 0.003 min

Lab File: BN014526.D

Acq: 04 May 2021 14:47





Tgt Ion:264 Resp: 13409

Ion Ratio Lower Upper

264 100

260 25.4 19.5 29.3

265 46.4 35.4 53.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.08 ng

RT: 25.598 min Scan# 3164

Delta R.T. 0.007 min

Lab File: BN014526.D

Acq: 04 May 2021 14:47




Tgt Ion:276 Resp: 3538


Ion Ratio Lower Upper

276 100

138 19.3 18.8 28.2

277 24.9 19.8 29.8





#37

Benzo(b)fluoranthene

Concen: 0.09 ng

RT: 22.750 min Scan# 2332

Delta R.T. 0.007 min

Lab File: BN014526.D

Acq: 04 May 2021 14:47





Tgt Ion:252 Resp: 3678

Ion Ratio Lower Upper

252 100

253 36.5 19.6 29.4#

125 40.3 10.3 15.5#



#38

Benzo(k)fluoranthene

Concen: 0.09 ng

RT: 22.791 min Scan# 2344

Delta R.T. 0.007 min

Lab File: BN014526.D

Acq: 04 May 2021 14:47




Tgt Ion:252 Resp: 4971


Ion Ratio Lower Upper

252 100

253 36.5 19.7 29.5#

125 23.6 9.7 14.5#









#39
Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.311 min Scan# 2496
Delta R.T. 0.007 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion:252 Resp: 4132
Ion Ratio Lower Upper
252 100
253 50.0 21.8 32.6#
125 83.9 14.9 22.3#



#40
Dibenzo(a,h)anthracene
Concen: 0.07 ng
RT: 25.625 min Scan# 3172
Delta R.T. 0.021 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47


Tgt Ion:278 Resp: 2717
Ion Ratio Lower Upper
278 100
139 26.1 13.9 20.9#
279 42.4 20.7 31.1#





#41
Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 26.262 min Scan# 3358
Delta R.T. 0.004 min
Lab File: BN014526.D
Acq: 04 May 2021 14:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1



Tgt Ion:276 Resp: 4183
Ion Ratio Lower Upper
276 100
277 28.5 19.6 29.4
138 28.1 17.5 26.3#

