

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050421\
 Data File : BN014533.D
 Acq On : 04 May 2021 18:58
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050421

Quant Time: May 05 05:58:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 04 19:13:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.620	152	5638	0.40	ng	0.00
7) Naphthalene-d8	10.391	136	20635	0.40	ng	# 0.00
13) Acenaphthene-d10	14.257	164	10645	0.40	ng	0.00
19) Phenanthrene-d10	17.007	188	20576	0.40	ng	# 0.00
29) Chrysene-d12	21.208	240	16837	0.40	ng	0.00
35) Perylene-d12	23.393	264	17178	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.220	112	4326	0.44	ng	0.00
5) Phenol-d6	6.790	99	4570	0.40	ng	0.00
8) Nitrobenzene-d5	8.769	82	5581	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	11.991	152	12838	0.43	ng	0.00
14) 2,4,6-Tribromophenol	15.755	330	1032	0.42	ng	0.00
15) 2-Fluorobiphenyl	12.874	172	15985	0.42	ng	0.00
27) Fluoranthene-d10	19.039	212	21331	0.40	ng	0.00
31) Terphenyl-d14	19.650	244	16141	0.43	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.158	88	2344	0.37	ng	# 70
3) n-Nitrosodimethylamine	3.545	42	1109	0.43	ng	# 38
6) bis(2-Chloroethyl)ether	7.056	93	5170	0.41	ng	94
9) Naphthalene	10.442	128	21961	0.42	ng	100
10) Hexachlorobutadiene	10.733	225	4456	0.43	ng	# 100
12) 2-Methylnaphthalene	12.062	142	13875	0.43	ng	98
16) Acenaphthylene	13.978	152	17156	0.40	ng	100
17) Acenaphthene	14.321	154	12070	0.41	ng	96
18) Fluorene	15.310	166	14411	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.412	198	333	0.40	ng	# 1
21) 4-Bromophenyl-phenylether	16.204	248	4385	0.42	ng	# 86
22) Hexachlorobenzene	16.313	284	6500	0.44	ng	97
23) Atrazine	16.483	200	2329	0.40	ng	96
24) Pentachlorophenol	16.666	266	697	0.46	ng	# 43
25) Phenanthrene	17.043	178	23484	0.42	ng	100
26) Anthracene	17.141	178	17003	0.40	ng	100
28) Fluoranthene	19.074	202	23999	0.40	ng	99
30) Pyrene	19.437	202	23462	0.44	ng	100
32) Benzo(a)anthracene	21.186	228	17115	0.43	ng	99
33) Chrysene	21.239	228	21999	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.123	149	7107	0.43	ng	98
36) Indeno(1,2,3-cd)pyrene	25.587	276	23336	0.39	ng	98
37) Benzo(b)fluoranthene	22.784	252	27639	0.50	ng	99
38) Benzo(k)fluoranthene	22.784	252	29680	0.43	ng	99
39) Benzo(a)pyrene	23.301	252	23276	0.42	ng	# 85
40) Dibenzo(a,h)anthracene	25.605	278	18277	0.38	ng	99
41) Benzo(g,h,i)perylene	26.252	276	23888	0.41	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050421\
Data File : BN014533.D
Acq On : 04 May 2021 18:58
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN050421

Quant Time: May 05 05:58:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 04 19:13:12 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.620 min Scan# 601
 Delta R.T. -0.000 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58




Instrument :
 BNA_N
 ClientSampled :
 ICVBN050421


Tgt Ion:152 Resp: 5638
 Ion Ratio Lower Upper
 152 100
 150 155.7 123.8 185.8
 115 60.4 47.0 70.4



#2
 1,4-Dioxane
 Concen: 0.37 ng
 RT: 3.158 min Scan# 47
 Delta R.T. -0.008 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58

Tgt Ion: 88 Resp: 2344
 Ion Ratio Lower Upper
 88 100
 58 77.5 44.4 66.6#
 43 33.2 15.9 23.9#









#3
n-Nitrosodimethylamine
Concen: 0.43 ng
RT: 3.545 min Scan# 95
Delta R.T. -0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion: 42 Resp: 1109
Ion Ratio Lower Upper
42 100
74 116.9 175.1 262.7#
44 0.0 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.44 ng
RT: 5.220 min Scan# 303
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion: 112 Resp: 4326
Ion Ratio Lower Upper
112 100
64 56.2 41.8 62.6
63 29.6 22.0 33.0









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.790 min Scan# 498
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion: 99 Resp: 4570
Ion Ratio Lower Upper
99 100
42 22.3 13.2 19.8#
71 33.1 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 7.056 min Scan# 531
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion: 93 Resp: 5170
Ion Ratio Lower Upper
93 100
63 79.1 58.0 87.0
95 34.8 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.391 min Scan# 851
Delta R.T. -0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:	136	Resp:	20635
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.9	13.3
54	7.8	3.8	5.8#
68	6.2	3.8	5.6#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.769 min Scan# 723
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:	82	Resp:	5581
Ion Ratio	Lower	Upper	
82	100		
128	43.9	29.5	44.3
54	50.8	36.9	55.3









#9
Naphthalene
Concen: 0.42 ng
RT: 10.442 min Scan# 855
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:	128	Resp:	21961
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.0	13.4
127	13.0	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.43 ng
RT: 10.733 min Scan# 878
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:	225	Resp:	4456
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.43 ng
RT: 11.991 min Scan# 1048
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:152 Resp: 12838
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.43 ng
RT: 12.062 min Scan# 1064
Delta R.T. -0.004 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:142 Resp: 13875
Ion Ratio Lower Upper
142 100
141 88.2 70.2 105.4
115 35.8 26.7 40.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.257 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:164 Resp: 10645
Ion Ratio Lower Upper
164 100
162 99.4 78.6 118.0
160 49.5 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.42 ng
RT: 15.755 min Scan# 1420
Delta R.T. 0.001 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:330 Resp: 1032
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 39.3 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.874 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:	172	Resp:	15985
Ion Ratio	Lower	Upper	
172	100		
171	34.9	27.1	40.7
170	24.1	18.3	27.5



#16
Acenaphthylene
Concen: 0.40 ng
RT: 13.978 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:	152	Resp:	17156
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.6	23.4
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.41 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:154 Resp: 12070
Ion Ratio Lower Upper
154 100
153 115.2 90.2 135.4
152 62.5 46.2 69.2



#18
Fluorene
Concen: 0.41 ng
RT: 15.310 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:166 Resp: 14411
Ion Ratio Lower Upper
166 100
165 98.1 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.007 min Scan# 1522
Delta R.T. 0.001 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:188 Resp: 20576
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.5 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.412 min Scan# 1393
Delta R.T. -0.012 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:198 Resp: 333
Ion Ratio Lower Upper
198 100
51 195.3 865.3 1297.9#
105 90.7 93.6 140.4#









#21
4-Bromophenyl-phenylether
Concen: 0.42 ng
RT: 16.204 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:	248	Resp:	4385
Ion	Ratio	Lower	Upper
248	100		
250	95.3	80.5	120.7
141	69.4	37.2	55.8#



#22
Hexachlorobenzene
Concen: 0.44 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:	284	Resp:	6500
Ion	Ratio	Lower	Upper
284	100		
142	36.3	26.6	39.8
249	29.3	23.6	35.4









#23
Atrazine
Concen: 0.40 ng
RT: 16.483 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:	200	Resp:	2329
Ion Ratio	Lower	Upper	
200	100		
173	26.3	23.0	34.6
215	50.9	38.6	57.8



#24
Pentachlorophenol
Concen: 0.46 ng
RT: 16.666 min Scan# 1494
Delta R.T. -0.012 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:	266	Resp:	697
Ion Ratio	Lower	Upper	
266	100		
264	61.0	24.2	36.4#
268	63.3	25.2	37.8#









#25
Phenanthrene
Concen: 0.42 ng
RT: 17.043 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:178 Resp: 23484
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.3 12.4 18.6



#26
Anthracene
Concen: 0.40 ng
RT: 17.141 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:178 Resp: 17003
Ion Ratio Lower Upper
178 100
176 18.4 14.8 22.2
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.039 min Scan# 1811
Delta R.T. -0.005 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:212 Resp: 21331
Ion Ratio Lower Upper
212 100
106 13.0 10.1 15.1
104 7.5 5.4 8.2



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.074 min Scan# 1818
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:202 Resp: 23999
Ion Ratio Lower Upper
202 100
101 11.3 8.4 12.6
203 17.0 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.208 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:240 Resp: 16837
Ion Ratio Lower Upper
240 100
120 8.5 5.8 8.8
236 25.9 20.6 30.8



#30
Pyrene
Concen: 0.44 ng
RT: 19.437 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:202 Resp: 23462
Ion Ratio Lower Upper
202 100
200 20.4 16.3 24.5
203 17.3 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.43 ng
 RT: 19.650 min Scan# 1934
 Delta R.T. 0.000 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050421


Tgt Ion:244	Resp:	16141
Ion Ratio	Lower	Upper
244	100	
212	8.0	6.0 9.0
122	13.2	9.3 13.9



#32
 Benzo(a)anthracene
 Concen: 0.43 ng
 RT: 21.186 min Scan# 2117
 Delta R.T. 0.000 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58

Tgt Ion:228	Resp:	17115
Ion Ratio	Lower	Upper
228	100	
226	26.6	21.7 32.5
229	20.1	15.7 23.5









#33
Chrysene
Concen: 0.42 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58




Instrument :
BNA_N
ClientSampleId :
ICVBN050421


Tgt Ion:228	Resp:	21999
Ion Ratio	Lower	Upper
228	100	
226	29.1	23.8 35.6
229	20.0	15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.43 ng
RT: 21.123 min Scan# 2111
Delta R.T. -0.010 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Tgt Ion:149	Resp:	7107
Ion Ratio	Lower	Upper
149	100	
167	26.5	20.5 30.7
279	4.1	2.9 4.3









#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.393 min Scan# 2520
 Delta R.T. -0.003 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050421


Tgt Ion:264	Resp:	17178
Ion Ratio	Lower	Upper
264	100	
260	24.2	19.5 29.3
265	39.2	35.4 53.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.39 ng
 RT: 25.587 min Scan# 3161
 Delta R.T. -0.003 min
 Lab File: BN014533.D
 Acq: 04 May 2021 18:58

Tgt Ion:276	Resp:	23336
Ion Ratio	Lower	Upper
276	100	
138	21.2	18.8 28.2
277	24.8	19.8 29.8





#37

Benzo(b)fluoranthene

Concen: 0.50 ng

RT: 22.784 min Scan# 2342

Delta R.T. 0.041 min

Lab File: BN014533.D

Acq: 04 May 2021 18:58





Tgt Ion:252 Resp: 27639

Ion Ratio Lower Upper

252 100

253 24.0 19.6 29.4

125 12.3 10.3 15.5



#38

Benzo(k)fluoranthene

Concen: 0.43 ng

RT: 22.784 min Scan# 2342

Delta R.T. 0.000 min

Lab File: BN014533.D

Acq: 04 May 2021 18:58




Tgt Ion:252 Resp: 29680


Ion Ratio Lower Upper

252 100

253 24.0 19.7 29.5

125 12.3 9.7 14.5





#39

Benzo(a)pyrene

Concen: 0.42 ng

RT: 23.301 min Scan# 2493

Delta R.T. -0.003 min

Lab File: BN014533.D

Acq: 04 May 2021 18:58





Tgt Ion:252	Resp:	23276
Ion Ratio	Lower	Upper
252	100	
253	27.9	21.8 32.6
125	33.8	14.9 22.3#

Instrument :

BNA_N

ClientSampleId :

ICVBN050421



#40

Dibenzo(a,h)anthracene

Concen: 0.38 ng




RT: 25.605 min Scan# 3166


Delta R.T. 0.000 min

Lab File: BN014533.D

Acq: 04 May 2021 18:58

Tgt Ion:278	Resp:	18277
Ion Ratio	Lower	Upper
278	100	
139	18.0	13.9 20.9
279	25.6	20.7 31.1





#41
Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 26.252 min Scan# 3355
Delta R.T. -0.007 min
Lab File: BN014533.D
Acq: 04 May 2021 18:58

Instrument :
BNA_N
ClientSampleId :
ICVBN050421

Tgt Ion:	276	Resp:	23888
Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.6	29.4
138	22.2	17.5	26.3

