




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050622\
Data File : BN019695.D
Acq On : 06 May 2022 10:57
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 07 01:10:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019695.D\data.ms

(70) Pentachlorophenol (C)




16.869min (-0.003) 18993.94 ng

response 359979

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.90
264.00	61.60	59.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050622\
Data File : BN019695.D
Acq On : 06 May 2022 10:57
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 07 01:10:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019695.D\data.ms

(77) Benzidine

19.457min (-0.009) 0.00 ng

response 2369282

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.19
183.00	13.20	11.75
0.00	0.00	0.00