

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
 Data File : BN025288.D
 Acq On : 05 May 2023 20:49
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: May 08 02:05:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 08 02:01:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	5972	0.400	ng	0.00
7) Naphthalene-d8	10.718	136	18361	0.400	ng	#-0.01
13) Acenaphthene-d10	14.555	164	12144	0.400	ng	0.00
19) Phenanthrene-d10	17.298	188	27632	0.400	ng	#-0.01
29) Chrysene-d12	21.500	240	28604	0.400	ng	0.00
35) Perylene-d12	23.942	264	31069	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.470	112	11007	0.798	ng	0.00
5) Phenol-d6	7.080	99	13864	0.797	ng	0.00
8) Nitrobenzene-d5	9.084	82	11197	0.757	ng	-0.01
11) 2-Methylnaphthalene-d10	12.313	152	19054	0.784	ng	0.00
14) 2,4,6-Tribromophenol	16.057	330	4560	0.835	ng	-0.01
15) 2-Fluorobiphenyl	13.176	172	31741	0.684	ng	-0.01
27) Fluoranthene-d10	19.338	212	49755	0.794	ng	0.00
31) Terphenyl-d14	19.932	244	42410	0.654	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.325	88	4832	0.736	ng	95
3) n-Nitrosodimethylamine	3.657	42	7513	0.675	ng	99
6) bis(2-Chloroethyl)ether	7.333	93	12196	0.691	ng	99
9) Naphthalene	10.771	128	34952	0.739	ng	99
10) Hexachlorobutadiene	11.049	225	6639	0.737	ng	# 99
12) 2-Methylnaphthalene	12.385	142	25036	0.761	ng	99
16) Acenaphthylene	14.277	152	40336	0.790	ng	99
17) Acenaphthene	14.619	154	25204	0.730	ng	100
18) Fluorene	15.603	166	34508	0.728	ng	99
20) 4,6-Dinitro-2-methylph...	15.734	198	2323	0.632	ng	# 43
21) 4-Bromophenyl-phenylether	16.492	248	11360	0.698	ng	# 90
22) Hexachlorobenzene	16.616	284	12404	0.673	ng	99
23) Atrazine	16.765	200	9611	0.866	ng	97
24) Pentachlorophenol	17.050	266	2299	0.455	ng	# 23
25) Phenanthrene	17.348	178	56437	0.703	ng	99
26) Anthracene	17.435	178	52798	0.755	ng	100
28) Fluoranthene	19.367	202	68807	0.755	ng	100
30) Pyrene	19.734	202	71532	0.704	ng	100
32) Benzo(a)anthracene	21.482	228	70843	0.732	ng	99
33) Chrysene	21.535	228	69315	0.684	ng	99
34) Bis(2-ethylhexyl)phtha...	21.401	149	52373	1.008	ng	99
36) Indeno(1,2,3-cd)pyrene	26.486	276	91464	0.671	ng	100
37) Benzo(b)fluoranthene	23.196	252	73668	0.625	ng	94
38) Benzo(k)fluoranthene	23.243	252	76690	0.657	ng	95
39) Benzo(a)pyrene	23.840	252	75731	0.672	ng	# 93
40) Dibenzo(a,h)anthracene	26.506	278	73348	0.684	ng	97
41) Benzo(g,h,i)perylene	27.266	276	79199	0.682	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
Data File : BN025288.D
Acq On : 05 May 2023 20:49
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: May 08 02:05:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 08 02:01:54 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.911 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN025288.D
 Acq: 05 May 2023 20:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:152 Resp: 5972
 Ion Ratio Lower Upper
 152 100
 150 156.6 123.8 185.8
 115 63.7 52.9 79.3



#2
 1,4-Dioxane
 Concen: 0.736 ng
 RT: 3.325 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: BN025288.D
 Acq: 05 May 2023 20:49

Tgt Ion: 88 Resp: 4832
 Ion Ratio Lower Upper
 88 100
 43 40.4 37.2 55.8
 58 87.1 71.5 107.3









#3
n-Nitrosodimethylamine
Concen: 0.675 ng
RT: 3.657 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion: 42 Resp: 7513
Ion Ratio Lower Upper
42 100
74 98.5 79.9 119.9
44 8.6 7.0 10.6



#4
2-Fluorophenol
Concen: 0.798 ng
RT: 5.470 min Scan# 317
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion: 112 Resp: 11007
Ion Ratio Lower Upper
112 100
64 62.8 51.5 77.3
63 33.7 27.8 41.6









#5
Phenol-d6
Concen: 0.797 ng
RT: 7.080 min Scan# 541
Delta R.T. -0.007 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion: 99 Resp: 13864
Ion Ratio Lower Upper
99 100
42 27.5 22.0 33.0
71 35.6 28.7 43.1



#6
bis(2-Chloroethyl)ether
Concen: 0.691 ng
RT: 7.333 min Scan# 575
Delta R.T. -0.007 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion: 93 Resp: 12196
Ion Ratio Lower Upper
93 100
63 83.5 67.8 101.6
95 32.4 26.7 40.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.718 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	18361
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.4	14.2
54	13.3	8.5	12.7#
68	8.8	5.8	8.8#



#8
Nitrobenzene-d5
Concen: 0.757 ng
RT: 9.084 min Scan# 783
Delta R.T. -0.011 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:	82	Resp:	11197
Ion Ratio	Lower	Upper	
82	100		
128	41.6	35.3	52.9
54	63.1	48.7	73.1









#9
Naphthalene
Concen: 0.739 ng
RT: 10.771 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 34952
Ion Ratio Lower Upper
128 100
129 11.3 9.4 14.2
127 13.1 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.737 ng
RT: 11.049 min Scan# 967
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:225 Resp: 6639
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.7 49.8 74.8





#11

2-Methylnaphthalene-d10

Concen: 0.784 ng

RT: 12.313 min Scan# 11

Delta R.T. -0.004 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49

Instrument :

BNA_N

ClientSampleId :





SSTDICC0.8

Tgt Ion:152 Resp: 19054

Ion Ratio Lower Upper

152 100

151 20.5 16.4 24.6



#12

2-Methylnaphthalene

Concen: 0.761 ng

RT: 12.385 min Scan# 1242

Delta R.T. -0.004 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49




Tgt Ion:142 Resp: 25036


Ion Ratio Lower Upper

142 100

141 88.4 70.5 105.7

115 37.9 31.6 47.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.555 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 12144
Ion Ratio Lower Upper
164 100
162 103.2 82.1 123.1
160 47.9 38.3 57.5



#14
2,4,6-Tribromophenol
Concen: 0.835 ng
RT: 16.057 min Scan# 1687
Delta R.T. -0.013 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:330 Resp: 4560
Ion Ratio Lower Upper
330 100
332 97.5 77.4 116.0
141 38.0 28.9 43.3









#15
2-Fluorobiphenyl
Concen: 0.684 ng
RT: 13.176 min Scan# 1422
Delta R.T. -0.011 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 31741
Ion Ratio Lower Upper
172 100
171 35.6 27.9 41.9
170 24.5 19.0 28.4



#16
Acenaphthylene
Concen: 0.790 ng
RT: 14.277 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:152 Resp: 40336
Ion Ratio Lower Upper
152 100
151 19.5 16.1 24.1
153 13.2 10.3 15.5





#17

Acenaphthene

Concen: 0.730 ng

RT: 14.619 min Scan# 1557

Delta R.T. -0.000 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 25204

Ion Ratio Lower Upper

154 100

153 114.2 91.3 136.9

152 56.0 44.7 67.1



#18

Fluorene

Concen: 0.728 ng

RT: 15.603 min Scan# 1649

Delta R.T. -0.000 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49




Tgt Ion:166 Resp: 34508


Ion Ratio Lower Upper

166 100

165 99.4 79.0 118.4

167 13.3 10.3 15.5









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1788
Delta R.T. -0.013 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 27632
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 15.1 8.0 12.0#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.632 ng
RT: 15.734 min Scan# 1661
Delta R.T. -0.013 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:198 Resp: 2323
Ion Ratio Lower Upper
198 100
51 100.9 162.1 243.1#
105 63.5 83.1 124.7#









#21
4-Bromophenyl-phenylether
Concen: 0.698 ng
RT: 16.492 min Scan# 1722
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 11360
Ion Ratio Lower Upper
248 100
250 96.9 76.0 114.0
141 56.9 59.9 89.9#



#22
Hexachlorobenzene
Concen: 0.673 ng
RT: 16.616 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:284 Resp: 12404
Ion Ratio Lower Upper
284 100
142 37.2 30.0 45.0
249 30.9 25.0 37.6





#23

Atrazine

Concen: 0.866 ng

RT: 16.765 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:200 Resp: 9611

Ion Ratio Lower Upper

200 100

173 28.6 26.3 39.5

215 46.3 37.2 55.8



#24

Pentachlorophenol

Concen: 0.455 ng

RT: 17.050 min Scan# 1767

Delta R.T. -0.025 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49




Tgt Ion:266 Resp: 2299


Ion Ratio Lower Upper

266 100

264 0.0 43.0 64.4#

268 0.0 47.4 71.2#









#25
Phenanthrene
Concen: 0.703 ng
RT: 17.348 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 56437
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.2 12.4 18.6



#26
Anthracene
Concen: 0.755 ng
RT: 17.435 min Scan# 1798
Delta R.T. -0.012 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:178 Resp: 52798
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.3 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.794 ng
RT: 19.338 min Scan# 2160
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 49755
Ion Ratio Lower Upper
212 100
106 15.4 12.6 19.0
104 8.7 7.0 10.6



#28
Fluoranthene
Concen: 0.755 ng
RT: 19.367 min Scan# 2167
Delta R.T. -0.004 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:202 Resp: 68807
Ion Ratio Lower Upper
202 100
101 12.1 9.8 14.8
203 17.0 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.500 min Scan# 2486
Delta R.T. -0.009 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:240 Resp: 28604
Ion Ratio Lower Upper
240 100
120 15.2 10.3 15.5
236 28.1 22.3 33.5



#30
Pyrene
Concen: 0.704 ng
RT: 19.734 min Scan# 2254
Delta R.T. -0.000 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:202 Resp: 71532
Ion Ratio Lower Upper
202 100
200 20.9 16.8 25.2
203 17.9 14.4 21.6









#31
Terphenyl-d14
Concen: 0.654 ng
RT: 19.932 min Scan# 21
Delta R.T. -0.004 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 42410
Ion Ratio Lower Upper
244 100
212 8.5 7.1 10.7
122 0.0 0.0 0.0



#32
Benzo(a)anthracene
Concen: 0.732 ng
RT: 21.482 min Scan# 2483
Delta R.T. -0.009 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:228 Resp: 70843
Ion Ratio Lower Upper
228 100
226 28.3 22.5 33.7
229 19.5 16.2 24.2





#33

Chrysene

Concen: 0.684 ng

RT: 21.535 min Scan# 2490

Delta R.T. -0.009 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 69315

Ion Ratio Lower Upper

228 100

226 31.0 24.6 36.8

229 19.5 16.2 24.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.008 ng

RT: 21.401 min Scan# 2474

Delta R.T. -0.000 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49




Tgt Ion:149 Resp: 52373


Ion Ratio Lower Upper

149 100

167 27.2 22.2 33.4

279 3.2 2.6 4.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.942 min Scan# 3109

Delta R.T. 0.009 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:264 Resp: 31069

Ion Ratio Lower Upper

264 100

260 26.2 21.5 32.3

265 46.9 46.2 69.4



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.671 ng

RT: 26.486 min Scan# 3982

Delta R.T. 0.029 min

Lab File: BN025288.D

Acq: 05 May 2023 20:49




Tgt Ion:276 Resp: 91464


Ion Ratio Lower Upper

276 100

138 27.9 22.5 33.7

277 24.8 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.625 ng
RT: 23.196 min Scan# 2856
Delta R.T. 0.003 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 73668
Ion Ratio Lower Upper
252 100
253 23.8 20.8 31.2
125 13.5 13.4 20.2



#38
Benzo(k)fluoranthene
Concen: 0.657 ng
RT: 23.243 min Scan# 2873
Delta R.T. 0.003 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:252 Resp: 76690
Ion Ratio Lower Upper
252 100
253 23.9 20.6 30.8
125 13.6 13.1 19.7









#39
Benzo(a)pyrene
Concen: 0.672 ng
RT: 23.840 min Scan# 3073
Delta R.T. 0.012 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 75731
Ion Ratio Lower Upper
252 100
253 23.9 21.4 32.0
125 15.3 15.4 23.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.684 ng
RT: 26.506 min Scan# 3989
Delta R.T. 0.026 min
Lab File: BN025288.D
Acq: 05 May 2023 20:49

Tgt Ion:278 Resp: 73348
Ion Ratio Lower Upper
278 100
139 20.1 17.6 26.4
279 25.4 21.6 32.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.682 ng
 RT: 27.266 min Scan# 41
 Delta R.T. 0.035 min
 Lab File: BN025288.D
 Acq: 05 May 2023 20:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	276	Resp:	79199
Ion Ratio	Lower	Upper	
276	100		
277	24.1	20.1	30.1
138	25.5	21.0	31.6

