

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
 Data File : BN025311.D
 Acq On : 06 May 2023 12:22
 Operator : CG/JU
 Sample : 02558-08MS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ZONE-A-3-6-05022023MS

Quant Time: May 08 04:20:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 08 04:14:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	5797	0.400	ng	0.00
7) Naphthalene-d8	10.718	136	19472	0.400	ng	0.00
13) Acenaphthene-d10	14.544	164	12327	0.400	ng	-0.01
19) Phenanthrene-d10	17.298	188	29131	0.400	ng	0.00
29) Chrysene-d12	21.491	240	29606	0.400	ng	# 0.00
35) Perylene-d12	23.913	264	28484	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.470	112	4128	0.293	ng	0.00
5) Phenol-d6	7.081	99	5554	0.314	ng	0.00
8) Nitrobenzene-d5	9.074	82	3951	0.247	ng	-0.01
11) 2-Methylnaphthalene-d10	12.306	152	7945	0.299	ng	0.00
14) 2,4,6-Tribromophenol	16.045	330	1665	0.264	ng	-0.01
15) 2-Fluorobiphenyl	13.176	172	11476	0.271	ng	0.00
27) Fluoranthene-d10	19.334	212	18886	0.262	ng	0.00
31) Terphenyl-d14	19.928	244	14111	0.236	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.318	88	1915	0.315	ng	# 56
3) n-Nitrosodimethylamine	3.650	42	3604	0.388	ng	# 98
6) bis(2-Chloroethyl)ether	7.333	93	6600	0.423	ng	99
9) Naphthalene	10.761	128	35409	0.709	ng	99
10) Hexachlorobutadiene	11.049	225	3606	0.382	ng	# 98
12) 2-Methylnaphthalene	12.381	142	20295	0.575	ng	98
16) Acenaphthylene	14.266	152	42041	0.762	ng	99
17) Acenaphthene	14.609	154	18586	0.541	ng	99
18) Fluorene	15.603	166	26008	0.553	ng	99
20) 4,6-Dinitro-2-methylph...	15.710	198	1131	0.458	ng	# 51
21) 4-Bromophenyl-phenylether	16.492	248	6132	0.377	ng	# 80
22) Hexachlorobenzene	16.603	284	7116	0.406	ng	100
23) Atrazine	16.752	200	5564	0.391	ng	97
24) Pentachlorophenol	16.963	266	2594	0.648	ng	91
25) Phenanthrene	17.336	178	181351	2.250	ng	100
26) Anthracene	17.435	178	62575	0.829	ng	100
28) Fluoranthene	19.363	202	426281	4.233	ng	99
30) Pyrene	19.726	202	395742	3.910	ng	97
32) Benzo(a)anthracene	21.473	228	244483	2.443	ng	100
33) Chrysene	21.527	228	246612	2.476	ng	98
34) Bis(2-ethylhexyl)phtha...	21.392	149	49607	0.632	ng	99
36) Indeno(1,2,3-cd)pyrene	26.427	276	141721	1.241	ng	96
37) Benzo(b)fluoranthene	23.182	252	306507	3.395	ng	94
39) Benzo(a)pyrene	23.702	252	131508	1.422	ng	96
40) Dibenzo(a,h)anthracene	26.442	278	53622	0.595	ng	98
41) Benzo(g,h,i)perylene	27.208	276	132003	1.333	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
Data File : BN025311.D
Acq On : 06 May 2023 12:22
Operator : CG/JU
Sample : 02558-08MS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Quant Time: May 08 04:20:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 08 04:14:26 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.911 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN025311.D
 Acq: 06 May 2023 12:22

Instrument :
 BNA_N
 ClientSampleId :
 ZONE-A-3-6-05022023MS






Tgt Ion:152 Resp: 5797
 Ion Ratio Lower Upper
 152 100
 150 151.5 123.8 185.8
 115 61.6 52.9 79.3



#2
 1,4-Dioxane
 Concen: 0.315 ng
 RT: 3.318 min Scan# 19
 Delta R.T. -0.007 min
 Lab File: BN025311.D
 Acq: 06 May 2023 12:22

Tgt Ion: 88 Resp: 1915
 Ion Ratio Lower Upper
 88 100
 43 121.6 37.2 55.8#
 58 97.0 71.5 107.3









#3
n-Nitrosodimethylamine
Concen: 0.388 ng
RT: 3.650 min Scan# 61
Delta R.T. -0.014 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS

Tgt Ion: 42 Resp: 3604
Ion Ratio Lower Upper
42 100
74 101.6 79.9 119.9
44 6.2 7.0 10.6#



#4
2-Fluorophenol
Concen: 0.293 ng
RT: 5.470 min Scan# 317
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion: 112 Resp: 4128
Ion Ratio Lower Upper
112 100
64 63.9 51.5 77.3
63 31.4 27.8 41.6





#5

Phenol-d6

Concen: 0.314 ng

RT: 7.081 min Scan# 541

Delta R.T. 0.000 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS





Tgt Ion: 99 Resp: 5554

Ion Ratio Lower Upper

99 100

42 33.3 22.0 33.0#

71 37.6 28.7 43.1



#6

bis(2-Chloroethyl)ether

Concen: 0.423 ng

RT: 7.333 min Scan# 575

Delta R.T. 0.000 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion: 93 Resp: 6600

Ion Ratio Lower Upper

93 100

63 83.4 67.8 101.6

95 33.4 26.7 40.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.718 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:	136	Resp:	19472
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.4	14.2
54	10.6	8.5	12.7
68	7.2	5.8	8.8



#8
Nitrobenzene-d5
Concen: 0.247 ng
RT: 9.074 min Scan# 782
Delta R.T. -0.011 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:	82	Resp:	3951
Ion Ratio	Lower	Upper	
82	100		
128	37.8	35.3	52.9
54	73.3	48.7	73.1#





#9

Naphthalene

Concen: 0.709 ng

RT: 10.761 min Scan# 941

Delta R.T. -0.011 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS





Tgt Ion:128 Resp: 35409

Ion Ratio Lower Upper

128 100

129 11.2 9.4 14.2

127 13.6 11.0 16.6



#10

Hexachlorobutadiene

Concen: 0.382 ng

RT: 11.049 min Scan# 967

Delta R.T. 0.000 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion:225 Resp: 3606


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.8 49.8 74.8









#11
2-Methylnaphthalene-d10
Concen: 0.299 ng
RT: 12.306 min Scan# 11
Delta R.T. -0.004 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:152 Resp: 7945
Ion Ratio Lower Upper
152 100
151 19.9 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.575 ng
RT: 12.381 min Scan# 1241
Delta R.T. -0.004 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:142 Resp: 20295
Ion Ratio Lower Upper
142 100
141 89.3 70.5 105.7
115 37.7 31.6 47.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.544 min Scan# 1550
Delta R.T. -0.011 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:164 Resp: 12327
Ion Ratio Lower Upper
164 100
162 106.2 82.1 123.1
160 50.9 38.3 57.5



#14
2,4,6-Tribromophenol
Concen: 0.264 ng
RT: 16.045 min Scan# 1686
Delta R.T. -0.012 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:330 Resp: 1665
Ion Ratio Lower Upper
330 100
332 97.5 77.4 116.0
141 40.7 28.9 43.3









#15
2-Fluorobiphenyl
Concen: 0.271 ng
RT: 13.176 min Scan# 1423
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:172 Resp: 11476
Ion Ratio Lower Upper
172 100
171 34.8 27.9 41.9
170 23.6 19.0 28.4



#16
Acenaphthylene
Concen: 0.762 ng
RT: 14.266 min Scan# 1524
Delta R.T. -0.011 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:152 Resp: 42041
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1
153 13.3 10.3 15.5





#17

Acenaphthene

Concen: 0.541 ng

RT: 14.609 min Scan# 1556

Delta R.T. -0.011 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS





Tgt Ion:154 Resp: 18586

Ion Ratio Lower Upper

154 100

153 112.9 91.3 136.9

152 56.5 44.7 67.1



#18

Fluorene

Concen: 0.553 ng

RT: 15.603 min Scan# 1649

Delta R.T. 0.000 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion:166 Resp: 26008


Ion Ratio Lower Upper

166 100

165 99.1 79.0 118.4

167 14.3 10.3 15.5









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:188 Resp: 29131
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.8 8.0 12.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.458 ng
RT: 15.710 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:198 Resp: 1131
Ion Ratio Lower Upper
198 100
51 105.2 162.1 243.1#
105 82.9 83.1 124.7#









#21
4-Bromophenyl-phenylether
Concen: 0.377 ng
RT: 16.492 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:248 Resp: 6132
Ion Ratio Lower Upper
248 100
250 102.4 76.0 114.0
141 43.7 59.9 89.9#



#22
Hexachlorobenzene
Concen: 0.406 ng
RT: 16.603 min Scan# 1731
Delta R.T. -0.012 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:284 Resp: 7116
Ion Ratio Lower Upper
284 100
142 37.2 30.0 45.0
249 31.1 25.0 37.6









#23
Atrazine
Concen: 0.391 ng
RT: 16.752 min Scan# 1743
Delta R.T. -0.012 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:200 Resp: 5564
Ion Ratio Lower Upper
200 100
173 36.7 26.3 39.5
215 46.6 37.2 55.8



#24
Pentachlorophenol
Concen: 0.648 ng
RT: 16.963 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:266 Resp: 2594
Ion Ratio Lower Upper
266 100
264 63.0 43.0 64.4
268 63.3 47.4 71.2









#25
Phenanthrene
Concen: 2.250 ng
RT: 17.336 min Scan# 1790
Delta R.T. -0.012 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:178 Resp: 181351
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.829 ng
RT: 17.435 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:178 Resp: 62575
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 14.9 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.262 ng
RT: 19.334 min Scan# 2160
Delta R.T. -0.004 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:212 Resp: 18886
Ion Ratio Lower Upper
212 100
106 15.9 12.6 19.0
104 9.6 7.0 10.6



#28
Fluoranthene
Concen: 4.233 ng
RT: 19.363 min Scan# 2166
Delta R.T. -0.004 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:202 Resp: 426281
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.8
203 17.3 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.491 min Scan# 2486

Delta R.T. -0.009 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS





Tgt Ion:240 Resp: 29606

Ion Ratio Lower Upper

240 100

120 18.4 10.3 15.5#

236 29.1 22.3 33.5



#30

Pyrene

Concen: 3.910 ng

RT: 19.726 min Scan# 2252

Delta R.T. -0.004 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion:202 Resp: 395742

Ion Ratio Lower Upper

202 100

200 21.1 16.8 25.2

203 20.8 14.4 21.6









#31
Terphenyl-d14
Concen: 0.236 ng
RT: 19.928 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:244 Resp: 14111
Ion Ratio Lower Upper
244 100
212 9.3 7.1 10.7
122 0.0 0.0 0.0



#32
Benzo(a)anthracene
Concen: 2.443 ng
RT: 21.473 min Scan# 2482
Delta R.T. -0.009 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:228 Resp: 244483
Ion Ratio Lower Upper
228 100
226 28.4 22.5 33.7
229 20.3 16.2 24.2





#33

Chrysene

Concen: 2.476 ng

RT: 21.527 min Scan# 2488

Delta R.T. -0.009 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS





Tgt Ion:228 Resp: 246612

Ion Ratio Lower Upper

228 100

226 31.6 24.6 36.8

229 20.9 16.2 24.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.632 ng

RT: 21.392 min Scan# 2473

Delta R.T. 0.000 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion:149 Resp: 49607


Ion Ratio Lower Upper

149 100

167 27.4 22.2 33.4

279 3.0 2.6 4.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.913 min Scan# 3109

Delta R.T. -0.023 min

Lab File: BN025311.D


Acq: 06 May 2023 12:22

Instrument :

BNA_N

ClientSampleId :

ZONE-A-3-6-05022023MS






Tgt Ion:264 Resp: 28484

Ion Ratio Lower Upper

264 100

260 27.2 21.5 32.3

265 50.9 46.2 69.4



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.241 ng

RT: 26.427 min Scan# 3962

Delta R.T. -0.052 min

Lab File: BN025311.D

Acq: 06 May 2023 12:22




Tgt Ion:276 Resp: 141721


Ion Ratio Lower Upper

276 100

138 24.7 22.5 33.7

277 24.7 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 3.395 ng
RT: 23.182 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22




Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS


Tgt Ion:252 Resp: 306507
Ion Ratio Lower Upper
252 100
253 23.1 20.8 31.2
125 14.0 13.4 20.2



#39
Benzo(a)pyrene
Concen: 1.422 ng
RT: 23.702 min Scan# 3030
Delta R.T. -0.131 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:252 Resp: 131508
Ion Ratio Lower Upper
252 100
253 23.6 21.4 32.0
125 19.7 15.4 23.2









#40
Dibenzo(a,h)anthracene
Concen: 0.595 ng
RT: 26.442 min Scan# 3980
Delta R.T. -0.055 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Instrument :
BNA_N
ClientSampleId :
ZONE-A-3-6-05022023MS

Tgt Ion:278 Resp: 53622
Ion Ratio Lower Upper
278 100
139 23.6 17.6 26.4
279 27.4 21.6 32.4



#41
Benzo(g,h,i)perylene
Concen: 1.333 ng
RT: 27.208 min Scan# 4229
Delta R.T. -0.050 min
Lab File: BN025311.D
Acq: 06 May 2023 12:22

Tgt Ion:276 Resp: 132003
Ion Ratio Lower Upper
276 100
277 24.8 20.1 30.1
138 25.5 21.0 31.6

