

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
 Data File : BN025313.D
 Acq On : 06 May 2023 13:33
 Operator : CG/JU
 Sample : 02588-14
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PILE-A-UNK-05022023

Quant Time: May 08 03:59:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 08 02:37:05 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 05/08/2023
 Supervised By :Jagrut Upadhyay 05/08/2023

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.911	152	5786	0.400 ng	0.00
7) Naphthalene-d8	10.718	136	18871	0.400 ng	-0.01
13) Acenaphthene-d10	14.555	164	12935	0.400 ng	0.00
19) Phenanthrene-d10	17.298	188	30394	0.400 ng	#-0.01
29) Chrysene-d12	21.495	240	29484	0.400 ng	#-0.01
35) Perylene-d12	23.932	264	28499	0.400 ng	0.00
System Monitoring Compounds					
4) 2-Fluorophenol	5.470	112	3552	0.252 ng	0.00
5) Phenol-d6	7.081	99	4606	0.261 ng	0.00
8) Nitrobenzene-d5	9.084	82	3214	0.207 ng	-0.01
11) 2-Methylnaphthalene-d10	12.309	152	5897	0.229 ng	0.00
14) 2,4,6-Tribromophenol	16.045	330	1422	0.215 ng	-0.02
15) 2-Fluorobiphenyl	13.176	172	9175	0.207 ng	-0.01
27) Fluoranthene-d10	19.334	212	14413	0.192 ng	0.00
31) Terphenyl-d14	19.928	244	11001	0.185 ng	0.00
Target Compounds					Qvalue
6) bis(2-Chloroethyl)ether	7.333	93	1088	0.070 ng	96
9) Naphthalene	10.771	128	36526	0.754 ng	98
12) 2-Methylnaphthalene	12.381	142	12642	0.369 ng	99
16) Acenaphthylene	14.266	152	10446	0.180 ng	97
17) Acenaphthene	14.609	154	2946	0.082 ng	94
18) Fluorene	15.603	166	5543	0.112 ng	# 95
25) Phenanthrene	17.336	178	69667	0.828 ng	99
26) Anthracene	17.435	178	15111	0.192 ng	98
28) Fluoranthene	19.363	202	156180	1.486 ng	100
30) Pyrene	19.726	202	141803	1.407 ng	# 96
32) Benzo(a)anthracene	21.477	228	81524	0.818 ng	98
33) Chrysene	21.531	228	79389	0.800 ng	97
34) Bis(2-ethylhexyl)phtha...	21.396	149	18627	0.238 ng	# 83
36) Indeno(1,2,3-cd)pyrene	26.472	276	36625	0.321 ng	95
37) Benzo(b)fluoranthene	23.195	252	99480	1.101 ng	96
38) Benzo(k)fluoranthene	23.236	252	34219m	0.362 ng	
39) Benzo(a)pyrene	23.829	252	66903	0.723 ng	97
40) Dibenzo(a,h)anthracene	26.484	278	9289	0.103 ng	# 67
41) Benzo(g,h,i)perylene	27.256	276	35682	0.360 ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050623\
Data File : BN025313.D
Acq On : 06 May 2023 13:33
Operator : CG/JU
Sample : 02588-14
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023

Quant Time: May 08 03:59:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 08 02:37:05 2023
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.911 min Scan# 655
Delta R.T. 0.000 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:152 Resp: 5780
Ion Ratio Lower Upper
152 100
150 152.0 123.8 185.8
115 64.5 52.9 79.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#4
2-Fluorophenol
Concen: 0.252 ng
RT: 5.470 min Scan# 317
Delta R.T. 0.000 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:112 Resp: 3552
Ion Ratio Lower Upper
112 100
64 66.0 51.5 77.3
63 32.0 27.8 41.6





#5
Phenol-d6
Concen: 0.261 ng
RT: 7.081 min Scan# 541
Delta R.T. -0.007 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion: 99 Resp: 4600
Ion Ratio Lower Upper
99 100
42 31.7 22.0 33.0
71 33.0 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#6
bis(2-Chloroethyl)ether
Concen: 0.070 ng
RT: 7.333 min Scan# 575
Delta R.T. -0.007 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33

Tgt Ion: 93 Resp: 1088
Ion Ratio Lower Upper
93 100
63 81.9 67.8 101.6
95 29.0 26.7 40.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.718 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33

Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023

Tgt Ion:136	Resp:	1887
Ion Ratio	Lower	Upper
136	100	
137	11.4	9.4 14.2
54	11.7	8.5 12.7
68	8.2	5.8 8.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#8
Nitrobenzene-d5
Concen: 0.207 ng
RT: 9.084 min Scan# 783
Delta R.T. -0.011 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion: 82	Resp:	3214
Ion Ratio	Lower	Upper
82	100	
128	46.3	35.3 52.9
54	67.0	48.7 73.1





#9
Naphthalene
Concen: 0.754 ng
RT: 10.771 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023




Tgt Ion:128 Resp: 36520
Ion Ratio Lower Upper
128 100
129 11.3 9.4 14.2
127 13.0 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023




#11
2-Methylnaphthalene-d10
Concen: 0.229 ng
RT: 12.309 min Scan# 1222
Delta R.T. -0.008 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Tgt Ion:152 Resp: 5897
Ion Ratio Lower Upper
152 100
151 20.7 16.4 24.6





#12
2-Methylnaphthalene
Concen: 0.369 ng
RT: 12.381 min Scan# 1241
Delta R.T. -0.008 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:142 Resp: 1264
Ion Ratio Lower Upper
142 100
141 88.8 70.5 105.7
115 38.9 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.555 min Scan# 1551
Delta R.T. 0.000 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:164 Resp: 12935
Ion Ratio Lower Upper
164 100
162 101.1 82.1 123.1
160 46.4 38.3 57.5





#14
2,4,6-Tribromophenol
Concen: 0.215 ng
RT: 16.045 min Scan# 1686
Delta R.T. -0.025 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion: 330 Resp: 142
Ion Ratio Lower Upper
330 100
332 101.7 77.4 116.0
141 38.3 28.9 43.3




Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 05/08/2023
Supervised By : Jagrut Upadhyay 05/08/2023



#15
2-Fluorobiphenyl
Concen: 0.207 ng
RT: 13.176 min Scan# 1422
Delta R.T. -0.011 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion: 172 Resp: 9175
Ion Ratio Lower Upper
172 100
171 36.0 27.9 41.9
170 23.8 19.0 28.4





#16
Acenaphthylene
Concen: 0.180 ng
RT: 14.266 min Scan# 1524
Delta R.T. -0.011 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument : BNA_N
ClientSampleId : PILE-A-UNK-05022023




Tgt Ion:152 Resp: 10440
Ion Ratio Lower Upper
152 100
151 20.9 16.1 24.1
153 14.9 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023




#17
Acenaphthene
Concen: 0.082 ng
RT: 14.609 min Scan# 1556
Delta R.T. -0.011 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Tgt Ion:154 Resp: 2946
Ion Ratio Lower Upper
154 100
153 117.6 91.3 136.9
152 64.8 44.7 67.1





#18
Fluorene
Concen: 0.112 ng
RT: 15.603 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:166 Resp: 554
Ion Ratio Lower Upper
166 100
165 102.1 79.0 118.4
167 19.3 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1787
Delta R.T. -0.012 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:188 Resp: 30394
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.7 8.0 12.0





#25
Phenanthrene
Concen: 0.828 ng
RT: 17.336 min Scan# 1790
Delta R.T. -0.012 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023




Tgt Ion:178 Resp: 6966
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.9 12.4 18.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023




#26
Anthracene
Concen: 0.192 ng
RT: 17.435 min Scan# 1798
Delta R.T. -0.012 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33



Tgt Ion:178 Resp: 15111
Ion Ratio Lower Upper
178 100
176 18.1 14.9 22.3
179 13.7 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.192 ng
RT: 19.334 min Scan# 2160
Delta R.T. -0.004 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33

Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023

Tgt Ion:212 Resp: 1441
Ion Ratio Lower Upper
212 100
106 17.0 12.6 19.0
104 9.6 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#28
Fluoranthene
Concen: 1.486 ng
RT: 19.363 min Scan# 2166
Delta R.T. -0.008 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:202 Resp: 156180
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.8
203 17.1 13.5 20.3





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.495 min Scan# 2486
Delta R.T. -0.014 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:240 Resp: 29484
Ion Ratio Lower Upper
240 100
120 19.0 10.3 15.5
236 29.0 22.3 33.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#30
Pyrene
Concen: 1.407 ng
RT: 19.726 min Scan# 2252
Delta R.T. -0.008 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:202 Resp: 141803
Ion Ratio Lower Upper
202 100
200 21.1 16.8 25.2
203 21.6 14.4 21.6#





#31
Terphenyl-d14
Concen: 0.185 ng
RT: 19.928 min Scan# 2110
Delta R.T. -0.008 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:244 Resp: 1100
Ion Ratio Lower Upper
244 100
212 10.4 7.1 10.7
122 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#32
Benzo(a)anthracene
Concen: 0.818 ng
RT: 21.477 min Scan# 2483
Delta R.T. -0.014 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:228 Resp: 81524
Ion Ratio Lower Upper
228 100
226 28.8 22.5 33.7
229 21.4 16.2 24.2





#33
Chrysene
Concen: 0.800 ng
RT: 21.531 min Scan# 2489
Delta R.T. -0.014 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:228 Resp: 79389
Ion Ratio Lower Upper
228 100
226 31.4 24.6 36.8
229 22.6 16.2 24.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.238 ng
RT: 21.396 min Scan# 2474
Delta R.T. -0.005 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33


Tgt Ion:149 Resp: 18627
Ion Ratio Lower Upper
149 100
167 18.3 22.2 33.4#
279 2.0 2.6 4.0#





#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.932 min Scan# 3109
 Delta R.T. -0.002 min
 Lab File: BN025313.D
 Acq: 06 May 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 PILE-A-UNK-05022023



Tgt Ion:264 Resp: 28495
 Ion Ratio Lower Upper
 264 100
 260 26.9 21.5 32.3
 265 49.0 46.2 69.4




Manual Integrations
 APPROVED


Reviewed By :Christian Giraldo 05/08/2023
 Supervised By :Jagrut Upadhyay 05/08/2023



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.321 ng
 RT: 26.472 min Scan# 3977
 Delta R.T. 0.016 min
 Lab File: BN025313.D
 Acq: 06 May 2023 13:33


Tgt Ion:276 Resp: 36625
 Ion Ratio Lower Upper
 276 100
 138 23.6 22.5 33.7
 277 24.6 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 1.101 ng
RT: 23.195 min Scan# 2856
Delta R.T. 0.001 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33




Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion:252 Resp: 99480
Ion Ratio Lower Upper
252 100
253 25.1 20.8 31.2
125 20.0 13.4 20.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/08/2023
Supervised By :Jagrut Upadhyay 05/08/2023



#38
Benzo(k)fluoranthene
Concen: 0.362 ng m
RT: 23.236 min Scan# 2870
Delta R.T. -0.005 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33

Tgt Ion:252 Resp: 34219
Ion Ratio Lower Upper
252 100
253 27.6 20.6 30.8
125 25.7 13.1 19.7#





#39

Benzo(a)pyrene

Concen: 0.723 ng

RT: 23.829 min Scan# 3073

Delta R.T. 0.001 min

Lab File: BN025313.D

Acq: 06 May 2023 13:33

Instrument :

BNA_N

ClientSampleId :

PILE-A-UNK-05022023





Tgt Ion:	252	Resp:	6690
Ion Ratio	Lower	Upper	
252	100		
253	25.6	21.4	32.0
125	21.5	15.4	23.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/08/2023

Supervised By :Jagrut Upadhyay 05/08/2023



#40

Dibenzo(a,h)anthracene

Concen: 0.103 ng




RT: 26.484 min Scan# 3981


Delta R.T. 0.004 min

Lab File: BN025313.D

Acq: 06 May 2023 13:33


Tgt Ion:	278	Resp:	9289
Ion Ratio	Lower	Upper	
278	100		
139	41.2	17.6	26.4#
279	41.3	21.6	32.4#





#41
Benzo(g,h,i)perylene
Concen: 0.360 ng
RT: 27.256 min Scan# 41
Delta R.T. 0.025 min
Lab File: BN025313.D
Acq: 06 May 2023 13:33

Instrument :
BNA_N
ClientSampleId :
PILE-A-UNK-05022023



Tgt Ion: 276 Resp: 3568

Ion	Ratio	Lower	Upper
276	100		
277	25.9	20.1	30.1
138	27.7	21.0	31.6

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/08/2023
Supervised By : Jagrut Upadhyay 05/08/2023

