

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
 Data File : BN014543.D
 Acq On : 07 May 2021 17:26
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050721

Quant Time: May 07 19:12:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 17:32:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.401	-0.3	106	0.00
3	n-Nitrosodimethylamine	0.400	0.078	80.5#	31	0.04
4 S	2-Fluorophenol	0.400	0.385	3.8	103	0.00
5 S	Phenol-d6	0.400	0.379	5.3	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	102	0.00
9	Naphthalene	0.400	0.374	6.5	100	0.00
10	Hexachlorobutadiene	0.400	0.377	5.8	103	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.369	7.8	100	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	98	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.372	7.0	101	0.00
16	Acenaphthylene	0.400	0.353	11.8	98	0.00
17	Acenaphthene	0.400	0.364	9.0	102	0.00
18	Fluorene	0.400	0.366	8.5	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.419	-4.7	113	0.00
21	4-Bromophenyl-phenylether	0.400	0.360	10.0	99	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	102	0.00
23	Atrazine	0.400	0.350	12.5	104	0.00
24	Pentachlorophenol	0.400	0.320	20.0	102	0.00
25	Phenanthrene	0.400	0.364	9.0	102	0.00
26	Anthracene	0.400	0.350	12.5	100	0.00
27 SURR	Fluoranthene-d10	0.400	0.356	11.0	102	0.00
28	Fluoranthene	0.400	0.358	10.5	103	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
30	Pyrene	0.400	0.361	9.8	104	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	104	0.00
32	Benzo(a)anthracene	0.400	0.357	10.8	106	0.00
33	Chrysene	0.400	0.366	8.5	107	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.307	23.3	99	0.00
35 I	Perylene-d12	0.400	0.400	0.0	108	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	110	0.01
37	Benzo(b)fluoranthene	0.400	0.357	10.8	106	0.00
38	Benzo(k)fluoranthene	0.400	0.376	6.0	111	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	111	0.00
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	110	0.01
41	Benzo(g,h,i)perylene	0.400	0.371	7.3	110	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN050721