

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
 Data File : BN014549.D
 Acq On : 07 May 2021 22:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 08 01:59:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 17:32:10 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.385	3.8	114	0.00
3	n-Nitrosodimethylamine	0.400	0.056	86.0#	25	-0.03
4 S	2-Fluorophenol	0.400	0.384	4.0	115	0.00
5 S	Phenol-d6	0.400	0.381	4.8	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	114	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.00
8 S	Nitrobenzene-d5	0.400	0.357	10.8	113	-0.01
9	Naphthalene	0.400	0.366	8.5	113	0.00
10	Hexachlorobutadiene	0.400	0.360	10.0	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	114	0.00
12	2-Methylnaphthalene	0.400	0.359	10.3	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	117	-0.01
16	Acenaphthylene	0.400	0.340	15.0	107	0.00
17	Acenaphthene	0.400	0.367	8.3	116	0.00
18	Fluorene	0.400	0.363	9.3	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.454	-13.5	151	0.00
21	4-Bromophenyl-phenylether	0.400	0.356	11.0	112	0.00
22	Hexachlorobenzene	0.400	0.366	8.5	113	0.00
23	Atrazine	0.400	0.310	22.5	106	0.00
24	Pentachlorophenol	0.400	0.331	17.3	120	0.00
25	Phenanthrene	0.400	0.358	10.5	115	0.00
26	Anthracene	0.400	0.337	15.8	110	-0.01
27 SURR	Fluoranthene-d10	0.400	0.342	14.5	112	0.00
28	Fluoranthene	0.400	0.343	14.2	113	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	117	0.00
30	Pyrene	0.400	0.368	8.0	115	0.00
31 S	Terphenyl-d14	0.400	0.377	5.8	113	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	115	-0.01
33	Chrysene	0.400	0.361	9.8	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.288	28.0#	101	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	120	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.373	6.8	124	0.00
37	Benzo(b)fluoranthene	0.400	0.354	11.5	117	0.00
38	Benzo(k)fluoranthene	0.400	0.372	7.0	123	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	120	0.00
40	Dibenzo(a,h)anthracene	0.400	0.374	6.5	124	0.00
41	Benzo(g,h,i)perylene	0.400	0.362	9.5	120	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4