

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
 Data File : BN014539.D
 Acq On : 07 May 2021 14:37
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

mohammad

5/11/2021 12:50:27 PM

Quant Time: May 07 15:16:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 14:41:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.757	152	654	0.40	ng	# 0.00
7) Naphthalene-d8	10.543	136	2378	0.40	ng	# 0.00
13) Acenaphthene-d10	14.397	164	1591	0.40	ng	0.00
19) Phenanthrene-d10	17.153	188	3568	0.40	ng	# 0.00
29) Chrysene-d12	21.345	240	4380	0.40	ng	# 0.00
35) Perylene-d12	23.633	264	4443	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.349	112	1455	1.29	ng	0.00
5) Phenol-d6	6.919	99	2045	1.56	ng	0.00
8) Nitrobenzene-d5	8.908	82	1535	0.95	ng	0.00
11) 2-Methylnaphthalene-d10	12.137	152	4434	1.27	ng	0.00
14) 2,4,6-Tribromophenol	15.894	330	543	1.44	ng	0.00
15) 2-Fluorobiphenyl	13.027	172	4655	0.81	ng	0.00
27) Fluoranthene-d10	19.188	212	12064	1.31	ng	0.00
31) Terphenyl-d14	19.794	244	8553	0.89	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.311	88	661	0.90	ng	96
3) n-Nitrosodimethylamine	3.690	42	196m	0.56	ng	
6) bis(2-Chloroethyl)ether	7.193	93	1473	1.00	ng	95
9) Naphthalene	10.594	128	4914	0.82	ng	96
10) Hexachlorobutadiene	10.885	225	1196	1.01	ng	# 99
12) 2-Methylnaphthalene	12.213	142	3504	0.94	ng	# 90
16) Acenaphthylene	14.118	152	4790	0.75	ng	97
17) Acenaphthene	14.460	154	3499	0.79	ng	96
18) Fluorene	15.450	166	4504	0.85	ng	98
20) 4,6-Dinitro-2-methylph...	15.526	198	215	0.95	ng	# 1
21) 4-Bromophenyl-phenylether	16.349	248	1739	0.96	ng	# 91
22) Hexachlorobenzene	16.459	284	1777	0.69	ng	92
23) Atrazine	16.617	200	1188	1.18	ng	94
24) Pentachlorophenol	16.800	266	615	1.74	ng	# 43
25) Phenanthrene	17.189	178	7488	0.77	ng	99
26) Anthracene	17.286	178	6774	0.92	ng	99
28) Fluoranthene	19.218	202	9639	0.92	ng	99
30) Pyrene	19.580	202	9840	0.71	ng	99
32) Benzo(a)anthracene	21.335	228	10416	1.01	ng	99
33) Chrysene	21.388	228	10829	0.80	ng	99
34) Bis(2-ethylhexyl)phtha...	21.282	149	3415	0.79	ng	# 82
36) Indeno(1,2,3-cd)pyrene	25.943	276	13256	0.86	ng	# 90
37) Benzo(b)fluoranthene	22.945	252	11787m	0.82	ng	
38) Benzo(k)fluoranthene	22.989	252	11405	0.64	ng	96
39) Benzo(a)pyrene	23.530	252	10109	0.71	ng	# 92
40) Dibenzo(a,h)anthracene	25.964	278	11473	0.93	ng	# 95
41) Benzo(g,h,i)perylene	26.642	276	11277	0.74	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
Data File : BN014539.D
Acq On : 07 May 2021 14:37
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Manual Integrations
APPROVED

mohammad

5/11/2021 12:50:27 PM

Quant Time: May 07 15:16:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 07 14:41:36 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.757 min Scan# 618
 Delta R.T. -0.000 min
 Lab File: BN014539.D
 Acq: 07 May 2021 14:37





Tgt Ion:	152	Resp:	654
Ion	Ratio	Lower	Upper
152	100		
150	147.6	123.8	185.8
115	89.7	47.0	70.4#

Instrument :
 BNA_N
 ClientSampled :
 SSTDICC0.8

Manual Integrations
 APPROVED




mohammad

5/11/2021 12:50:27 PM




#2
 1,4-Dioxane
 Concen: 0.90 ng
 RT: 3.311 min Scan# 66
 Delta R.T. -0.008 min
 Lab File: BN014539.D
 Acq: 07 May 2021 14:37

Tgt Ion:	88	Resp:	661
Ion	Ratio	Lower	Upper
88	100		
58	58.9	44.4	66.6
43	18.2	15.9	23.9



Abundance Scan 118 (3.730 min): BN014538.D\data.ms (-10



#3

n-Nitrosodimethylamine
Concen: 0.56 ng m
RT: 3.690 min Scan# 113
Delta R.T. -0.040 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:	42	Resp:	196
Ion	Ratio	Lower	Upper
42	100		
74	0.0	175.1	262.7#
44	0.0	7.5	11.3#



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED


mohammad

5/11/2021 12:50:27 PM


Abundance Scan 113 (3.690 min): BN014539.D\data.ms



Abundance Scan 113 (3.690 min): BN014539.D\data.ms (-10



Abundance Scan 320 (5.357 min): BN014538.D\data.ms (-31





#4


2-Fluorophenol
Concen: 1.29 ng
RT: 5.349 min Scan# 319
Delta R.T. -0.008 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion:	112	Resp:	1455
Ion	Ratio	Lower	Upper
112	100		
64	62.6	41.8	62.6
63	33.8	22.0	33.0#

Abundance Scan 319 (5.349 min): BN014539.D\data.ms



Abundance Scan 319 (5.349 min): BN014539.D\data.ms (-30





#5
Phenol-d6
Concen: 1.56 ng
RT: 6.919 min Scan# 514
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion: 99 Resp: 2045
Ion Ratio Lower Upper
99 100
42 12.9 13.2 19.8#
71 43.9 26.2 39.2#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#6
bis(2-Chloroethyl)ether
Concen: 1.00 ng
RT: 7.193 min Scan# 548
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion: 93 Resp: 1473
Ion Ratio Lower Upper
93 100
63 69.3 58.0 87.0
95 37.9 27.0 40.6





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Tgt Ion: 136 Resp: 2378
Ion Ratio Lower Upper
136 100
137 14.9 8.9 13.3#
54 11.0 3.8 5.8#
68 11.2 3.8 5.6#

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#8
Nitrobenzene-d5
Concen: 0.95 ng
RT: 8.908 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion: 82 Resp: 1535
Ion Ratio Lower Upper
82 100
128 39.0 29.5 44.3
54 37.8 36.9 55.3





#9
Naphthalene
Concen: 0.82 ng
RT: 10.594 min Scan# 867
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:	128	Resp:	4914
Ion	Ratio	Lower	Upper
128	100		
129	12.4	9.0	13.4
127	15.1	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#10
Hexachlorobutadiene
Concen: 1.01 ng
RT: 10.885 min Scan# 890
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:	225	Resp:	1196
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.5	50.9	76.3





#11
2-Methylnaphthalene-d10
Concen: 1.27 ng
RT: 12.137 min Scan# 1081
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:152 Resp: 4434
Ion Ratio Lower Upper
152 100
151 21.1 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#12
2-Methylnaphthalene
Concen: 0.94 ng
RT: 12.213 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion:142 Resp: 3504
Ion Ratio Lower Upper
142 100
141 90.5 70.2 105.4
115 49.2 26.7 40.1#





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.397 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Tgt Ion:164 Resp: 1591
Ion Ratio Lower Upper
164 100
162 102.5 78.6 118.0
160 48.8 35.8 53.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#14
2,4,6-Tribromophenol
Concen: 1.44 ng
RT: 15.894 min Scan# 1431
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:330 Resp: 543
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.4 27.0 40.4





#15
2-Fluorobiphenyl
Concen: 0.81 ng
RT: 13.027 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:	172	Resp:	4655
Ion	Ratio	Lower	Upper
172	100		
171	36.6	27.1	40.7
170	23.4	18.3	27.5

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#16
Acenaphthylene
Concen: 0.75 ng
RT: 14.118 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion:	152	Resp:	4790
Ion	Ratio	Lower	Upper
152	100		
151	21.4	15.6	23.4
153	13.3	10.5	15.7





#17
Acenaphthene
Concen: 0.79 ng
RT: 14.460 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:154 Resp: 3499
Ion Ratio Lower Upper
154 100
153 110.5 90.2 135.4
152 52.9 46.2 69.2


Manual Integrations
APPROVED

mohammad



5/11/2021 12:50:27 PM




#18
Fluorene
Concen: 0.85 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37




Tgt Ion:166 Resp: 4504
Ion Ratio Lower Upper
166 100
165 99.7 78.4 117.6
167 13.4 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.153 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Tgt Ion:188 Resp: 3568
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.9 5.2 7.8#


Manual Integrations
APPROVED

mohammad



5/11/2021 12:50:27 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.95 ng
RT: 15.526 min Scan# 1402
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37




Tgt Ion:198 Resp: 215
Ion Ratio Lower Upper
198 100
51 59.7 865.3 1297.9#
105 69.9 93.6 140.4#








#21
4-Bromophenyl-phenylether
Concen: 0.96 ng
RT: 16.349 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:248 Resp: 1739
Ion Ratio Lower Upper
248 100
250 99.0 80.5 120.7
141 62.2 37.2 55.8#




#22
Hexachlorobenzene
Concen: 0.69 ng
RT: 16.459 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37




Tgt Ion:284 Resp: 1777
Ion Ratio Lower Upper
284 100
142 29.8 26.6 39.8
249 34.5 23.6 35.4





#23
Atrazine
Concen: 1.18 ng
RT: 16.617 min Scan# 1490
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Tgt Ion:200 Resp: 1188
Ion Ratio Lower Upper
200 100
173 32.8 23.0 34.6
215 51.3 38.6 57.8


Manual Integrations
APPROVED

mohammad



5/11/2021 12:50:27 PM




#24
Pentachlorophenol
Concen: 1.74 ng
RT: 16.800 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37



Tgt Ion:266 Resp: 615
Ion Ratio Lower Upper
266 100
264 57.2 24.2 36.4#
268 67.5 25.2 37.8#





#25
Phenanthrene
Concen: 0.77 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:	178	Resp:	7488
Ion Ratio	Lower	Upper	
178	100		
176	19.5	15.2	22.8
179	15.7	12.4	18.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#26
Anthracene
Concen: 0.92 ng
RT: 17.286 min Scan# 1545
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:	178	Resp:	6774
Ion Ratio	Lower	Upper	
178	100		
176	18.7	14.8	22.2
179	15.6	12.2	18.2





#27
Fluoranthene-d10
Concen: 1.31 ng
RT: 19.188 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion	Ratio	Lower	Upper
212	100		
106	10.1	10.1	15.1#
104	6.1	5.4	8.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM




#28
Fluoranthene
Concen: 0.92 ng
RT: 19.218 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.8	8.4	12.6
203	17.3	13.5	20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.345 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:240 Resp: 4380
Ion Ratio Lower Upper
240 100
120 13.8 5.8 8.8#
236 28.6 20.6 30.8



#30
Pyrene
Concen: 0.71 ng
RT: 19.580 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37



Tgt Ion:202 Resp: 9840
Ion Ratio Lower Upper
202 100
200 21.1 16.3 24.5
203 17.7 14.0 21.0




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED

mohammad


5/11/2021 12:50:27 PM



#31
Terphenyl-d14
Concen: 0.89 ng
RT: 19.794 min Scan# 1963
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37



Tgt Ion:244 Resp: 8553
Ion Ratio Lower Upper
244 100
212 7.8 6.0 9.0
122 8.8 9.3 13.9#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED

mohammad



5/11/2021 12:50:27 PM




#32
Benzo(a)anthracene
Concen: 1.01 ng
RT: 21.335 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37



Tgt Ion:228 Resp: 10416
Ion Ratio Lower Upper
228 100
226 27.3 21.7 32.5
229 20.4 15.7 23.5





#33
Chrysene
Concen: 0.80 ng
RT: 21.388 min Scan# 2136
Delta R.T. 0.010 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:	228	Resp:	10829
Ion Ratio	Lower	Upper	
228	100		
226	29.6	23.8	35.6
229	20.6	15.6	23.4

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.79 ng
RT: 21.282 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion:	149	Resp:	3415
Ion Ratio	Lower	Upper	
149	100		
167	35.3	20.5	30.7#
279	7.7	2.9	4.3#





#35
Perylene-d12
Concen: 0.40 ng
RT: 23.633 min Scan# 2590
Delta R.T. 0.003 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Instrument :
BNA_N
ClientSampled :
SSTDIC0.8






Tgt Ion:264 Resp: 4443
Ion Ratio Lower Upper
264 100
260 27.1 19.5 29.3
265 100.8 35.4 53.0#

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.86 ng
RT: 25.943 min Scan# 3265
Delta R.T. 0.003 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:276 Resp: 13256
Ion Ratio Lower Upper
276 100
138 14.6 18.8 28.2#
277 25.5 19.8 29.8





#37
Benzo(b)fluoranthene
Concen: 0.82 ng m
RT: 22.945 min Scan# 2389
Delta R.T. 0.003 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:252 Resp: 11787
Ion Ratio Lower Upper
252 100
253 26.2 19.6 29.4
125 8.7 10.3 15.5#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#38
Benzo(k)fluoranthene
Concen: 0.64 ng
RT: 22.989 min Scan# 2402
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:252 Resp: 11405
Ion Ratio Lower Upper
252 100
253 27.1 19.7 29.5
125 11.7 9.7 14.5





#39
Benzo(a)pyrene
Concen: 0.71 ng
RT: 23.530 min Scan# 2560
Delta R.T. 0.003 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37





Tgt Ion:	252	Resp:	10109
Ion Ratio	Lower	Upper	
252	100		
253	27.8	21.8	32.6
125	10.1	14.9	22.3#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad


5/11/2021 12:50:27 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.93 ng
RT: 25.964 min Scan# 3271
Delta R.T. 0.007 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37


Tgt Ion:	278	Resp:	11473
Ion Ratio	Lower	Upper	
278	100		
139	13.3	13.9	20.9#
279	27.0	20.7	31.1





#41
Benzo(g,h,i)perylene
Concen: 0.74 ng
RT: 26.642 min Scan# 3469
Delta R.T. -0.000 min
Lab File: BN014539.D
Acq: 07 May 2021 14:37

Tgt Ion:	276	Resp:	11277
Ion Ratio	Lower	Upper	
276	100		
277	26.7	19.6	29.4
138	13.1	17.5	26.3#



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

mohammad

5/11/2021 12:50:27 PM