

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
 Data File : BN014543.D
 Acq On : 07 May 2021 17:26
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050721

Manual Integrations
 APPROVED

mohammad
 5/11/2021 2:04:20 PM

Quant Time: May 07 19:12:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 17:32:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.765	152	758	0.40	ng	# 0.00
7) Naphthalene-d8	10.543	136	2731	0.40	ng	# 0.00
13) Acenaphthene-d10	14.397	164	1851	0.40	ng	0.00
19) Phenanthrene-d10	17.153	188	4073	0.40	ng	# 0.00
29) Chrysene-d12	21.345	240	5046	0.40	ng	# 0.00
35) Perylene-d12	23.636	264	5028	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.357	112	816	0.38	ng	0.00
5) Phenol-d6	6.919	99	1118	0.38	ng	0.00
8) Nitrobenzene-d5	8.908	82	832	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.142	152	2396	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.894	330	281	0.34	ng	0.00
15) 2-Fluorobiphenyl	13.027	172	2505	0.37	ng	0.00
27) Fluoranthene-d10	19.188	212	6224	0.36	ng	0.00
31) Terphenyl-d14	19.799	244	4700	0.38	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.319	88	383	0.40	ng	98
3) n-Nitrosodimethylamine	3.770	42	26	0.08	ng	# 1
6) bis(2-Chloroethyl)ether	7.193	93	810	0.38	ng	95
9) Naphthalene	10.594	128	2708	0.37	ng	# 93
10) Hexachlorobutadiene	10.898	225	655	0.38	ng	# 97
12) 2-Methylnaphthalene	12.217	142	1872	0.36	ng	# 90
16) Acenaphthylene	14.118	152	2531	0.35	ng	98
17) Acenaphthene	14.460	154	1820	0.36	ng	99
18) Fluorene	15.450	166	2413	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.526	198	110	0.42	ng	# 1
21) 4-Bromophenyl-phenylether	16.349	248	909	0.36	ng	# 88
22) Hexachlorobenzene	16.459	284	956	0.38	ng	92
23) Atrazine	16.617	200	605	0.35	ng	# 90
24) Pentachlorophenol	16.800	266	280	0.32	ng	# 36
25) Phenanthrene	17.189	178	3955	0.36	ng	99
26) Anthracene	17.286	178	3474	0.35	ng	99
28) Fluoranthene	19.218	202	4955	0.36	ng	98
30) Pyrene	19.585	202	5108	0.36	ng	99
32) Benzo(a)anthracene	21.335	228	5388	0.36	ng	100
33) Chrysene	21.388	228	5619	0.37	ng	98
34) Bis(2-ethylhexyl)phtha...	21.282	149	1671	0.31	ng	# 82
36) Indeno(1,2,3-cd)pyrene	25.950	276	7010	0.37	ng	# 91
37) Benzo(b)fluoranthene	22.948	252	5945	0.36	ng	# 91
38) Benzo(k)fluoranthene	22.993	252	6038m	0.38	ng	
39) Benzo(a)pyrene	23.534	252	5314	0.37	ng	# 91
40) Dibenzo(a,h)anthracene	25.967	278	6071	0.37	ng	94
41) Benzo(g,h,i)perylene	26.652	276	5992	0.37	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050721\
Data File : BN014543.D
Acq On : 07 May 2021 17:26
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM

Quant Time: May 07 19:12:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 07 17:32:10 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.765 min Scan# 619
Delta R.T. 0.008 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion: 152 Resp: 758
Ion Ratio Lower Upper
152 100
150 145.6 123.8 185.8
115 81.7 47.0 70.4#



#2
1,4-Dioxane
Concen: 0.40 ng
RT: 3.319 min Scan# 67
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26



Tgt Ion: 88 Resp: 383
Ion Ratio Lower Upper
88 100
58 56.7 44.4 66.6
43 19.1 15.9 23.9




Instrument :
BNA_N
ClientSampled :
ICVBN050721

Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM

Abundance Scan 118 (3.730 min): BN014538.D\data.ms (-10







#3
n-Nitrosodimethylamine
Concen: 0.08 ng
RT: 3.770 min Scan# 123
Delta R.T. 0.040 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion: 42 Resp: 26
Ion Ratio Lower Upper
42 100
74 0.0 175.1 262.7#
44 69.2 7.5 11.3#

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#4
2-Fluorophenol
Concen: 0.38 ng
RT: 5.357 min Scan# 320
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion: 112 Resp: 816
Ion Ratio Lower Upper
112 100
64 60.8 41.8 62.6
63 35.5 22.0 33.0#









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.919 min Scan# 514
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	99	Resp:	1118
Ion	Ratio	Lower	Upper
99	100		
42	13.0	13.2	19.8#
71	46.9	26.2	39.2#

Instrument :
BNA_N
ClientSampled :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.193 min Scan# 548
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26


Tgt Ion:	93	Resp:	810
Ion	Ratio	Lower	Upper
93	100		
63	67.9	58.0	87.0
95	31.5	27.0	40.6





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26



Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Tgt Ion:	136	Resp:	2731
Ion	Ratio	Lower	Upper
136	100		
137	14.0	8.9	13.3#
54	12.0	3.8	5.8#
68	10.9	3.8	5.6#

Manual Integrations
APPROVED



mohammad
5/11/2021 2:04:20 PM




#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.908 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26



Tgt Ion:	82	Resp:	832
Ion	Ratio	Lower	Upper
82	100		
128	34.1	29.5	44.3
54	46.6	36.9	55.3









#9
Naphthalene
Concen: 0.37 ng
RT: 10.594 min Scan# 867
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	128	Resp:	2708
Ion Ratio	Lower	Upper	
128	100		
129	13.2	9.0	13.4
127	16.2	10.4	15.6#

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.898 min Scan# 891
Delta R.T. 0.012 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	225	Resp:	655
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	61.5	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.142 min Scan# 1082
Delta R.T. 0.004 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:152 Resp: 2396
Ion Ratio Lower Upper
152 100
151 21.0 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM




#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.217 min Scan# 1099
Delta R.T. 0.004 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:142 Resp: 1872
Ion Ratio Lower Upper
142 100
141 89.4 70.2 105.4
115 51.3 26.7 40.1#








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.397 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:164 Resp: 1851

Ion	Ratio	Lower	Upper
164	100		
162	107.7	78.6	118.0
160	47.8	35.8	53.8





#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.894 min Scan# 1431
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26



Tgt Ion:330 Resp: 281


Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	30.6	27.0	40.4



Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM







#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 13.027 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	172	Resp:	2505
Ion Ratio	Lower	Upper	
172	100		
171	37.9	27.1	40.7
170	24.6	18.3	27.5

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM




#16
Acenaphthylene
Concen: 0.35 ng
RT: 14.118 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	152	Resp:	2531
Ion Ratio	Lower	Upper	
152	100		
151	21.2	15.6	23.4
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.36 ng
RT: 14.460 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:154 Resp: 1820
Ion Ratio Lower Upper
154 100
153 112.0 90.2 135.4
152 56.1 46.2 69.2



#18
Fluorene
Concen: 0.37 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:166 Resp: 2413
Ion Ratio Lower Upper
166 100
165 98.5 78.4 117.6
167 14.5 10.7 16.1




Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED




mohammad
5/11/2021 2:04:20 PM




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.153 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:188 Resp: 4073
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.5 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.42 ng
RT: 15.526 min Scan# 1402
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:198 Resp: 110
Ion Ratio Lower Upper
198 100
51 80.5 865.3 1297.9#
105 85.6 93.6 140.4#



Instrument :
BNA_N
ClientSampleId :
ICVBN050721


Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM



#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.349 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Tgt Ion:248 Resp: 909
Ion Ratio Lower Upper
248 100
250 100.3 80.5 120.7
141 71.7 37.2 55.8#

Manual Integrations
APPROVED



mohammad
5/11/2021 2:04:20 PM




#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.459 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:284 Resp: 956
Ion Ratio Lower Upper
284 100
142 28.7 26.6 39.8
249 33.9 23.6 35.4









#23
Atrazine
Concen: 0.35 ng
RT: 16.617 min Scan# 1490
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:200 Resp: 605
Ion Ratio Lower Upper
200 100
173 34.7 23.0 34.6#
215 54.4 38.6 57.8



#24
Pentachlorophenol
Concen: 0.32 ng
RT: 16.800 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:266 Resp: 280
Ion Ratio Lower Upper
266 100
264 58.9 24.2 36.4#
268 72.5 25.2 37.8#




Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED




mohammad
5/11/2021 2:04:20 PM




#25
Phenanthrene
Concen: 0.36 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:178 Resp: 3955
Ion Ratio Lower Upper
178 100
176 19.6 15.2 22.8
179 15.9 12.4 18.6



#26
Anthracene
Concen: 0.35 ng
RT: 17.286 min Scan# 1545
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:178 Resp: 3474
Ion Ratio Lower Upper
178 100
176 18.7 14.8 22.2
179 15.6 12.2 18.2




Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED




mohammad
5/11/2021 2:04:20 PM




#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.188 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:212 Resp: 6224
Ion Ratio Lower Upper
212 100
106 9.9 10.1 15.1#
104 6.2 5.4 8.2



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.218 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:202 Resp: 4955
Ion Ratio Lower Upper
202 100
101 9.4 8.4 12.6
203 17.8 13.5 20.3




Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED




mohammad
5/11/2021 2:04:20 PM




#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.345 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:240 Resp: 5046
Ion Ratio Lower Upper
240 100
120 15.5 5.8 8.8#
236 29.2 20.6 30.8



#30
Pyrene
Concen: 0.36 ng
RT: 19.585 min Scan# 1921
Delta R.T. 0.005 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:202 Resp: 5108
Ion Ratio Lower Upper
202 100
200 21.0 16.3 24.5
203 18.0 14.0 21.0



Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM







#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.799 min Scan# 1964
Delta R.T. 0.005 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:244 Resp: 4700
Ion Ratio Lower Upper
244 100
212 8.4 6.0 9.0
122 8.2 9.3 13.9#

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.335 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:228 Resp: 5388
Ion Ratio Lower Upper
228 100
226 27.2 21.7 32.5
229 19.9 15.7 23.5









#33
Chrysene
Concen: 0.37 ng
RT: 21.388 min Scan# 2136
Delta R.T. 0.010 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	228	Resp:	5619
Ion Ratio	Lower	Upper	
228	100		
226	30.9	23.8	35.6
229	20.7	15.6	23.4

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.31 ng
RT: 21.282 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	149	Resp:	1671
Ion Ratio	Lower	Upper	
149	100		
167	35.0	20.5	30.7#
279	6.9	2.9	4.3#









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.636 min Scan# 2591
Delta R.T. 0.007 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26





Tgt Ion:264 Resp: 5028
Ion Ratio Lower Upper
264 100
260 26.5 19.5 29.3
265 87.4 35.4 53.0#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 25.950 min Scan# 3267
Delta R.T. 0.010 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26




Tgt Ion:276 Resp: 7010
Ion Ratio Lower Upper
276 100
138 15.0 18.8 28.2#
277 25.7 19.8 29.8



Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM







#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.948 min Scan# 2390
Delta R.T. 0.007 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion	Ratio	Lower	Upper
252	100		
253	29.6	19.6	29.4#
125	10.7	10.3	15.5

Instrument :
BNA_N
ClientSampled :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#38
Benzo(k)fluoranthene
Concen: 0.38 ng m
RT: 22.993 min Scan# 2403
Delta R.T. 0.003 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion	Ratio	Lower	Upper
252	100		
253	30.4	19.7	29.5#
125	14.6	9.7	14.5#









#39
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.534 min Scan# 2561
Delta R.T. 0.007 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	252	Resp:	5314
Ion Ratio	Lower	Upper	
252	100		
253	30.8	21.8	32.6
125	13.1	14.9	22.3#

Instrument :
BNA_N
ClientSampleId :
ICVBN050721




Manual Integrations
APPROVED


mohammad
5/11/2021 2:04:20 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.967 min Scan# 3272
Delta R.T. 0.010 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26


Tgt Ion:	278	Resp:	6071
Ion Ratio	Lower	Upper	
278	100		
139	16.3	13.9	20.9
279	29.9	20.7	31.1





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.652 min Scan# 3472
Delta R.T. 0.010 min
Lab File: BN014543.D
Acq: 07 May 2021 17:26

Tgt Ion:	276	Resp:	5992
Ion Ratio	Lower	Upper	
276	100		
277	27.2	19.6	29.4
138	14.0	17.5	26.3#



Instrument :
BNA_N
ClientSampleId :
ICVBN050721

Manual Integrations
APPROVED

mohammad
5/11/2021 2:04:20 PM