

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050823\
 Data File : BN025331.D
 Acq On : 08 May 2023 14:07
 Operator : CG/JU
 Sample : 02588-12
 Misc :
 ALS Vial : 4 Sample Multiplier: 50

Instrument :
 BNA_N
 ClientSampleId :
 ZONE-D-6-9-05022023

Quant Time: May 09 04:02:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 08 18:27:02 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 05/09/2023
 Supervised By :Jagrut Upadhyay 05/09/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.897	152	281038	20.000	ng	0.00
7) Naphthalene-d8	10.707	136	916568	20.000	ng	-0.01
13) Acenaphthene-d10	14.544	164	586222	20.000	ng	0.00
19) Phenanthrene-d10	17.299	188	1353557	20.000	ng	0.00
29) Chrysene-d12	21.500	240	1309778	20.000	ng	0.00
35) Perylene-d12	23.928	264	1661966	20.000	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.456	112	1387069	101.385	ng	0.00
5) Phenol-d6	7.066	99	2000522	116.532	ng	0.00
8) Nitrobenzene-d5	9.063	82	990422	65.687	ng	-0.02
11) 2-Methylnaphthalene-d10	12.272	152	85	0.003	ng	-0.03
14) 2,4,6-Tribromophenol	16.045	330	731966	122.122	ng	-0.01
15) 2-Fluorobiphenyl	13.176	172	2184594	54.334	ng	0.00
27) Fluoranthene-d10	19.325	212	112	0.002	ng	0.00
31) Terphenyl-d14	19.928	244	2845973	53.883	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.319	93	1054	0.070	ng	84
9) Naphthalene	10.761	128	47189	1.003	ng	99
12) 2-Methylnaphthalene	12.377	142	14037	0.422	ng	99
16) Acenaphthylene	14.267	152	65560	1.249	ng	99
17) Acenaphthene	14.609	154	15334	0.470	ng	96
18) Fluorene	15.603	166	43947	0.982	ng	# 97
25) Phenanthrene	17.336	178	530994	7.089	ng	99
26) Anthracene	17.435	178	164164	2.340	ng	100
28) Fluoranthene	19.363	202	857001	9.157	ng	99
30) Pyrene	19.726	202	702652	7.846	ng	98
32) Benzo(a)anthracene	21.482	228	428445	4.839	ng	97
33) Chrysene	21.536	228	354904	4.027	ng	97
34) Bis(2-ethylhexyl)phtha...	21.392	149	53499	0.770	ng	100
36) Indeno(1,2,3-cd)pyrene	26.442	276	249247	1.870	ng	95
37) Benzo(b)fluoranthene	23.188	252	498273	4.729	ng	95
38) Benzo(k)fluoranthene	23.232	252	177383m	1.608	ng	
39) Benzo(a)pyrene	23.817	252	390796m	3.620	ng	
40) Dibenzo(a,h)anthracene	26.451	278	62500	0.594	ng	# 85
41) Benzo(g,h,i)perylene	27.220	276	233518	2.021	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050823\
Data File : BN025331.D
Acq On : 08 May 2023 14:07
Operator : CG/JU
Sample : 02588-12
Misc :
ALS Vial : 4 Sample Multiplier: 50


Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/09/2023
Supervised By : Jagrut Upadhyay 05/09/2023


Quant Time: May 09 04:02:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 08 18:27:02 2023
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.897 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:152 Resp: 281038
Ion Ratio Lower Upper
152 100
150 156.9 123.8 185.8
115 60.2 52.9 79.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#4
2-Fluorophenol
Concen: 101.385 ng
RT: 5.456 min Scan# 315
Delta R.T. -0.007 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07


Tgt Ion:112 Resp: 1387069
Ion Ratio Lower Upper
112 100
64 64.0 51.5 77.3
63 34.4 27.8 41.6





#5
Phenol-d6
Concen: 116.532 ng
RT: 7.066 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion: 99 Resp: 200052
Ion Ratio Lower Upper
99 100
42 26.4 22.0 33.0
71 35.1 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#6
bis(2-Chloroethyl)ether
Concen: 0.070 ng
RT: 7.319 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07


Tgt Ion: 93 Resp: 1054
Ion Ratio Lower Upper
93 100
63 68.1 67.8 101.6
95 27.6 26.7 40.1





#7
Naphthalene-d8
Concen: 20.000 ng
RT: 10.707 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:	136	Resp:	91656
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.4	14.2
54	11.5	8.5	12.7
68	7.4	5.8	8.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#8
Nitrobenzene-d5
Concen: 65.687 ng
RT: 9.063 min Scan# 781
Delta R.T. -0.021 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07

Tgt Ion:	82	Resp:	990422
Ion	Ratio	Lower	Upper
82	100		
128	38.7	35.3	52.9
54	64.5	48.7	73.1





#9

Naphthalene

Concen: 1.003 ng

RT: 10.761 min Scan# 941

Delta R.T. -0.000 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07

Instrument :

BNA_N

ClientSampleId :





ZONE-D-6-9-05022023

Tgt Ion:	128	Resp:	47189
Ion Ratio	Lower	Upper	
128	100		
129	11.6	9.4	14.2
127	13.3	11.0	16.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#11

2-Methylnaphthalene-d10

Concen: 0.003 ng




RT: 12.272 min Scan# 1212


Delta R.T. -0.034 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07


Tgt Ion:	152	Resp:	85
Ion Ratio	Lower	Upper	
152	100		
151	140.0	16.4	24.6





#12
2-Methylnaphthalene
Concen: 0.422 ng
RT: 12.377 min Scan# 11
Delta R.T. -0.004 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:142 Resp: 1403
Ion Ratio Lower Upper
142 100
141 87.4 70.5 105.7
115 39.3 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#13
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.544 min Scan# 1550
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07


Tgt Ion:164 Resp: 586222
Ion Ratio Lower Upper
164 100
162 104.4 82.1 123.1
160 48.4 38.3 57.5





#14
2,4,6-Tribromophenol
Concen: 122.122 ng
RT: 16.045 min Scan# 1688
Delta R.T. -0.012 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:330 Resp: 731960
Ion Ratio Lower Upper
330 100
332 97.2 77.4 116.0
141 40.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#15
2-Fluorobiphenyl
Concen: 54.334 ng
RT: 13.176 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07


Tgt Ion:172 Resp: 2184594
Ion Ratio Lower Upper
172 100
171 33.5 27.9 41.9
170 21.9 19.0 28.4





#16
Acenaphthylene
Concen: 1.249 ng
RT: 14.267 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:152 Resp: 65560
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.2 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#17
Acenaphthene
Concen: 0.470 ng
RT: 14.609 min Scan# 1556
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07


Tgt Ion:154 Resp: 15334
Ion Ratio Lower Upper
154 100
153 118.2 91.3 136.9
152 58.7 44.7 67.1





#18
Fluorene
Concen: 0.982 ng
RT: 15.603 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:166 Resp: 4394
Ion Ratio Lower Upper
166 100
165 100.2 79.0 118.4
167 17.2 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#19
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.299 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07

Tgt Ion:188 Resp: 1353557
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.5 8.0 12.0





#25

Phenanthrene

Concen: 7.089 ng

RT: 17.336 min Scan# 1790

Delta R.T. -0.000 min

Lab File: BN025331.D


Acq: 08 May 2023 14:07

Instrument :

BNA_N

ClientSampleId :

ZONE-D-6-9-05022023



Tgt Ion:178 Resp: 530994

Ion Ratio Lower Upper

178 100

176 19.2 15.5 23.3




179 15.9 12.4 18.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#26

Anthracene

Concen: 2.340 ng

RT: 17.435 min Scan# 1798

Delta R.T. -0.000 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07




Tgt Ion:178 Resp: 164164


Ion Ratio Lower Upper

178 100

176 18.5 14.9 22.3


179 14.9 12.0 18.0





#27
Fluoranthene-d10
Concen: 0.002 ng
RT: 19.325 min Scan# 2157
Delta R.T. -0.008 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:212 Resp: 112
Ion Ratio Lower Upper
212 100
106 78.6 12.6 19.0
104 0.0 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#28
Fluoranthene
Concen: 9.157 ng
RT: 19.363 min Scan# 2166
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07

Tgt Ion:202 Resp: 857001
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.8
203 17.5 13.5 20.3





#29

Chrysene-d12

Concen: 20.000 ng

RT: 21.500 min Scan# 2486

Delta R.T. -0.000 min

Lab File: BN025331.D


Acq: 08 May 2023 14:07

Instrument :

BNA_N

ClientSampleId :

ZONE-D-6-9-05022023



Tgt Ion:240 Resp: 1309778

Ion Ratio Lower Upper

240 100

120 12.6 10.3 15.5




236 28.7 22.3 33.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#30

Pyrene

Concen: 7.846 ng

RT: 19.726 min Scan# 2252

Delta R.T. -0.004 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07




Tgt Ion:202 Resp: 702652


Ion Ratio Lower Upper

202 100

200 21.0 16.8 25.2


203 20.3 14.4 21.6





#31
Terphenyl-d14
Concen: 53.883 ng
RT: 19.928 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:244 Resp: 284597
Ion Ratio Lower Upper
244 100
212 9.3 7.1 10.7
122 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#32
Benzo(a)anthracene
Concen: 4.839 ng
RT: 21.482 min Scan# 2483
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07

Tgt Ion:228 Resp: 428445
Ion Ratio Lower Upper
228 100
226 29.5 22.5 33.7
229 21.9 16.2 24.2





#33

Chrysene

Concen: 4.027 ng

RT: 21.536 min Scan# 2490

Delta R.T. -0.000 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07

Instrument :

BNA_N

ClientSampleId :





ZONE-D-6-9-05022023

Tgt Ion:228	Resp: 35490
Ion Ratio	Lower Upper
228 100	
226 31.5	24.6 36.8
229 23.3	16.2 24.4

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.770 ng




RT: 21.392 min Scan# 2473


Delta R.T. -0.000 min

Lab File: BN025331.D

Acq: 08 May 2023 14:07


Tgt Ion:149	Resp: 53499
Ion Ratio	Lower Upper
149 100	
167 28.0	22.2 33.4
279 3.1	2.6 4.0





#35
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.928 min Scan# 3107
 Delta R.T. -0.003 min
 Lab File: BN025331.D
 Acq: 08 May 2023 14:07




Instrument :
 BNA_N
 ClientSampleId :
 ZONE-D-6-9-05022023



Tgt Ion:264 Resp: 1661960
 Ion Ratio Lower Upper
 264 100
 260 26.0 21.5 32.3
 265 22.0 46.2 69.4




Manual Integrations
 APPROVED


Reviewed By :Christian Giraldo 05/09/2023
 Supervised By :Jagrut Upadhyay 05/09/2023



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.870 ng
 RT: 26.442 min Scan# 3967
 Delta R.T. -0.006 min
 Lab File: BN025331.D
 Acq: 08 May 2023 14:07


Tgt Ion:276 Resp: 249247
 Ion Ratio Lower Upper
 276 100
 138 23.6 22.5 33.7
 277 24.8 19.8 29.6





#37
Benzo(b)fluoranthene
Concen: 4.729 ng
RT: 23.188 min Scan# 2854
Delta R.T. -0.000 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07




Instrument :
BNA_N
ClientSampleId :
ZONE-D-6-9-05022023



Tgt Ion:252 Resp: 49827
Ion Ratio Lower Upper
252 100
253 23.2 20.8 31.2
125 14.8 13.4 20.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#38
Benzo(k)fluoranthene
Concen: 1.608 ng m
RT: 23.232 min Scan# 2869
Delta R.T. -0.006 min
Lab File: BN025331.D
Acq: 08 May 2023 14:07

Tgt Ion:252 Resp: 177383
Ion Ratio Lower Upper
252 100
253 25.1 20.6 30.8
125 17.9 13.1 19.7





#39

Benzo(a)pyrene

Concen: 3.620 ng m

RT: 23.817 min Scan# 3073

Delta R.T. -0.006 min

Lab File: BN025331.D


Acq: 08 May 2023 14:07

Instrument :

BNA_N

ClientSampleId :

ZONE-D-6-9-05022023



Tgt Ion:252 Resp: 390790



Ion	Ratio	Lower	Upper
252	100		
253	23.9	21.4	32.0
125	15.3	15.4	23.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#40

Dibenzo(a,h)anthracene

Concen: 0.594 ng

RT: 26.451 min Scan# 3970




Delta R.T. -0.015 min


Lab File: BN025331.D

Acq: 08 May 2023 14:07

Tgt Ion:278 Resp: 62500


Ion	Ratio	Lower	Upper
278	100		
139	29.9	17.6	26.4#
279	33.8	21.6	32.4#





#41
 Benzo(g,h,i)perylene
 Concen: 2.021 ng
 RT: 27.220 min Scan# 41
 Delta R.T. -0.009 min
 Lab File: BN025331.D
 Acq: 08 May 2023 14:07

Instrument :
 BNA_N
 ClientSampleId :
 ZONE-D-6-9-05022023



Tgt Ion: 276 Resp: 233518

Ion	Ratio	Lower	Upper
276	100		
277	24.5	20.1	30.1
138	25.3	21.0	31.6

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/09/2023
 Supervised By : Jagrut Upadhyay 05/09/2023

