




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050922\
Data File : BN019704.D
Acq On : 09 May 2022 15:32
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 10 04:03:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019704.D\data.ms

(70) Pentachlorophenol (C)




16.869min (-0.003) 20903.86 ng

response 370749

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.13
264.00	61.60	62.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050922\
Data File : BN019704.D
Acq On : 09 May 2022 15:32
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 10 04:03:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019704.D\data.ms

(77) Benzidine

19.463min (-0.003) 0.00 ng

response 2271110

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.42
183.00	13.20	12.36
0.00	0.00	0.00