

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050922\
 Data File : BN019712.D
 Acq On : 09 May 2022 20:52
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050922

Quant Time: May 10 01:37:17 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 10 01:35:26 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.510	0.497	2.5	100	0.00
3	n-Nitrosodimethylamine	0.524	0.475	9.4	102	0.00
4 S	2-Fluorophenol	0.905	0.913	-0.9	103	0.00
5 S	Phenol-d6	1.139	1.116	2.0	104	0.00
6	bis(2-Chloroethyl)ether	1.226	1.195	2.5	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.334	0.328	1.8	102	0.00
9	Naphthalene	1.221	1.194	2.2	103	0.00
10	Hexachlorobutadiene	0.228	0.225	1.3	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.685	0.664	3.1	105	0.00
12	2-Methylnaphthalene	0.819	0.793	3.2	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.145	0.122	15.9	100	0.00
15 S	2-Fluorobiphenyl	1.805	1.782	1.3	103	0.00
16	Acenaphthylene	2.013	1.875	6.9	107	0.00
17	Acenaphthene	1.405	1.329	5.4	105	0.00
18	Fluorene	1.784	1.672	6.3	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.053	3.6	108	0.00
21	4-Bromophenyl-phenylether	0.262	0.245	6.5	103	0.00
22	Hexachlorobenzene	0.298	0.291	2.3	107	0.00
23	Atrazine	0.161	0.139	13.7	104	0.00
24	Pentachlorophenol	0.090	0.076	15.6	98	0.00
25	Phenanthrene	1.417	1.357	4.2	105	0.00
26	Anthracene	1.171	1.080	7.8	105	0.00
27 SURR	Fluoranthene-d10	1.154	1.096	5.0	107	0.00
28	Fluoranthene	1.512	1.430	5.4	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.701	1.645	3.3	106	0.00
31 S	Terphenyl-d14	0.930	0.926	0.4	105	0.00
32	Benzo(a)anthracene	1.447	1.330	8.1	107	0.00
33	Chrysene	1.688	1.645	2.5	106	0.00
34	Bis(2-ethylhexyl)phthalate	0.651	0.574	11.8	105	0.00
35 I	Perylene-d12	1.000	1.000	0.0	106	0.00
36	Indeno(1,2,3-cd)pyrene	2.025	1.906	5.9	104	0.00
37	Benzo(b)fluoranthene	1.796	1.643	8.5	105	0.00
38	Benzo(k)fluoranthene	1.867	1.763	5.6	103	0.00
39 C	Benzo(a)pyrene	1.389	1.274	8.3	111	0.00
40	Dibenzo(a,h)anthracene	1.657	1.559	5.9	103	0.00
41	Benzo(g,h,i)perylene	1.727	1.620	6.2	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN050922