

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050922\
 Data File : BN019719.D
 Acq On : 10 May 2022 02:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 02:47:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 10 01:35:26 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.510	0.466	8.6	87	0.00
3	n-Nitrosodimethylamine	0.524	0.484	7.6	97	0.00
4 S	2-Fluorophenol	0.905	0.789	12.8	83	0.00
5 S	Phenol-d6	1.139	0.968	15.0	85	0.00
6	bis(2-Chloroethyl)ether	1.226	1.158	5.5	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.334	0.313	6.3	90	0.00
9	Naphthalene	1.221	1.187	2.8	95	0.00
10	Hexachlorobutadiene	0.228	0.228	0.0	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.685	0.649	5.3	95	0.00
12	2-Methylnaphthalene	0.819	0.783	4.4	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.145	0.114	21.4	83	0.00
15 S	2-Fluorobiphenyl	1.805	1.834	-1.6	94	0.00
16	Acenaphthylene	2.013	1.845	8.3	93	0.00
17	Acenaphthene	1.405	1.336	4.9	93	0.00
18	Fluorene	1.784	1.662	6.8	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.050	9.1	89	0.00
21	4-Bromophenyl-phenylether	0.262	0.245	6.5	91	0.00
22	Hexachlorobenzene	0.298	0.297	0.3	96	0.00
23	Atrazine	0.161	0.126	21.7	83	0.00
24	Pentachlorophenol	0.090	0.069	23.3	78	0.00
25	Phenanthrene	1.417	1.373	3.1	94	0.00
26	Anthracene	1.171	1.053	10.1	90	0.00
27 SURR	Fluoranthene-d10	1.154	1.056	8.5	91	0.00
28	Fluoranthene	1.512	1.407	6.9	92	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
30	Pyrene	1.701	1.666	2.1	92	0.00
31 S	Terphenyl-d14	0.930	0.933	-0.3	90	0.00
32	Benzo(a)anthracene	1.447	1.303	10.0	90	0.00
33	Chrysene	1.688	1.675	0.8	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.651	0.606	6.9	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	2.025	1.965	3.0	93	0.00
37	Benzo(b)fluoranthene	1.796	1.655	7.9	92	0.00
38	Benzo(k)fluoranthene	1.867	1.855	0.6	94	0.00
39 C	Benzo(a)pyrene	1.389	1.233	11.2	93	0.00
40	Dibenzo(a,h)anthracene	1.657	1.616	2.5	93	0.00
41	Benzo(g,h,i)perylene	1.727	1.631	5.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4