

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050922\
 Data File : BN019719.D
 Acq On : 10 May 2022 02:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 02:47:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 10 01:35:26 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.365	8.8	87	0.00
3	n-Nitrosodimethylamine	0.400	0.370	7.5	97	0.00
4 S	2-Fluorophenol	0.400	0.349	12.8	83	0.00
5 S	Phenol-d6	0.400	0.340	15.0	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.375	6.3	90	0.00
9	Naphthalene	0.400	0.389	2.8	95	0.00
10	Hexachlorobutadiene	0.400	0.400	0.0	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	95	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.314	21.5	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	94	0.00
16	Acenaphthylene	0.400	0.367	8.3	93	0.00
17	Acenaphthene	0.400	0.380	5.0	93	0.00
18	Fluorene	0.400	0.373	6.8	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.358	10.5	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	91	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	96	0.00
23	Atrazine	0.400	0.314	21.5	83	0.00
24	Pentachlorophenol	0.400	0.306	23.5	78	0.00
25	Phenanthrene	0.400	0.388	3.0	94	0.00
26	Anthracene	0.400	0.360	10.0	90	0.00
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	91	0.00
28	Fluoranthene	0.400	0.372	7.0	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.392	2.0	92	0.00
31 S	Terphenyl-d14	0.400	0.401	-0.3	90	0.00
32	Benzo(a)anthracene	0.400	0.360	10.0	90	0.00
33	Chrysene	0.400	0.397	0.8	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.372	7.0	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.388	3.0	93	0.00
37	Benzo(b)fluoranthene	0.400	0.369	7.8	92	0.00
38	Benzo(k)fluoranthene	0.400	0.397	0.8	94	0.00
39 C	Benzo(a)pyrene	0.400	0.355	11.3	93	0.00
40	Dibenzo(a,h)anthracene	0.400	0.390	2.5	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.378	5.5	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4