

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050923\
 Data File : BN025361.D
 Acq On : 09 May 2023 11:27
 Operator : CG/JU
 Sample : PB152568BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB152568BL

Quant Time: May 10 03:57:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 09 10:18:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	5593	0.400	ng	0.00
7) Naphthalene-d8	10.718	136	16388	0.400	ng	# 0.01
13) Acenaphthene-d10	14.544	164	10380	0.400	ng	0.00
19) Phenanthrene-d10	17.311	188	20811	0.400	ng	0.01
29) Chrysene-d12	21.504	240	19961	0.400	ng	0.02
35) Perylene-d12	23.935	264	18695	0.400	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.456	112	5911	0.434	ng	0.00
5) Phenol-d6	7.088	99	5939	0.348	ng	0.01
8) Nitrobenzene-d5	9.116	82	3924	0.291	ng	0.04
11) 2-Methylnaphthalene-d10	12.313	152	9135	0.408	ng	0.01
14) 2,4,6-Tribromophenol	16.070	330	1680	0.317	ng	0.02
15) 2-Fluorobiphenyl	13.187	172	13000	0.365	ng	0.01
27) Fluoranthene-d10	19.334	212	23749	0.461	ng	0.00
31) Terphenyl-d14	19.924	244	19385	0.482	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050923\
Data File : BN025361.D
Acq On : 09 May 2023 11:27
Operator : CG/JU
Sample : PB152568BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB152568BL

Quant Time: May 10 03:57:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 09 10:18:02 2023
Response via : Initial Calibration

