

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051023\
 Data File : BN025378.D
 Acq On : 10 May 2023 13:21
 Operator : CG/JU
 Sample : 02619-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFZE1

Quant Time: May 11 01:11:16 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 12:11:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	8276	0.400	ng/ul	0.00
4) Naphthalene-d8	10.704	136	27363	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.547	164	16557	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.296	188	35273	0.400	ng/ul	0.00
17) Chrysene-d12	21.488	240	31533	0.400	ng/ul	# 0.00
23) Perylene-d12	23.908	264	32047	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	37449	4.370	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.304	152	12896	0.367	ng/ul	0.00
18) Fluoranthene-d10	19.326	212	38328	0.461	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051023\
Data File : BN025378.D
Acq On : 10 May 2023 13:21
Operator : CG/JU
Sample : 02619-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFZE1

Quant Time: May 11 01:11:16 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 10 12:11:16 2023
Response via : Initial Calibration

