

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051024\
 Data File : BN031490.D
 Acq On : 11 May 2024 09:24
 Operator : MA/JU
 Sample : P2428-01
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RMW-06B-38-050824

Quant Time: May 13 01:26:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 02:12:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.711	152	12003	0.400	ng	0.00
7) Naphthalene-d8	10.485	136	37904	0.400	ng	#-0.01
13) Acenaphthene-d10	14.338	164	20968	0.400	ng	-0.01
19) Phenanthrene-d10	17.078	188	47374	0.400	ng	-0.01
29) Chrysene-d12	21.265	240	36531	0.400	ng	0.00
35) Perylene-d12	23.501	264	32230	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.299	112	2906	0.085	ng	0.00
5) Phenol-d6	6.866	99	2514	0.055	ng	0.00
8) Nitrobenzene-d5	8.851	82	7611	0.280	ng	-0.01
11) 2-Methylnaphthalene-d10	12.081	152	13351	0.219	ng	-0.01
14) 2,4,6-Tribromophenol	15.824	330	1682	0.211	ng	-0.01
15) 2-Fluorobiphenyl	12.970	172	25827	0.319	ng	-0.01
27) Fluoranthene-d10	19.114	212	36200	0.275	ng	0.00
31) Terphenyl-d14	19.718	244	37558	0.420	ng	-0.01

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051024\
Data File : BN031490.D
Acq On : 11 May 2024 09:24
Operator : MA/JU
Sample : P2428-01
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RMW-06B-38-050824

Quant Time: May 13 01:26:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 02:12:24 2024
Response via : Initial Calibration

