

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051024\
 Data File : BN031456.D
 Acq On : 10 May 2024 08:47
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 10:44:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN050924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 02:12:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	0.00
2	1,4-Dioxane	0.400	0.386	3.5	122	0.00
3	n-Nitrosodimethylamine	0.400	0.401	-0.3	127	0.00
4 S	2-Fluorophenol	0.400	0.366	8.5	110	0.00
5 S	Phenol-d6	0.400	0.362	9.5	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	117	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	119	-0.01
9	Naphthalene	0.400	0.394	1.5	119	-0.01
10	Hexachlorobutadiene	0.400	0.396	1.0	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	115	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	116	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.400	0.332	17.0	101	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	119	-0.01
16	Acenaphthylene	0.400	0.377	5.8	111	-0.01
17	Acenaphthene	0.400	0.390	2.5	116	-0.01
18	Fluorene	0.400	0.390	2.5	116	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.363	9.3	122	-0.01
21	4-Bromophenyl-phenylether	0.400	0.394	1.5	116	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	122	0.00
23	Atrazine	0.400	0.343	14.2	102	-0.01
24	Pentachlorophenol	0.400	0.348	13.0	111	0.00
25	Phenanthrene	0.400	0.395	1.3	115	0.00
26	Anthracene	0.400	0.376	6.0	112	-0.01
27 SURR	Fluoranthene-d10	0.400	0.375	6.3	110	0.00
28	Fluoranthene	0.400	0.386	3.5	113	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	108	0.00
30	Pyrene	0.400	0.405	-1.3	113	0.00
31 S	Terphenyl-d14	0.400	0.379	5.3	107	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	104	0.00
33	Chrysene	0.400	0.392	2.0	110	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.279	30.3#	80	0.00
35 I	Perylene-d12	0.400	0.400	0.0	106	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	99	0.00
37	Benzo(b)fluoranthene	0.400	0.363	9.3	102	0.00
38	Benzo(k)fluoranthene	0.400	0.389	2.8	108	0.00
39 C	Benzo(a)pyrene	0.400	0.348	13.0	97	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	101	0.00
41	Benzo(g,h,i)perylene	0.400	0.367	8.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4