

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
 Data File : BN014564.D
 Acq On : 11 May 2021 16:23
 Operator : CG/JU
 Sample : SSTDIC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.2

Quant Time: May 11 17:46:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 17:45:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.741	152	663	0.40	ng	# 0.00
7) Naphthalene-d8	10.530	136	2387	0.40	ng	# 0.00
13) Acenaphthene-d10	14.384	164	1673	0.40	ng	0.00
19) Phenanthrene-d10	17.128	188	4048	0.40	ng	# 0.00
29) Chrysene-d12	21.335	240	5229	0.40	ng	0.00
35) Perylene-d12	23.606	264	5224	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.333	112	398	0.21	ng	0.00
5) Phenol-d6	6.903	99	514	0.20	ng	0.00
8) Nitrobenzene-d5	8.883	82	430	0.22	ng	0.00
11) 2-Methylnaphthalene-d10	12.120	152	1185	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	158	0.21	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	1286	0.21	ng	0.00
27) Fluoranthene-d10	19.173	212	3655	0.21	ng	0.00
31) Terphenyl-d14	19.779	244	2567	0.20	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.295	88	247	0.30	ng	# 85
3) n-Nitrosodimethylamine	3.730	42	7	0.02	ng	# 1
6) bis(2-Chloroethyl)ether	7.177	93	409	0.22	ng	# 86
9) Naphthalene	10.581	128	1303	0.21	ng	# 84
10) Hexachlorobutadiene	10.872	225	307	0.20	ng	# 96
12) 2-Methylnaphthalene	12.195	142	915	0.20	ng	# 88
16) Acenaphthylene	14.105	152	1193	0.18	ng	98
17) Acenaphthene	14.448	154	908	0.20	ng	98
18) Fluorene	15.437	166	1251	0.21	ng	99
20) 4,6-Dinitro-2-methylph...	15.513	198	84	0.25	ng	# 1
21) 4-Bromophenyl-phenylether	16.337	248	488	0.19	ng	93
22) Hexachlorobenzene	16.447	284	557	0.22	ng	90
23) Atrazine	16.605	200	315	0.18	ng	# 82
24) Pentachlorophenol	16.788	266	181	0.21	ng	# 23
25) Phenanthrene	17.177	178	2304	0.21	ng	98
26) Anthracene	17.262	178	1774	0.18	ng	98
28) Fluoranthene	19.203	202	2916	0.21	ng	97
30) Pyrene	19.566	202	3024	0.21	ng	99
32) Benzo(a)anthracene	21.314	228	3094	0.20	ng	97
33) Chrysene	21.367	228	3315	0.21	ng	99
34) Bis(2-ethylhexyl)phtha...	21.261	149	924	0.16	ng	# 75
36) Indeno(1,2,3-cd)pyrene	25.899	276	4122	0.21	ng	# 92
37) Benzo(b)fluoranthene	22.962	252	3679	0.21	ng	92
38) Benzo(k)fluoranthene	22.962	252	3679	0.22	ng	# 91
39) Benzo(a)pyrene	23.503	252	2959	0.20	ng	# 90
40) Dibenzo(a,h)anthracene	25.919	278	3481	0.20	ng	91
41) Benzo(g,h,i)perylene	26.601	276	3380	0.20	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
Data File : BN014564.D
Acq On : 11 May 2021 16:23
Operator : CG/JU
Sample : SSTDIC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Quant Time: May 11 17:46:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 11 17:45:36 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.741 min Scan# 616
 Delta R.T. -0.000 min
 Lab File: BN014564.D
 Acq: 11 May 2021 16:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion: 152 Resp: 663
 Ion Ratio Lower Upper
 152 100
 150 144.3 123.8 185.8
 115 84.7 47.0 70.4#



#2
 1,4-Dioxane
 Concen: 0.30 ng
 RT: 3.295 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN014564.D
 Acq: 11 May 2021 16:23

Tgt Ion: 88 Resp: 247
 Ion Ratio Lower Upper
 88 100
 58 44.9 44.4 66.6
 43 11.7 15.9 23.9#









#3
n-Nitrosodimethylamine
Concen: 0.02 ng
RT: 3.730 min Scan# 118
Delta R.T. -0.040 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 7
Ion Ratio Lower Upper
42 100
74 0.0 175.1 262.7#
44 171.4 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.21 ng
RT: 5.333 min Scan# 317
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion: 112 Resp: 398
Ion Ratio Lower Upper
112 100
64 58.0 41.8 62.6
63 28.9 22.0 33.0









#5
Phenol-d6
Concen: 0.20 ng
RT: 6.903 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 99 Resp: 514
Ion Ratio Lower Upper
99 100
42 12.6 13.2 19.8#
71 43.0 26.2 39.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.22 ng
RT: 7.177 min Scan# 546
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion: 93 Resp: 409
Ion Ratio Lower Upper
93 100
63 57.7 58.0 87.0#
95 31.1 27.0 40.6





#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.530 min Scan# 862

Delta R.T. 0.000 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23





Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2

Tgt Ion:	136	Resp:	2387
Ion	Ratio	Lower	Upper
136	100		
137	15.1	8.9	13.3#
54	8.6	3.8	5.8#
68	8.3	3.8	5.6#



#8

Nitrobenzene-d5

Concen: 0.22 ng




RT: 8.883 min Scan# 732


Delta R.T. 0.000 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23

Tgt Ion:	82	Resp:	430
Ion	Ratio	Lower	Upper
82	100		
128	36.6	29.5	44.3
54	52.3	36.9	55.3









#9
Naphthalene
Concen: 0.21 ng
RT: 10.581 min Scan# 866
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 1303
Ion Ratio Lower Upper
128 100
129 16.6 9.0 13.4#
127 20.0 10.4 15.6#



#10
Hexachlorobutadiene
Concen: 0.20 ng
RT: 10.872 min Scan# 889
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:225 Resp: 307
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 66.8 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.21 ng
RT: 12.120 min Scan# 1077
Delta R.T. -0.004 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:152 Resp: 1185
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.20 ng
RT: 12.195 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:142 Resp: 915
Ion Ratio Lower Upper
142 100
141 91.2 70.2 105.4
115 52.9 26.7 40.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:164 Resp: 1673
Ion Ratio Lower Upper
164 100
162 103.9 78.6 118.0
160 49.1 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.21 ng
RT: 15.881 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:330 Resp: 158
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.1 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.21 ng
RT: 13.001 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:172 Resp: 1286
Ion Ratio Lower Upper
172 100
171 41.2 27.1 40.7#
170 29.0 18.3 27.5#



#16
Acenaphthylene
Concen: 0.18 ng
RT: 14.105 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:152 Resp: 1193
Ion Ratio Lower Upper
152 100
151 20.0 15.6 23.4
153 13.9 10.5 15.7









#17
Acenaphthene
Concen: 0.20 ng
RT: 14.448 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:154 Resp: 908
Ion Ratio Lower Upper
154 100
153 115.1 90.2 135.4
152 58.5 46.2 69.2



#18
Fluorene
Concen: 0.21 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:166 Resp: 1251
Ion Ratio Lower Upper
166 100
165 99.0 78.4 117.6
167 14.1 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:188 Resp: 4048
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.0 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.25 ng
RT: 15.513 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:198 Resp: 84
Ion Ratio Lower Upper
198 100
51 109.2 865.3 1297.9#
105 89.7 93.6 140.4#









#21
4-Bromophenyl-phenylether
Concen: 0.19 ng
RT: 16.337 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampled :
SSTDIC0.2


Tgt Ion:248 Resp: 488
Ion Ratio Lower Upper
248 100
250 108.4 80.5 120.7
141 49.6 37.2 55.8



#22
Hexachlorobenzene
Concen: 0.22 ng
RT: 16.447 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:284 Resp: 557
Ion Ratio Lower Upper
284 100
142 27.3 26.6 39.8
249 34.6 23.6 35.4









#23
Atrazine
Concen: 0.18 ng
RT: 16.605 min Scan# 1489
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:200 Resp: 315
Ion Ratio Lower Upper
200 100
173 39.1 23.0 34.6#
215 59.8 38.6 57.8#



#24
Pentachlorophenol
Concen: 0.21 ng
RT: 16.788 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:266 Resp: 181
Ion Ratio Lower Upper
266 100
264 63.5 24.2 36.4#
268 82.9 25.2 37.8#









#25
Phenanthrene
Concen: 0.21 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:178 Resp: 2304
Ion Ratio Lower Upper
178 100
176 19.4 15.2 22.8
179 16.8 12.4 18.6



#26
Anthracene
Concen: 0.18 ng
RT: 17.262 min Scan# 1543
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:178 Resp: 1774
Ion Ratio Lower Upper
178 100
176 19.8 14.8 22.2
179 15.6 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.21 ng
RT: 19.173 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 3655
Ion Ratio Lower Upper
212 100
106 10.7 10.1 15.1
104 6.1 5.4 8.2



#28
Fluoranthene
Concen: 0.21 ng
RT: 19.203 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:202 Resp: 2916
Ion Ratio Lower Upper
202 100
101 9.4 8.4 12.6
203 18.2 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.335 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 5229
Ion Ratio Lower Upper
240 100
120 7.1 5.8 8.8
236 27.2 20.6 30.8



#30
Pyrene
Concen: 0.21 ng
RT: 19.566 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23


Tgt Ion:202 Resp: 3024
Ion Ratio Lower Upper
202 100
200 20.7 16.3 24.5
203 17.8 14.0 21.0








#31
 Terphenyl-d14
 Concen: 0.20 ng
 RT: 19.779 min Scan# 1960
 Delta R.T. 0.005 min
 Lab File: BN014564.D
 Acq: 11 May 2021 16:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2







Tgt Ion:244 Resp: 2567
 Ion Ratio Lower Upper
 244 100
 212 9.6 6.0 9.0#
 122 9.7 9.3 13.9



#32
 Benzo(a)anthracene
 Concen: 0.20 ng
 RT: 21.314 min Scan# 2129
 Delta R.T. 0.000 min
 Lab File: BN014564.D
 Acq: 11 May 2021 16:23

Tgt Ion:228 Resp: 3094
 Ion Ratio Lower Upper
 228 100
 226 28.7 21.7 32.5
 229 21.0 15.7 23.5





#33

Chrysene

Concen: 0.21 ng

RT: 21.367 min Scan# 2134

Delta R.T. 0.000 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:228 Resp: 3315

Ion Ratio Lower Upper

228 100

226 29.4 23.8 35.6

229 20.1 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.16 ng

RT: 21.261 min Scan# 2124

Delta R.T. 0.000 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23




Tgt Ion:149 Resp: 924


Ion Ratio Lower Upper

149 100

167 39.3 20.5 30.7#

279 6.9 2.9 4.3#









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.606 min Scan# 2582
Delta R.T. 0.003 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:264 Resp: 5224
Ion Ratio Lower Upper
264 100
260 25.8 19.5 29.3
265 63.5 35.4 53.0#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.21 ng
RT: 25.899 min Scan# 3252
Delta R.T. -0.003 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:276 Resp: 4122
Ion Ratio Lower Upper
276 100
138 16.2 18.8 28.2#
277 25.8 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.21 ng
RT: 22.962 min Scan# 2394
Delta R.T. 0.044 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 3679
Ion Ratio Lower Upper
252 100
253 29.3 19.6 29.4
125 15.0 10.3 15.5



#38
Benzo(k)fluoranthene
Concen: 0.22 ng
RT: 22.962 min Scan# 2394
Delta R.T. 0.000 min
Lab File: BN014564.D
Acq: 11 May 2021 16:23

Tgt Ion:252 Resp: 3679
Ion Ratio Lower Upper
252 100
253 29.3 19.7 29.5
125 15.0 9.7 14.5#





#39

Benzo(a)pyrene

Concen: 0.20 ng

RT: 23.503 min Scan# 2552

Delta R.T. 0.000 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23





Tgt Ion:252	Resp:	2959
Ion Ratio	Lower	Upper
252	100	
253	33.3	21.8 32.6#
125	15.7	14.9 22.3

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2



#40

Dibenzo(a,h)anthracene

Concen: 0.20 ng




RT: 25.919 min Scan# 3258


Delta R.T. 0.003 min

Lab File: BN014564.D

Acq: 11 May 2021 16:23

Tgt Ion:278	Resp:	3481
Ion Ratio	Lower	Upper
278	100	
139	14.0	13.9 20.9
279	30.9	20.7 31.1





#41
 Benzo(g,h,i)perylene
 Concen: 0.20 ng
 RT: 26.601 min Scan# 3457
 Delta R.T. 0.000 min
 Lab File: BN014564.D
 Acq: 11 May 2021 16:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	276	Resp:	3380
Ion	Ratio	Lower	Upper
276	100		
277	28.3	19.6	29.4
138	16.4	17.5	26.3#

