

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
 Data File : BN014565.D
 Acq On : 11 May 2021 16:58
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: May 11 17:47:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 17:45:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.741	152	749	0.40	ng	# 0.00
7) Naphthalene-d8	10.530	136	2786	0.40	ng	# 0.00
13) Acenaphthene-d10	14.384	164	2022	0.40	ng	0.00
19) Phenanthrene-d10	17.128	188	4980	0.40	ng	# 0.00
29) Chrysene-d12	21.335	240	6459	0.40	ng	0.00
35) Perylene-d12	23.602	264	6341	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.333	112	732	0.35	ng	0.00
5) Phenol-d6	6.903	99	976	0.33	ng	0.00
8) Nitrobenzene-d5	8.883	82	856	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.124	152	2483	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	300	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	2723	0.37	ng	0.00
27) Fluoranthene-d10	19.173	212	7736	0.36	ng	0.00
31) Terphenyl-d14	19.774	244	5730	0.36	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.295	88	393	0.42	ng	99
3) n-Nitrosodimethylamine	3.770	42	14	0.04	ng	# 1
6) bis(2-Chloroethyl)ether	7.177	93	805	0.38	ng	94
9) Naphthalene	10.581	128	2721	0.37	ng	# 93
10) Hexachlorobutadiene	10.872	225	660	0.37	ng	# 99
12) 2-Methylnaphthalene	12.195	142	1962	0.37	ng	# 92
16) Acenaphthylene	14.105	152	2573	0.33	ng	98
17) Acenaphthene	14.448	154	1994	0.37	ng	98
18) Fluorene	15.437	166	2690	0.37	ng	98
20) 4,6-Dinitro-2-methylph...	15.513	198	183	0.45	ng	# 1
21) 4-Bromophenyl-phenylether	16.337	248	1091	0.35	ng	97
22) Hexachlorobenzene	16.447	284	1142	0.37	ng	92
23) Atrazine	16.605	200	642	0.30	ng	91
24) Pentachlorophenol	16.788	266	337	0.32	ng	# 40
25) Phenanthrene	17.177	178	4797	0.36	ng	100
26) Anthracene	17.262	178	3963	0.33	ng	98
28) Fluoranthene	19.203	202	6281	0.37	ng	98
30) Pyrene	19.566	202	6339	0.35	ng	99
32) Benzo(a)anthracene	21.314	228	6639	0.34	ng	98
33) Chrysene	21.367	228	7064	0.36	ng	98
34) Bis(2-ethylhexyl)phtha...	21.261	149	1771	0.25	ng	# 79
36) Indeno(1,2,3-cd)pyrene	25.902	276	9081	0.38	ng	# 92
37) Benzo(b)fluoranthene	22.962	252	8049	0.38	ng	97
38) Benzo(k)fluoranthene	22.962	252	8049	0.40	ng	98
39) Benzo(a)pyrene	23.503	252	6357	0.35	ng	# 93
40) Dibenzo(a,h)anthracene	25.916	278	7800	0.38	ng	# 94
41) Benzo(g,h,i)perylene	26.601	276	7272	0.36	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
Data File : BN014565.D
Acq On : 11 May 2021 16:58
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: May 11 17:47:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 11 17:45:36 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.741 min Scan# 616
 Delta R.T. -0.000 min
 Lab File: BN014565.D
 Acq: 11 May 2021 16:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:152 Resp: 749
 Ion Ratio Lower Upper
 152 100
 150 154.7 123.8 185.8
 115 87.4 47.0 70.4#



#2
 1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.295 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN014565.D
 Acq: 11 May 2021 16:58

Tgt Ion: 88 Resp: 393
 Ion Ratio Lower Upper
 88 100
 58 54.5 44.4 66.6
 43 20.6 15.9 23.9









#3
n-Nitrosodimethylamine
Concen: 0.04 ng
RT: 3.770 min Scan# 123
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 14
Ion Ratio Lower Upper
42 100
74 0.0 175.1 262.7#
44 171.4 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.35 ng
RT: 5.333 min Scan# 317
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion: 112 Resp: 732
Ion Ratio Lower Upper
112 100
64 58.5 41.8 62.6
63 29.4 22.0 33.0









#5
Phenol-d6
Concen: 0.33 ng
RT: 6.903 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampled :
SSTDICCC0.4


Tgt Ion:	99	Resp:	976
Ion	Ratio	Lower	Upper
99	100		
42	14.5	13.2	19.8
71	44.8	26.2	39.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.177 min Scan# 546
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:	93	Resp:	805
Ion	Ratio	Lower	Upper
93	100		
63	65.8	58.0	87.0
95	34.5	27.0	40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.530 min Scan# 862
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	2786
Ion	Ratio	Lower	Upper
136	100		
137	14.1	8.9	13.3#
54	7.8	3.8	5.8#
68	6.7	3.8	5.6#



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.883 min Scan# 732
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:	82	Resp:	856
Ion	Ratio	Lower	Upper
82	100		
128	32.5	29.5	44.3
54	47.8	36.9	55.3









#9
Naphthalene
Concen: 0.37 ng
RT: 10.581 min Scan# 866
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 2721
Ion Ratio Lower Upper
128 100
129 13.7 9.0 13.4#
127 15.6 10.4 15.6#



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.872 min Scan# 889
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:225 Resp: 660
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2483
Ion Ratio Lower Upper
152 100
151 21.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.195 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:142 Resp: 1962
Ion Ratio Lower Upper
142 100
141 88.6 70.2 105.4
115 48.8 26.7 40.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 2022
Ion Ratio Lower Upper
164 100
162 102.1 78.6 118.0
160 47.5 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.881 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:330 Resp: 300
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 32.0 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 13.001 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	2723
Ion	Ratio	Lower	Upper
172	100		
171	37.7	27.1	40.7
170	26.1	18.3	27.5



#16
Acenaphthylene
Concen: 0.33 ng
RT: 14.105 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:	152	Resp:	2573
Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.6	23.4
153	13.6	10.5	15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.448 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 1994
Ion Ratio Lower Upper
154 100
153 110.7 90.2 135.4
152 55.4 46.2 69.2



#18
Fluorene
Concen: 0.37 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:166 Resp: 2690
Ion Ratio Lower Upper
166 100
165 99.6 78.4 117.6
167 14.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 4980
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.7 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.45 ng
RT: 15.513 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:198 Resp: 183
Ion Ratio Lower Upper
198 100
51 68.7 865.3 1297.9#
105 62.0 93.6 140.4#









#21
4-Bromophenyl-phenylether
Concen: 0.35 ng
RT: 16.337 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampled :
SSTDICCC0.4


Tgt Ion:248 Resp: 1091
Ion Ratio Lower Upper
248 100
250 101.7 80.5 120.7
141 41.9 37.2 55.8



#22
Hexachlorobenzene
Concen: 0.37 ng
RT: 16.447 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:284 Resp: 1142
Ion Ratio Lower Upper
284 100
142 28.4 26.6 39.8
249 33.7 23.6 35.4





#23

Atrazine

Concen: 0.30 ng

RT: 16.605 min Scan# 1489

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58





Tgt Ion:	200	Resp:	642
Ion Ratio	Lower	Upper	
200	100		
173	31.4	23.0	34.6
215	55.6	38.6	57.8

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#24

Pentachlorophenol

Concen: 0.32 ng




RT: 16.788 min Scan# 1504


Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58

Tgt Ion:	266	Resp:	337
Ion Ratio	Lower	Upper	
266	100		
264	64.7	24.2	36.4#
268	63.5	25.2	37.8#









#25
Phenanthrene
Concen: 0.36 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 4797
Ion Ratio Lower Upper
178 100
176 19.1 15.2 22.8
179 15.8 12.4 18.6



#26
Anthracene
Concen: 0.33 ng
RT: 17.262 min Scan# 1543
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:178 Resp: 3963
Ion Ratio Lower Upper
178 100
176 18.1 14.8 22.2
179 16.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.173 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 7736
Ion Ratio Lower Upper
212 100
106 10.1 10.1 15.1#
104 5.9 5.4 8.2



#28
Fluoranthene
Concen: 0.37 ng
RT: 19.203 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:202 Resp: 6281
Ion Ratio Lower Upper
202 100
101 9.6 8.4 12.6
203 17.3 13.5 20.3





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.335 min Scan# 2131

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58

Tgt Ion:240 Resp: 6459

Ion Ratio Lower Upper

240 100

120 5.9 5.8 8.8





236 26.4 20.6 30.8

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#30

Pyrene

Concen: 0.35 ng

RT: 19.566 min Scan# 1917

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58




Tgt Ion:202 Resp: 6339


Ion Ratio Lower Upper

202 100

200 20.8 16.3 24.5

203 17.6 14.0 21.0









#31
Terphenyl-d14
Concen: 0.36 ng
RT: 19.774 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:244 Resp: 5730
Ion Ratio Lower Upper
244 100
212 8.6 6.0 9.0
122 9.5 9.3 13.9



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.314 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:228 Resp: 6639
Ion Ratio Lower Upper
228 100
226 27.9 21.7 32.5
229 20.5 15.7 23.5





#33

Chrysene

Concen: 0.36 ng

RT: 21.367 min Scan# 2134

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58

Tgt Ion:228 Resp: 7064

Ion Ratio Lower Upper

228 100

226 30.4 23.8 35.6





229 20.3 15.6 23.4

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.25 ng

RT: 21.261 min Scan# 2124

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58




Tgt Ion:149 Resp: 1771


Ion Ratio Lower Upper

149 100

167 37.0 20.5 30.7#

279 7.2 2.9 4.3#





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.602 min Scan# 2581

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58





Tgt Ion:264 Resp: 6341

Ion Ratio Lower Upper

264 100

260 26.6 19.5 29.3

265 55.9 35.4 53.0#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.38 ng

RT: 25.902 min Scan# 3253

Delta R.T. 0.000 min

Lab File: BN014565.D

Acq: 11 May 2021 16:58




Tgt Ion:276 Resp: 9081


Ion Ratio Lower Upper

276 100

138 16.8 18.8 28.2#

277 25.8 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 22.962 min Scan# 2394
Delta R.T. 0.044 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252	Resp:	8049
Ion Ratio	Lower	Upper
252	100	
253	25.7	19.6 29.4
125	11.1	10.3 15.5



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.962 min Scan# 2394
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:252	Resp:	8049
Ion Ratio	Lower	Upper
252	100	
253	25.7	19.7 29.5
125	11.1	9.7 14.5









#39
Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.503 min Scan# 2552
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 6357
Ion Ratio Lower Upper
252 100
253 28.0 21.8 32.6
125 11.9 14.9 22.3#



#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.916 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BN014565.D
Acq: 11 May 2021 16:58

Tgt Ion:278 Resp: 7800
Ion Ratio Lower Upper
278 100
139 11.8 13.9 20.9#
279 26.5 20.7 31.1





#41
 Benzo(g,h,i)perylene
 Concen: 0.36 ng
 RT: 26.601 min Scan# 3457
 Delta R.T. 0.000 min
 Lab File: BN014565.D
 Acq: 11 May 2021 16:58

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	7272
Ion	Ratio	Lower	Upper
276	100		
277	26.5	19.6	29.4
138	13.6	17.5	26.3#

