

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
 Data File : BN014570.D
 Acq On : 11 May 2021 19:55
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: May 12 09:54:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 12 09:51:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.741	152	768	0.40	ng	# 0.00
7) Naphthalene-d8	10.530	136	2978	0.40	ng	# 0.00
13) Acenaphthene-d10	14.384	164	2157	0.40	ng	0.00
19) Phenanthrene-d10	17.128	188	5273	0.40	ng	# 0.00
29) Chrysene-d12	21.324	240	6853	0.40	ng	#-0.01
35) Perylene-d12	23.602	264	6731	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.333	112	748	0.37	ng	0.00
5) Phenol-d6	6.903	99	1041	0.38	ng	0.00
8) Nitrobenzene-d5	8.883	82	893	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.120	152	2616	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	312	0.31	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	2848	0.36	ng	0.00
27) Fluoranthene-d10	19.168	212	8183	0.35	ng	0.00
31) Terphenyl-d14	19.774	244	6038	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.295	88	392	0.30	ng	95
3) n-Nitrosodimethylamine	3.746	42	35	0.09	ng	# 1
6) bis(2-Chloroethyl)ether	7.169	93	863	0.38	ng	90
9) Naphthalene	10.568	128	2829	0.36	ng	# 92
10) Hexachlorobutadiene	10.872	225	680	0.36	ng	# 99
12) 2-Methylnaphthalene	12.195	142	2056	0.36	ng	# 92
16) Acenaphthylene	14.105	152	2679	0.34	ng	98
17) Acenaphthene	14.448	154	2109	0.36	ng	99
18) Fluorene	15.437	166	2763	0.35	ng	97
20) 4,6-Dinitro-2-methylph...	15.513	198	185	0.36	ng	# 1
21) 4-Bromophenyl-phenylether	16.325	248	1121	0.35	ng	# 72
22) Hexachlorobenzene	16.447	284	1242	0.36	ng	91
23) Atrazine	16.605	200	627	0.31	ng	# 90
24) Pentachlorophenol	16.788	266	320	0.28	ng	# 40
25) Phenanthrene	17.177	178	5090	0.35	ng	99
26) Anthracene	17.262	178	4042	0.33	ng	99
28) Fluoranthene	19.203	202	6631	0.35	ng	98
30) Pyrene	19.566	202	6576	0.35	ng	99
32) Benzo(a)anthracene	21.314	228	6825	0.34	ng	99
33) Chrysene	21.367	228	7700	0.37	ng	98
34) Bis(2-ethylhexyl)phtha...	21.261	149	1770	0.32	ng	# 81
36) Indeno(1,2,3-cd)pyrene	25.899	276	9752	0.36	ng	# 92
37) Benzo(b)fluoranthene	22.962	252	8433	0.37	ng	97
38) Benzo(k)fluoranthene	22.962	252	8433	0.35	ng	97
39) Benzo(a)pyrene	23.503	252	6763	0.35	ng	# 94
40) Dibenzo(a,h)anthracene	25.913	278	8327	0.36	ng	# 95
41) Benzo(g,h,i)perylene	26.597	276	7752	0.35	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051121\
Data File : BN014570.D
Acq On : 11 May 2021 19:55
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: May 12 09:54:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 12 09:51:41 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.741 min Scan# 616
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 768
 Ion Ratio Lower Upper
 152 100
 150 153.7 123.8 185.8
 115 86.6 47.0 70.4#



1,4-Dioxane
 Concen: 0.30 ng
 RT: 3.295 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Tgt Ion: 88 Resp: 392
 Ion Ratio Lower Upper
 88 100
 58 52.8 44.4 66.6
 43 16.6 15.9 23.9









n-Nitrosodimethylamine
 Concen: 0.09 ng
 RT: 3.746 min Scan# 120
 Delta R.T. -0.024 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion: 42 Resp: 35




Ion	Ratio	Lower	Upper
42	100		
74	22.9	175.1	262.7#
44	31.4	7.5	11.3#




2-Fluorophenol
 Concen: 0.37 ng
 RT: 5.333 min Scan# 317
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Tgt Ion: 112 Resp: 748

Ion	Ratio	Lower	Upper
112	100		
64	58.2	41.8	62.6
63	31.0	22.0	33.0









Phenol-d6
 Concen: 0.38 ng
 RT: 6.903 min Scan# 512
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	99	Resp:	1041
Ion	Ratio	Lower	Upper
99	100		
42	12.7	13.2	19.8#
71	43.3	26.2	39.2#




bis(2-Chloroethyl)ether
 Concen: 0.38 ng
 RT: 7.169 min Scan# 545
 Delta R.T. -0.008 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Tgt Ion:	93	Resp:	863
Ion	Ratio	Lower	Upper
93	100		
63	61.0	58.0	87.0
95	33.3	27.0	40.6






Abundance Scan 862 (10.530 min): BN014565.D\data.ms (-8567-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.530 min Scan# 862
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	136	Resp:	2978
Ion	Ratio	Lower	Upper
136	100		
137	13.1	8.9	13.3
54	7.1	3.8	5.8#
68	7.0	3.8	5.6#






Abundance Scan 732 (8.883 min): BN014565.D\data.ms (-72978)




Nitrobenzene-d5
Concen: 0.35 ng
RT: 8.883 min Scan# 732
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	82	Resp:	893
Ion	Ratio	Lower	Upper
82	100		
128	33.5	29.5	44.3
54	46.2	36.9	55.3






Abundance Scan 866 (10.581 min): BN014565.D\data.ms (-86#9-)




Naphthalene
Concen: 0.36 ng
RT: 10.568 min Scan# 865
Delta R.T. -0.013 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	128	Resp:	2829
Ion	Ratio	Lower	Upper
128	100		
129	13.9	9.0	13.4#
127	16.6	10.4	15.6#






Abundance Scan 889 (10.872 min): BN014565.D\data.ms (-88#10-)



Hexachlorobutadiene
Concen: 0.36 ng
RT: 10.872 min Scan# 889
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55


Tgt Ion:	225	Resp:	680
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.6	50.9	76.3






Abundance Scan 1078 (12.124 min): BN014565.D\data.ms (-1014) (-)

2-Methylnaphthalene-d10
Concen: 0.36 ng
RT: 12.120 min Scan# 1077
Delta R.T. -0.005 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:152 Resp: 2616
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2







Abundance Scan 1094 (12.195 min): BN014565.D\data.ms (-1015) (-)

2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.195 min Scan# 1094
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55




Tgt Ion:142 Resp: 2056
Ion Ratio Lower Upper
142 100
141 87.7 70.2 105.4
115 49.0 26.7 40.1#








#106 (-)
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0





Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.9	78.6	118.0
160	45.8	35.8	53.8




#144 (-)
2,4,6-Tribromophenol
Concen: 0.31 ng
RT: 15.881 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55



Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	29.2	27.0	40.4









Scan 1203 (13.001 min): BN014565.D\data.ms (-1195) (-)
 2-Fluorobiphenyl
 Concen: 0.36 ng
 RT: 13.001 min Scan# 1203
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	172	Resp:	2848
Ion	Ratio	Lower	Upper
172	100		
171	38.1	27.1	40.7
170	26.2	18.3	27.5



Scan 1290 (14.105 min): BN014565.D\data.ms (-1210) (-)
 Acenaphthylene
 Concen: 0.34 ng
 RT: 14.105 min Scan# 1290
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Tgt Ion:	152	Resp:	2679
Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.6	23.4
153	14.1	10.5	15.7









Acenaphthene
Concen: 0.36 ng
RT: 14.448 min Scan# 1317
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	154	Resp:	2109
Ion	Ratio	Lower	Upper
154	100		
153	112.6	90.2	135.4
152	55.4	46.2	69.2




Fluorene
Concen: 0.35 ng
RT: 15.437 min Scan# 1395
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	166	Resp:	2763
Ion	Ratio	Lower	Upper
166	100		
165	100.7	78.4	117.6
167	14.0	10.7	16.1






Abundance Scan 1532 (17.128 min): BN014565.D\data.ms (-1516) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.128 min Scan# 1532
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	188	Resp:	5273
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	13.0	5.2	7.8#






Abundance Scan 1401 (15.513 min): BN014565.D\data.ms (-1326) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.36 ng
RT: 15.513 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	198	Resp:	185
Ion Ratio	Lower	Upper	
198	100		
51	66.7	865.3	1297.9#
105	59.5	93.6	140.4#





Abundance Scan 1467 (16.337 min): BN014565.D\data.ms (-1467) (-)




4-Bromophenyl-phenylether
Concen: 0.35 ng
RT: 16.325 min Scan# 1466
Delta R.T. -0.012 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	248	Resp:	1121
Ion	Ratio	Lower	Upper
248	100		
250	91.2	80.5	120.7
141	92.3	37.2	55.8#



Abundance Scan 1466 (16.325 min): BN014570.D\data.ms (-1459) (-)





Abundance Scan 1476 (16.447 min): BN014565.D\data.ms (-1472) (-)





Hexachlorobenzene
Concen: 0.36 ng
RT: 16.447 min Scan# 1476
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	284	Resp:	1242
Ion	Ratio	Lower	Upper
284	100		
142	27.8	26.6	39.8
249	34.1	23.6	35.4



Abundance Scan 1476 (16.447 min): BN014570.D\data.ms (-1468) (-)









Scan 1489 (16.605 min): BN014565.D\data.ms (-1423) (-)
 Atrazine
 Concen: 0.31 ng
 RT: 16.605 min Scan# 1489
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	200	Resp:	627
Ion Ratio	Lower	Upper	
200	100		
173	30.3	23.0	34.6
215	58.0	38.6	57.8#



Scan 1504 (16.788 min): BN014565.D\data.ms (-1429) (-)
 Pentachlorophenol
 Concen: 0.28 ng
 RT: 16.788 min Scan# 1504
 Delta R.T. -0.000 min
 Lab File: BN014570.D
 Acq: 11 May 2021 19:55

Tgt Ion:	266	Resp:	320
Ion Ratio	Lower	Upper	
266	100		
264	59.4	24.2	36.4#
268	67.8	25.2	37.8#









#25 (-)
Phenanthrene
Concen: 0.35 ng
RT: 17.177 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	178	Resp:	5090
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.2	22.8
179	15.2	12.4	18.6




#26 (-)
Anthracene
Concen: 0.33 ng
RT: 17.262 min Scan# 1543
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	178	Resp:	4042
Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.8	22.2
179	15.6	12.2	18.2





Abundance Scan 1838 (19.173 min): BN014565.D\data.ms (-1825) (-)




Fluoranthene-d10
Concen: 0.35 ng
RT: 19.168 min Scan# 1837
Delta R.T. -0.005 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	212	Resp:	8183
Ion	Ratio	Lower	Upper
212	100		
106	10.1	10.1	15.1
104	6.2	5.4	8.2



Abundance Scan 1837 (19.168 min): BN014570.D\data.ms (-1798) (-)





Abundance Scan 1844 (19.203 min): BN014565.D\data.ms (-1825) (-)




Fluoranthene
Concen: 0.35 ng
RT: 19.203 min Scan# 1844
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55


Tgt Ion:	202	Resp:	6631
Ion	Ratio	Lower	Upper
202	100		
101	9.4	8.4	12.6
203	17.3	13.5	20.3



Abundance Scan 1844 (19.203 min): BN014570.D\data.ms (-1804) (-)






Abundance Scan 2131 (21.335 min): BN014565.D\data.ms (-2129) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.324 min Scan# 2130
Delta R.T. -0.011 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	240	Resp:	6853
Ion	Ratio	Lower	Upper
240	100		
120	11.7	5.8	8.8#
236	28.7	20.6	30.8






Abundance Scan 1917 (19.566 min): BN014565.D\data.ms (-1939) (-)




Pyrene
Concen: 0.35 ng
RT: 19.566 min Scan# 1917
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	202	Resp:	6576
Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.3	24.5
203	18.0	14.0	21.0






Abundance Scan 1959 (19.774 min): BN014565.D\data.ms (-1929) (-)




Terphenyl-d14
Concen: 0.37 ng
RT: 19.774 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.1	6.0	9.0
122	9.0	9.3	13.9#







Abundance Scan 2129 (21.314 min): BN014565.D\data.ms (-2129) (-)



Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.7	32.5
229	20.5	15.7	23.5









#32 (-)
Chrysene
Concen: 0.37 ng
RT: 21.367 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	228	Resp:	7700
Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.8	35.6
229	20.6	15.6	23.4



#24 (-)
Bis(2-ethylhexyl)phthalate
Concen: 0.32 ng
RT: 21.261 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55


Tgt Ion:	149	Resp:	1770
Ion	Ratio	Lower	Upper
149	100		
167	36.0	20.5	30.7#
279	7.3	2.9	4.3#






Abundance Scan 2581 (23.602 min): BN014565.D\data.ms (-2535) (-)

Perylene-d12
Concen: 0.40 ng
RT: 23.602 min Scan# 2581
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:	264	Resp:	6731
Ion	Ratio	Lower	Upper
264	100		
260	25.9	19.5	29.3
265	51.7	35.4	53.0






Abundance Scan 3253 (25.902 min): BN014565.D\data.ms (-3230) (-)


Indeno(1,2,3-cd)pyrene
Concen: 0.36 ng
RT: 25.899 min Scan# 3252
Delta R.T. -0.003 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55



Tgt Ion:	276	Resp:	9752
Ion	Ratio	Lower	Upper
276	100		
138	16.8	18.8	28.2#
277	25.8	19.8	29.8






Abundance Scan 2381 (22.918 min): BN014565.D\data.ms (-2381) (-)




Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 22.962 min Scan# 2394
Delta R.T. 0.044 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion:	252	Resp:	8433
Ion Ratio	Lower	Upper	
252	100		
253	25.4	19.6	29.4
125	10.4	10.3	15.5

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0






Abundance Scan 2394 (22.962 min): BN014565.D\data.ms (-2388) (-)



Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.962 min Scan# 2394
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55


Tgt Ion:	252	Resp:	8433
Ion Ratio	Lower	Upper	
252	100		
253	25.4	19.7	29.5
125	10.4	9.7	14.5






Abundance Scan 2552 (23.503 min): BN014565.D\data.ms (-2529) (-)

Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.503 min Scan# 2552
Delta R.T. -0.000 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.3	21.8	32.6
125	12.3	14.9	22.3#




Abundance Scan 3257 (25.916 min): BN014565.D\data.ms (-3249) (-)

Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.913 min Scan# 3256
Delta R.T. -0.004 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Tgt Ion	Ratio	Lower	Upper
278	100		
139	12.1	13.9	20.9#
279	26.1	20.7	31.1




Abundance Scan 3457 (26.601 min): BN014565.D\data.ms (-3433) (-)




Benzo(g,h,i)perylene
Concen: 0.35 ng
RT: 26.597 min Scan# 3456
Delta R.T. -0.003 min
Lab File: BN014570.D
Acq: 11 May 2021 19:55

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Abundance Scan 3456 (26.597 min): BN014570.D\data.ms



Tgt Ion:	276	Resp:	7752
Ion	Ratio	Lower	Upper
276	100		
277	25.8	19.6	29.4
138	14.0	17.5	26.3#



Abundance Scan 3456 (26.597 min): BN014570.D\data.ms (-3399) (-)

