

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051122\
 Data File : BN019742.D
 Acq On : 12 May 2022 04:12
 Operator : CG/JU
 Sample : N2451-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 50

Instrument :
 BNA_N
 ClientSampleId :
 SB-3145

Quant Time: May 12 05:53:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 10 01:35:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.862	152	209218	20.000	ng	0.00
7) Naphthalene-d8	10.648	136	670066	20.000	ng	# 0.00
13) Acenaphthene-d10	14.474	164	408822	20.000	ng	0.00
19) Phenanthrene-d10	17.215	188	886536	20.000	ng	0.00
29) Chrysene-d12	21.404	240	825624	20.000	ng	0.00
35) Perylene-d12	23.744	264	806880	20.000	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1347003	142.304	ng	0.00
5) Phenol-d6	7.024	99	1769092	148.425	ng	0.00
8) Nitrobenzene-d5	9.003	82	883291	78.878	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	8	0.000	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	463324	156.050	ng	0.00
15) 2-Fluorobiphenyl	13.105	172	2181035	59.124	ng	0.00
27) Fluoranthene-d10	19.253	212	14	0.000	ng	0.00
31) Terphenyl-d14	19.847	244	2126100	55.398	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.324	149	1507	0.056	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051122\
Data File : BN019742.D
Acq On : 12 May 2022 04:12
Operator : CG/JU
Sample : N2451-04
Misc :
ALS Vial : 21 Sample Multiplier: 50

Instrument :
BNA_N
ClientSampleId :
SB-3145

Quant Time: May 12 05:53:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 10 01:35:26 2022
Response via : Initial Calibration

