

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051123\
 Data File : BN025400.D
 Acq On : 11 May 2023 14:38
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN051123

Quant Time: May 12 00:58:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 00:57:16 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.476	-19.0	111	0.00
3	n-Nitrosodimethylamine	0.400	0.451	-12.7	113	0.00
4 S	2-Fluorophenol	0.400	0.449	-12.2	106	0.00
5 S	Phenol-d6	0.400	0.452	-13.0	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.429	-7.2	105	-0.01
9	Naphthalene	0.400	0.445	-11.2	106	0.00
10	Hexachlorobutadiene	0.400	0.445	-11.2	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.441	-10.2	105	0.00
12	2-Methylnaphthalene	0.400	0.440	-10.0	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.417	-4.2	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.449	-12.2	105	0.00
16	Acenaphthylene	0.400	0.439	-9.7	105	0.00
17	Acenaphthene	0.400	0.452	-13.0	105	0.00
18	Fluorene	0.400	0.447	-11.7	104	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.431	-7.7	142	0.00
21	4-Bromophenyl-phenylether	0.400	0.444	-11.0	106	0.00
22	Hexachlorobenzene	0.400	0.450	-12.5	104	0.00
23	Atrazine	0.400	0.418	-4.5	103	0.00
24	Pentachlorophenol	0.400	0.386	3.5	100	0.00
25	Phenanthrene	0.400	0.440	-10.0	102	0.00
26	Anthracene	0.400	0.425	-6.2	102	0.00
27 SURR	Fluoranthene-d10	0.400	0.425	-6.2	100	0.00
28	Fluoranthene	0.400	0.422	-5.5	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.470	-17.5	100	0.00
31 S	Terphenyl-d14	0.400	0.467	-16.8	102	0.00
32	Benzo(a)anthracene	0.400	0.439	-9.7	95	0.00
33	Chrysene	0.400	0.475	-18.7	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.441	-10.2	94	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.446	-11.5	98	0.02
37	Benzo(b)fluoranthene	0.400	0.425	-6.2	97	0.00
38	Benzo(k)fluoranthene	0.400	0.438	-9.5	97	0.00
39 C	Benzo(a)pyrene	0.400	0.440	-10.0	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.447	-11.7	98	0.02
41	Benzo(g,h,i)perylene	0.400	0.445	-11.2	98	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN051123