

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051123\
 Data File : BN025393.D
 Acq On : 11 May 2023 10:28
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: May 12 00:52:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 00:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.896	152	11837	0.400	ng	0.00
7) Naphthalene-d8	10.707	136	38089	0.400	ng	0.00
13) Acenaphthene-d10	14.544	164	22436	0.400	ng	0.00
19) Phenanthrene-d10	17.298	188	44372	0.400	ng	0.00
29) Chrysene-d12	21.495	240	34548	0.400	ng	# 0.00
35) Perylene-d12	23.932	264	30706	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.455	112	2753	0.100	ng	0.00
5) Phenol-d6	7.080	99	3495	0.094	ng	0.00
8) Nitrobenzene-d5	9.084	82	2361	0.077	ng	0.00
11) 2-Methylnaphthalene-d10	12.309	152	4580	0.091	ng	0.00
14) 2,4,6-Tribromophenol	16.057	330	1020	0.085	ng	0.00
15) 2-Fluorobiphenyl	13.176	172	6948	0.089	ng	0.00
27) Fluoranthene-d10	19.334	212	9846	0.097	ng	0.00
31) Terphenyl-d14	19.924	244	7899	0.095	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.318	88	1067	0.098	ng	# 91
3) n-Nitrosodimethylamine	3.657	42	1438	0.082	ng	# 87
6) bis(2-Chloroethyl)ether	7.326	93	2586	0.091	ng	97
9) Naphthalene	10.760	128	9268	0.097	ng	98
10) Hexachlorobutadiene	11.038	225	1830	0.098	ng	# 98
12) 2-Methylnaphthalene	12.381	142	6107	0.092	ng	94
16) Acenaphthylene	14.266	152	9814	0.094	ng	99
17) Acenaphthene	14.608	154	5884	0.095	ng	98
18) Fluorene	15.603	166	7627	0.093	ng	100
21) 4-Bromophenyl-phenylether	16.491	248	2324	0.093	ng	# 90
22) Hexachlorobenzene	16.603	284	2661	0.096	ng	98
23) Atrazine	16.765	200	1916	0.094	ng	93
25) Phenanthrene	17.336	178	11499	0.096	ng	99
26) Anthracene	17.435	178	10636	0.091	ng	99
28) Fluoranthene	19.363	202	13736	0.099	ng	100
30) Pyrene	19.726	202	14092	0.097	ng	99
32) Benzo(a)anthracene	21.477	228	11387	0.097	ng	# 89
33) Chrysene	21.531	228	10739	0.094	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.387	149	13053	0.115	ng	99
36) Indeno(1,2,3-cd)pyrene	26.481	276	10991	0.097	ng	98
37) Benzo(b)fluoranthene	23.189	252	10087	0.100	ng	# 63
38) Benzo(k)fluoranthene	23.239	252	9838	0.095	ng	# 62
39) Benzo(a)pyrene	23.829	252	9392	0.096	ng	# 51
40) Dibenzo(a,h)anthracene	26.493	278	8641	0.096	ng	# 68
41) Benzo(g,h,i)perylene	27.253	276	9617	0.101	ng	# 80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051123\
Data File : BN025393.D
Acq On : 11 May 2023 10:28
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: May 12 00:52:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 12 00:52:05 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.896 min Scan# 61
Delta R.T. -0.008 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 11837
Ion Ratio Lower Upper
152 100
150 157.8 121.4 182.2
115 63.3 50.6 76.0



#2
1,4-Dioxane
Concen: 0.098 ng
RT: 3.318 min Scan# 19
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion: 88 Resp: 1067
Ion Ratio Lower Upper
88 100
43 58.9 37.4 56.2#
58 92.7 71.5 107.3









#3
n-Nitrosodimethylamine
Concen: 0.082 ng
RT: 3.657 min Scan# 60
Delta R.T. 0.007 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 1438
Ion Ratio Lower Upper
42 100
74 114.8 82.2 123.4
44 0.0 7.8 11.8#



#4
2-Fluorophenol
Concen: 0.100 ng
RT: 5.455 min Scan# 315
Delta R.T. -0.007 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion: 112 Resp: 2753
Ion Ratio Lower Upper
112 100
64 64.9 51.9 77.9
63 35.6 27.4 41.2





#5

Phenol-d6

Concen: 0.094 ng

RT: 7.080 min Scan# 540

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion: 99 Resp: 3495

Ion Ratio Lower Upper

99 100

42 29.7 23.5 35.3

71 35.5 28.1 42.1



#6

bis(2-Chloroethyl)ether

Concen: 0.091 ng

RT: 7.326 min Scan# 574

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion: 93 Resp: 2586


Ion Ratio Lower Upper

93 100

63 86.3 71.3 106.9

95 36.1 31.0 46.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	38089
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.1	13.7
54	13.0	11.2	16.8
68	8.4	7.4	11.0



#8
Nitrobenzene-d5
Concen: 0.077 ng
RT: 9.084 min Scan# 783
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:	82	Resp:	2361
Ion	Ratio	Lower	Upper
82	100		
128	52.3	35.8	53.8
54	75.3	50.4	75.6









#9
Naphthalene
Concen: 0.097 ng
RT: 10.760 min Scan# 940
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 9268
Ion Ratio Lower Upper
128 100
129 13.0 9.3 13.9
127 14.1 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.098 ng
RT: 11.038 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:225 Resp: 1830
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 65.0 50.8 76.2





#11

2-Methylnaphthalene-d10

Concen: 0.091 ng

RT: 12.309 min Scan# 1222

Delta R.T. 0.004 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :





SSTDIC0.1

Tgt Ion:152 Resp: 4580

Ion Ratio Lower Upper

152 100

151 19.4 16.1 24.1



#12

2-Methylnaphthalene

Concen: 0.092 ng

RT: 12.381 min Scan# 1241

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion:142 Resp: 6107


Ion Ratio Lower Upper

142 100

141 90.8 69.3 103.9

115 43.3 30.2 45.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.544 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 22436
Ion Ratio Lower Upper
164 100
162 104.3 83.2 124.8
160 48.1 38.8 58.2



#14
2,4,6-Tribromophenol
Concen: 0.085 ng
RT: 16.057 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:330 Resp: 1020
Ion Ratio Lower Upper
330 100
332 96.8 78.0 117.0
141 39.6 30.2 45.2





#15

2-Fluorobiphenyl

Concen: 0.089 ng

RT: 13.176 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:172 Resp: 6948

Ion Ratio Lower Upper

172 100

171 35.6 26.7 40.1

170 25.1 17.8 26.8



#16

Acenaphthylene

Concen: 0.094 ng

RT: 14.266 min Scan# 1524

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion:152 Resp: 9814


Ion Ratio Lower Upper

152 100

151 20.1 15.9 23.9

153 12.7 10.7 16.1





#17

Acenaphthene

Concen: 0.095 ng

RT: 14.608 min Scan# 1556

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:154 Resp: 5884

Ion Ratio Lower Upper

154 100

153 116.1 91.7 137.5

152 57.3 44.6 66.8



#18

Fluorene

Concen: 0.093 ng

RT: 15.603 min Scan# 1649

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion:166 Resp: 7627


Ion Ratio Lower Upper

166 100

165 97.9 78.7 118.1

167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.298 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 44372
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.5 9.4 14.2



#21
4-Bromophenyl-phenylether
Concen: 0.093 ng
RT: 16.491 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:248 Resp: 2324
Ion Ratio Lower Upper
248 100
250 97.5 83.1 124.7
141 55.9 35.0 52.6#









#22
Hexachlorobenzene
Concen: 0.096 ng
RT: 16.603 min Scan# 1731
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:284 Resp: 2661
Ion Ratio Lower Upper
284 100
142 38.2 29.7 44.5
249 32.1 24.7 37.1



#23
Atrazine
Concen: 0.094 ng
RT: 16.765 min Scan# 1744
Delta R.T. 0.012 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:200 Resp: 1916
Ion Ratio Lower Upper
200 100
173 27.9 26.5 39.7
215 49.7 37.0 55.4









#25
Phenanthrene
Concen: 0.096 ng
RT: 17.336 min Scan# 1790
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 11499
Ion Ratio Lower Upper
178 100
176 19.4 15.3 22.9
179 15.9 12.1 18.1



#26
Anthracene
Concen: 0.091 ng
RT: 17.435 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:178 Resp: 10636
Ion Ratio Lower Upper
178 100
176 18.2 15.0 22.4
179 15.1 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.097 ng
RT: 19.334 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 9846
Ion Ratio Lower Upper
212 100
106 16.1 12.6 19.0
104 9.1 7.0 10.4



#28
Fluoranthene
Concen: 0.099 ng
RT: 19.363 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:202 Resp: 13736
Ion Ratio Lower Upper
202 100
101 12.3 9.8 14.8
203 17.0 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.495 min Scan# 2485

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:240 Resp: 34548

Ion Ratio Lower Upper

240 100

120 18.2 19.0 28.6#

236 29.6 25.0 37.6



#30

Pyrene

Concen: 0.097 ng

RT: 19.726 min Scan# 2252

Delta R.T. -0.004 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion:202 Resp: 14092


Ion Ratio Lower Upper

202 100

200 21.2 16.8 25.2

203 18.4 14.4 21.6









#31
 Terphenyl-d14
 Concen: 0.095 ng
 RT: 19.924 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN025393.D
 Acq: 11 May 2023 10:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 7899
 Ion Ratio Lower Upper
 244 100
 212 12.1 7.5 11.3#
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 0.097 ng
 RT: 21.477 min Scan# 2483
 Delta R.T. -0.009 min
 Lab File: BN025393.D
 Acq: 11 May 2023 10:28

Tgt Ion:228 Resp: 11387
 Ion Ratio Lower Upper
 228 100
 226 34.1 23.2 34.8
 229 29.4 18.2 27.2#





#33

Chrysene

Concen: 0.094 ng

RT: 21.531 min Scan# 2489

Delta R.T. -0.009 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 10739

Ion Ratio Lower Upper

228 100

226 36.2 25.8 38.6

229 28.4 17.9 26.9#



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.115 ng

RT: 21.387 min Scan# 2473

Delta R.T. 0.000 min

Lab File: BN025393.D

Acq: 11 May 2023 10:28




Tgt Ion:149 Resp: 13053


Ion Ratio Lower Upper

149 100

167 27.9 21.7 32.5

279 3.6 3.0 4.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.932 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:264 Resp: 30706
Ion Ratio Lower Upper
264 100
260 29.1 23.5 35.3
265 50.7 50.9 76.3#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.097 ng
RT: 26.481 min Scan# 3980
Delta R.T. 0.032 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:276 Resp: 10991
Ion Ratio Lower Upper
276 100
138 27.8 22.7 34.1
277 26.7 20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 0.100 ng
RT: 23.189 min Scan# 2853
Delta R.T. 0.003 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 10087
Ion Ratio Lower Upper
252 100
253 43.4 23.4 35.0#
125 49.6 19.8 29.8#



#38
Benzo(k)fluoranthene
Concen: 0.095 ng
RT: 23.239 min Scan# 2871
Delta R.T. 0.006 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:252 Resp: 9838
Ion Ratio Lower Upper
252 100
253 44.3 23.6 35.4#
125 50.0 20.2 30.2#









#39
Benzo(a)pyrene
Concen: 0.096 ng
RT: 23.829 min Scan# 3073
Delta R.T. 0.009 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 9392
Ion Ratio Lower Upper
252 100
253 49.3 25.4 38.2#
125 66.4 23.5 35.3#



#40
Dibenzo(a,h)anthracene
Concen: 0.096 ng
RT: 26.493 min Scan# 3984
Delta R.T. 0.029 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Tgt Ion:278 Resp: 8641
Ion Ratio Lower Upper
278 100
139 48.7 23.8 35.8#
279 49.1 26.0 39.0#





#41
Benzo(g,h,i)perylene
Concen: 0.101 ng
RT: 27.253 min Scan# 41
Delta R.T. 0.023 min
Lab File: BN025393.D
Acq: 11 May 2023 10:28

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion:	276	Resp:	9617
Ion Ratio	Lower	Upper	
276	100		
277	39.5	23.5	35.3#
138	42.6	24.7	37.1#

