

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
 Data File : BN010693.D
 Acq On : 12 May 2020 10:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 12 11:00:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 10:58:34 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	1,4-Dioxane	0.294	0.307	-4.4	71	0.00
3	n-Nitrosodimethylamine	0.191	0.179	6.3	75	0.00
4 S	2-Fluorophenol	0.876	0.810	7.5	62	0.00
5 S	Phenol-d6	1.145	0.981	14.3	58	0.00
6	bis(2-Chloroethyl)ether	0.992	1.036	-4.4	69	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
8 S	Nitrobenzene-d5	0.280	0.269	3.9	69	0.00
9	Naphthalene	0.993	1.000	-0.7	70	0.00
10	Hexachlorobutadiene	0.239	0.221	7.5	63	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.664	-0.9	70	0.00
12	2-Methylnaphthalene	0.714	0.721	-1.0	70	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.189	0.128	32.3#	51	0.00
15 S	2-Fluorobiphenyl	1.456	1.439	1.2	70	0.00
16	Acenaphthylene	1.728	1.505	12.9	64	0.00
17	Acenaphthene	1.110	1.063	4.2	69	0.00
18	Fluorene	1.454	1.393	4.2	69	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
20	4-Bromophenyl-phenylether	0.222	0.211	5.0	66	0.00
21	Hexachlorobenzene	0.254	0.258	-1.6	68	0.00
22	Pentachlorophenol	0.100	0.088	12.0	71	0.00
23	Phenanthrene	1.080	1.029	4.7	65	0.00
24	Anthracene	0.988	0.835	15.5	59	0.00
25 SURR	Fluoranthene-d10	1.213	1.040	14.3	58	0.00
26	Fluoranthene	1.345	1.161	13.7	59	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
28	Pyrene	1.346	1.516	-12.6	57	0.00
29 S	Terphenyl-d14	0.915	0.972	-6.2	55	0.00
30	Benzo(a)anthracene	1.305	1.106	15.2	43#	0.00
31	Chrysene	1.300	1.280	1.5	51	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.510	23.9	41#	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	0.972	15.4	43#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	53	0.00
35	Benzo(b)fluoranthene	1.317	1.282	2.7	51	0.00
36	Benzo(k)fluoranthene	1.320	1.366	-3.5	55	0.00
37 C	Benzo(a)pyrene	1.166	1.077	7.6	49#	0.00
38	Dibenzo(a,h)anthracene	1.066	0.756	29.1#	37#	0.00
39	Benzo(g,h,i)perylene	1.073	1.045	2.6	51	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
Data File : BN010693.D
Acq On : 12 May 2020 10:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 11:00:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 10:58:34 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		