

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
 Data File : BN010693.D
 Acq On : 12 May 2020 10:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 12 11:00:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 10:58:34 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	67	0.00
2	1,4-Dioxane	0.400	0.417	-4.2	71	0.00
3	n-Nitrosodimethylamine	0.400	0.375	6.3	75	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	62	0.00
5 S	Phenol-d6	0.400	0.343	14.2	58	0.00
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	69	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	69	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	69	0.00
9	Naphthalene	0.400	0.403	-0.8	70	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	63	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.404	-1.0	70	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	70	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.400	0.271	32.3#	51	0.00
15 S	2-Fluorobiphenyl	0.400	0.396	1.0	70	0.00
16	Acenaphthylene	0.400	0.348	13.0	64	0.00
17	Acenaphthene	0.400	0.383	4.3	69	0.00
18	Fluorene	0.400	0.383	4.3	69	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	68	0.00
20	4-Bromophenyl-phenylether	0.400	0.379	5.3	66	0.00
21	Hexachlorobenzene	0.400	0.407	-1.7	68	0.00
22	Pentachlorophenol	0.400	0.354	11.5	71	0.00
23	Phenanthrene	0.400	0.381	4.8	65	0.00
24	Anthracene	0.400	0.338	15.5	59	0.00
25 SURR	Fluoranthene-d10	0.400	0.343	14.2	58	0.00
26	Fluoranthene	0.400	0.345	13.8	59	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	51	0.00
28	Pyrene	0.400	0.450	-12.5	57	0.00
29 S	Terphenyl-d14	0.400	0.425	-6.2	55	0.00
30	Benzo(a)anthracene	0.400	0.339	15.3	43	0.00
31	Chrysene	0.400	0.394	1.5	51	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.305	23.8	41	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.338	15.5	43	0.00
34 I	Perylene-d12	0.400	0.400	0.0	53	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	51	0.00
36	Benzo(k)fluoranthene	0.400	0.414	-3.5	55	0.00
37 C	Benzo(a)pyrene	0.400	0.369	7.8	49	0.00
38	Dibenzo(a,h)anthracene	0.400	0.284	29.0#	37	0.00
39	Benzo(g,h,i)perylene	0.400	0.390	2.5	51	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
Data File : BN010693.D
Acq On : 12 May 2020 10:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 11:00:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 10:58:34 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				