

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
 Data File : BN010698.D
 Acq On : 12 May 2020 13:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 12 14:17:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 10:58:34 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.294	0.340	-15.6	90	0.00
3	n-Nitrosodimethylamine	0.191	0.208	-8.9	100	0.00
4 S	2-Fluorophenol	0.876	0.796	9.1	70	0.00
5 S	Phenol-d6	1.145	0.975	14.8	65	0.00
6	bis(2-Chloroethyl)ether	0.992	1.044	-5.2	79	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
8 S	Nitrobenzene-d5	0.280	0.276	1.4	76	-0.01
9	Naphthalene	0.993	0.999	-0.6	75	0.00
10	Hexachlorobutadiene	0.239	0.233	2.5	71	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.640	2.7	72	0.00
12	2-Methylnaphthalene	0.714	0.691	3.2	72	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.189	0.132	30.2#	53	0.00
15 S	2-Fluorobiphenyl	1.456	1.459	-0.2	72	0.00
16	Acenaphthylene	1.728	1.538	11.0	66	0.00
17	Acenaphthene	1.110	1.083	2.4	71	0.00
18	Fluorene	1.454	1.398	3.9	70	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
20	4-Bromophenyl-phenylether	0.222	0.213	4.1	67	0.00
21	Hexachlorobenzene	0.254	0.261	-2.8	69	0.00
22	Pentachlorophenol	0.100	0.072	28.0#	58	0.00
23	Phenanthrene	1.080	1.035	4.2	65	0.00
24	Anthracene	0.988	0.846	14.4	59	0.00
25 SURR	Fluoranthene-d10	1.213	1.041	14.2	58	0.00
26	Fluoranthene	1.345	1.170	13.0	59	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	54	-0.01
28	Pyrene	1.346	1.411	-4.8	57	0.00
29 S	Terphenyl-d14	0.915	0.946	-3.4	57	0.00
30	Benzo(a)anthracene	1.305	1.160	11.1	49#	0.00
31	Chrysene	1.300	1.272	2.2	54	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.492	26.6#	42#	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.145	0.3	54	0.00
34 I	Perylene-d12	1.000	1.000	0.0	58	0.00
35	Benzo(b)fluoranthene	1.317	1.275	3.2	56	0.00
36	Benzo(k)fluoranthene	1.320	1.389	-5.2	62	0.00
37 C	Benzo(a)pyrene	1.166	1.098	5.8	55	0.00
38	Dibenzo(a,h)anthracene	1.066	0.967	9.3	53	0.00
39	Benzo(g,h,i)perylene	1.073	1.086	-1.2	59	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051220\
Data File : BN010698.D
Acq On : 12 May 2020 13:47
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 14:17:53 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 10:58:34 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			