

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051221\
 Data File : BN014579.D
 Acq On : 12 May 2021 15:03
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV007

Quant Time: May 12 15:38:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 12 14:39:16 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	680	0.40	ng/ul	0.00
4) Naphthalene-d8	10.53	136	2244	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.39	164	1453	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.14	188	3619	0.40	ng/ul	0.00
17) Chrysene-d12	21.34	240	4090	0.40	ng/ul	0.00
23) Perylene-d12	23.61	264	3692	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.27	96	291	0.41	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.13	152	1523	0.40	ng/ul	0.00
18) Fluoranthene-d10	19.17	212	4344	0.39	ng/ul	0.00
Target Compounds					Ovalue	
2) 1,4-Dioxane	3.30	88	296	0.382	ng/ul	99
5) Naphthalene	10.58	128	2631	0.407	ng/ul	98
7) 2-Methylnaphthalene	12.20	142	1811	0.391	ng/ul	97
8) 1-Methylnaphthalene	12.42	142	1785	0.401	ng/ul	96
10) Acenaphthylene	14.11	152	2215	0.400	ng/ul	99
11) Acenaphthene	14.45	153	1993	0.400	ng/ul	97
12) Fluorene	15.44	166	2402	0.399	ng/ul	98
14) Pentachlorophenol	16.79	266	263	0.371	ng/ul	99
15) Phenanthrene	17.18	178	4292	0.393	ng/ul	98
16) Anthracene	17.27	178	3472	0.384	ng/ul	97
19) Fluoranthene	19.20	202	5297	0.396	ng/ul	99
20) Pyrene	19.57	202	5233	0.395	ng/ul	99
21) Benzo(a)anthracene	21.32	228	4817	0.389	ng/ul	99
22) Chrysene	21.37	228	5801	0.394	ng/ul	99
24) Benzo(b)fluoranthene	22.92	252	5756	0.372	ng/ul	100
25) Benzo(k)fluoranthene	22.97	252	6466	0.401	ng/ul	99
26) Benzo(a)pyrene	23.51	252	4624	0.379	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.91	276	7257	0.407	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.92	278	5908	0.403	ng/ul	99
29) Benzo(a,h,i)perylene	26.60	276	5688	0.384	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051221\
Data File : BN014579.D
Acq On : 12 May 2021 15:03
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SICV007

Quant Time: May 12 15:38:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 12 14:39:16 2021
Response via : Initial Calibration

