

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051321\
 Data File : BN014593.D
 Acq On : 13 May 2021 17:00
 Operator : CG/JU
 Sample : M2197-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 YARG3

Manual Integrations
 APPROVED

mohammad
 5/14/2021 12:50:49 PM

Quant Time: May 13 17:38:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 13 12:42:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	699	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	2535	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1822	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	4330m	0.40	ng/ul	0.00
17) Chrysene-d12	21.33	240	4356	0.40	ng/ul	0.00
23) Perylene-d12	23.60	264	4383	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	2507	3.47	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.12	152	1231	0.28	ng/ul	0.00
18) Fluoranthene-d10	19.17	212	4330	0.37	ng/ul	0.00
Target Compounds						
2) 1,4-Dioxane	3.29	88	58	0.073	ng/ul#	Ovalue 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051321\
Data File : BN014593.D
Acq On : 13 May 2021 17:00
Operator : CG/JU
Sample : M2197-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
YARG3

Manual Integrations
APPROVED

mohammad
5/14/2021 12:50:49 PM

Quant Time: May 13 17:38:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 13 12:42:54 2021
Response via : Initial Calibration

