

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051420\  
 Data File : BN010710.D  
 Acq On : 14 May 2020 12:14  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: May 14 12:57:50 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu May 14 12:56:11 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | AvqRF | CCRF  | %Dev | Area% | Dev(min) |
|---------|----------------------------|-------|-------|------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 1.000 | 1.000 | 0.0  | 111   | 0.00     |
| 2       | 1,4-Dioxane                | 0.345 | 0.348 | -0.9 | 118   | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.199 | 0.202 | -1.5 | 120   | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.783 | 0.809 | -3.3 | 116   | 0.00     |
| 5 S     | Phenol-d6                  | 0.957 | 0.953 | 0.4  | 113   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.986 | 1.052 | -6.7 | 114   | 0.00     |
| 7 I     | Naphthalene-d8             | 1.000 | 1.000 | 0.0  | 109   | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.278 | 0.271 | 2.5  | 109   | 0.00     |
| 9       | Naphthalene                | 1.006 | 0.996 | 1.0  | 109   | -0.01    |
| 10      | Hexachlorobutadiene        | 0.230 | 0.232 | -0.9 | 109   | -0.01    |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.652 | 0.639 | 2.0  | 108   | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.709 | 0.696 | 1.8  | 109   | 0.00     |
| 13 I    | Acenaphthene-d10           | 1.000 | 1.000 | 0.0  | 110   | -0.01    |
| 14 S    | 2,4,6-Tribromophenol       | 0.145 | 0.134 | 7.6  | 114   | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 1.468 | 1.494 | -1.8 | 111   | 0.00     |
| 16      | Acenaphthylene             | 1.634 | 1.544 | 5.5  | 110   | 0.00     |
| 17      | Acenaphthene               | 1.104 | 1.100 | 0.4  | 109   | 0.00     |
| 18      | Fluorene                   | 1.447 | 1.406 | 2.8  | 110   | -0.01    |
| 19 I    | Phenanthrene-d10           | 1.000 | 1.000 | 0.0  | 114   | 0.00     |
| 20      | 4-Bromophenyl-phenylether  | 0.221 | 0.213 | 3.6  | 110   | 0.00     |
| 21      | Hexachlorobenzene          | 0.263 | 0.262 | 0.4  | 112   | 0.00     |
| 22      | Pentachlorophenol          | 0.081 | 0.083 | -2.5 | 133   | -0.01    |
| 23      | Phenanthrene               | 1.073 | 1.054 | 1.8  | 115   | -0.01    |
| 24      | Anthracene                 | 0.902 | 0.860 | 4.7  | 116   | -0.01    |
| 25 SURR | Fluoranthene-d10           | 1.085 | 1.061 | 2.2  | 120   | 0.00     |
| 26      | Fluoranthene               | 1.219 | 1.191 | 2.3  | 119   | 0.00     |
| 27 I    | Chrysene-d12               | 1.000 | 1.000 | 0.0  | 126   | 0.00     |
| 28      | Pvrene                     | 1.434 | 1.386 | 3.3  | 120   | 0.00     |
| 29 S    | Terphenyl-d14              | 0.949 | 0.948 | 0.1  | 125   | 0.00     |
| 30      | Benzo(a)anthracene         | 1.167 | 1.089 | 6.7  | 127   | 0.00     |
| 31      | Chrysene                   | 1.321 | 1.347 | -2.0 | 128   | 0.00     |
| 32      | Bis(2-ethylhexyl)phthalate | 0.511 | 0.473 | 7.4  | 125   | 0.01     |
| 33      | Indeno(1,2,3-cd)pyrene     | 1.175 | 1.180 | -0.4 | 135   | 0.00     |
| 34 I    | Perylene-d12               | 1.000 | 1.000 | 0.0  | 127   | 0.00     |
| 35      | Benzo(b)fluoranthene       | 1.322 | 1.262 | 4.5  | 128   | 0.00     |
| 36      | Benzo(k)fluoranthene       | 1.424 | 1.412 | 0.8  | 130   | 0.00     |
| 37 C    | Benzo(a)pyrene             | 1.144 | 1.101 | 3.8  | 126   | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 0.978 | 0.969 | 0.9  | 138   | 0.00     |
| 39      | Benzo(a,h,i)perylene       | 1.147 | 1.106 | 3.6  | 129   | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051420\  
Data File : BN010710.D  
Acq On : 14 May 2020 12:14  
Operator : CG/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_N  
LabSampleId :  
SSTDCCC0.4

Quant Time: May 14 12:57:50 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 14 12:56:11 2020  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | AvgRF                        | CCRF | %Dev Area | % Dev(min) |
|--------------------|------------------------------|------|-----------|------------|
| -----              |                              |      |           |            |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |      |           |            |