

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
 Data File : BN014623.D
 Acq On : 14 May 2021 13:31
 Operator : CG/JU
 Sample : SSTD0.404
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4004

Quant Time: May 14 14:01:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 14:00:50 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	470	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	1435	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1027	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	2617	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	3372	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	2633	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	204	0.42	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	1015	0.41	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	3375	0.37	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.29	88	204	0.381	ng/ul	100
5) Naphthalene	10.57	128	1684	0.408	ng/ul	100
7) 2-Methylnaphthalene	12.18	142	1215	0.411	ng/ul	100
8) 1-Methylnaphthalene	12.40	142	1179	0.414	ng/ul	100
10) Acenaphthylene	14.09	152	1408	0.359	ng/ul	100
11) Acenaphthene	14.44	153	1375	0.391	ng/ul	100
12) Fluorene	15.43	166	1655	0.389	ng/ul	100
14) Pentachlorophenol	16.78	266	124	0.242	ng/ul	100
15) Phenanthrene	17.17	178	3240	0.410	ng/ul	100
16) Anthracene	17.26	178	2541	0.388	ng/ul	100
19) Fluoranthene	19.19	202	4112	0.373	ng/ul	100
20) Pyrene	19.56	202	4076	0.374	ng/ul	100
21) Benzo(a)anthracene	21.31	228	4267	0.418	ng/ul	100
22) Chrysene	21.36	228	4959	0.409	ng/ul	100
24) Benzo(b)fluoranthene	22.91	252	4630	0.420	ng/ul	100
25) Benzo(k)fluoranthene	22.95	252	4832	0.420	ng/ul	100
26) Benzo(a)pyrene	23.49	252	3568	0.411	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.88	276	5004	0.394	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.90	278	4082	0.390	ng/ul	100
29) Benzo(g,h,i)perylene	26.58	276	4322	0.409	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
 Data File : BN014623.D
 Acq On : 14 May 2021 13:31
 Operator : CG/JU
 Sample : SSTD0.404
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 Client Sampled :
 SSTD0.4004

Quant Time: May 14 14:01:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 14:00:50 2021
 Response via : Initial Calibration

