

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
 Data File : BN014624.D
 Acq On : 14 May 2021 14:06
 Operator : CG/JU
 Sample : SST00.805
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST00.8005

Quant Time: May 14 14:36:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 14:00:50 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	465	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	1446	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	988	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	2597	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	3473	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	2616	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	375	0.78	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	2020	0.82	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	6704	0.71	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.28	88	379	0.715	ng/ul	94
5) Naphthalene	10.57	128	3367	0.809	ng/ul	95
7) 2-Methylnaphthalene	12.18	142	2428	0.814	ng/ul	98
8) 1-Methylnaphthalene	12.41	142	2366	0.825	ng/ul	100
10) Acenaphthylene	14.09	152	2963	0.786	ng/ul	94
11) Acenaphthene	14.44	153	2761	0.816	ng/ul	98
12) Fluorene	15.43	166	3411	0.834	ng/ul	97
14) Pentachlorophenol	16.78	266	292	0.573	ng/ul	95
15) Phenanthrene	17.17	178	6252	0.798	ng/ul	99
16) Anthracene	17.26	178	5141	0.792	ng/ul	97
19) Fluoranthene	19.19	202	8221	0.724	ng/ul	99
20) Pyrene	19.56	202	8128	0.723	ng/ul	97
21) Benzo(a)anthracene	21.31	228	8765	0.834	ng/ul	99
22) Chrysene	21.36	228	9868	0.790	ng/ul	98
24) Benzo(b)fluoranthene	22.91	252	9230	0.843	ng/ul	96
25) Benzo(k)fluoranthene	22.95	252	9746	0.852	ng/ul	97
26) Benzo(a)pyrene	23.49	252	7097	0.822	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	25.88	276	10322	0.818	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.90	278	8410	0.809	ng/ul	94
29) Benzo(g,h,i)perylene	26.58	276	8602	0.819	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
Data File : BN014624.D
Acq On : 14 May 2021 14:06
Operator : CG/JU
Sample : SST00.805
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.8005

Quant Time: May 14 14:36:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 14:00:50 2021
Response via : Initial Calibration

