

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
 Data File : BN014626.D
 Acq On : 14 May 2021 15:17
 Operator : CG/JU
 Sample : SSTD3.201
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2001

Quant Time: May 14 15:46:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 14:00:50 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	446	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	1490	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1033	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	2605	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	3703	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	2727	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	1429	3.10	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	7952	3.12	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	27395	2.73	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.28	88	1535	3.018	ng/ul#	84
5) Naphthalene	10.57	128	12800	2.984	ng/ul#	91
7) 2-Methylnaphthalene	12.18	142	9652	3.141	ng/ul	98
8) 1-Methylnaphthalene	12.41	142	9219	3.120	ng/ul	99
10) Acenaphthylene	14.09	152	12965	3.290	ng/ul#	92
11) Acenaphthene	14.44	153	11087	3.134	ng/ul	99
12) Fluorene	15.43	166	13674	3.197	ng/ul	98
14) Pentachlorophenol	16.78	266	1524	2.983	ng/ul	92
15) Phenanthrene	17.17	178	24727	3.146	ng/ul	98
16) Anthracene	17.26	178	22267	3.419	ng/ul	94
19) Fluoranthene	19.19	202	34174	2.824	ng/ul	97
20) Pyrene	19.55	202	33419	2.789	ng/ul	98
21) Benzo(a)anthracene	21.31	228	37497	3.344	ng/ul	98
22) Chrysene	21.36	228	39539	2.967	ng/ul	97
24) Benzo(b)fluoranthene	22.91	252	38658	3.385	ng/ul	93
25) Benzo(k)fluoranthene	22.95	252	41115	3.449	ng/ul#	93
26) Benzo(a)pyrene	23.49	252	31170	3.463	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	25.88	276	44592	3.388	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.90	278	36763	3.393	ng/ul#	92
29) Benzo(a,h,i)perylene	26.58	276	36408	3.325	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN051421\
Data File : BN014626.D
Acq On : 14 May 2021 15:17
Operator : CG/JU
Sample : SSTD3.201
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2001

Quant Time: May 14 15:46:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 14:00:50 2021
Response via : Initial Calibration

