

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051421\
 Data File : BN014651.D
 Acq On : 15 May 2021 16:52
 Operator : CG/JU
 Sample : M2397-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG1Z0

Manual Integrations
 APPROVED

mohammad
 5/17/2021 11:20:57 AM

Quant Time: May 16 01:59:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.730	152	414	0.40	ng/ul	0.00
4) Naphthalene-d8	10.511	136	1410	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.376	164	1123	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.123	188	2646m	0.40	ng/ul	0.00
17) Chrysene-d12	21.321	240	2933	0.40	ng/ul	0.00
23) Perylene-d12	23.592	264	3098	0.40	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.250	96	601	1.42	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.105	152	817	0.33	ng/ul	0.00
18) Fluoranthene-d10	19.159	212	3250	0.44	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.560	128	332	0.08	ng/ul#	75
7) 2-Methylnaphthalene	12.188	142	79	0.03	ng/ul	97
12) Fluorene	15.426	166	101	0.02	ng/ul#	86
15) Phenanthrene	17.165	178	751	0.09	ng/ul#	86
16) Anthracene	17.254	178	152	0.02	ng/ul#	48
19) Fluoranthene	19.191	202	414	0.05	ng/ul#	70
20) Pyrene	19.554	202	1214m	0.13	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051421\
Data File : BN014651.D
Acq On : 15 May 2021 16:52
Operator : CG/JU
Sample : M2397-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG1Z0

Manual Integrations
APPROVED

mohammad
5/17/2021 11:20:57 AM

Quant Time: May 16 01:59:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

