

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051421\
 Data File : BN014644.D
 Acq On : 15 May 2021 12:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4014

Quant Time: May 15 12:39:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.730	152	442	0.40	ng/ul	0.00
4) Naphthalene-d8	10.510	136	1447	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.375	164	1017	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.127	188	2509	0.40	ng/ul	0.00
17) Chrysene-d12	21.323	240	3102	0.40	ng/ul	0.00
23) Perylene-d12	23.586	264	2916	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.254	96	189	0.42	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.111	152	1017	0.41	ng/ul	0.00
18) Fluoranthene-d10	19.158	212	3200	0.41	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.291	88	188	0.40	ng/ul	97
5) Naphthalene	10.560	128	1667	0.40	ng/ul	97
7) 2-Methylnaphthalene	12.182	142	1192	0.39	ng/ul	97
8) 1-Methylnaphthalene	12.402	142	1152	0.39	ng/ul	99
10) Acenaphthylene	14.093	152	1521	0.41	ng/ul	99
11) Acenaphthene	14.435	153	1375	0.39	ng/ul	97
12) Fluorene	15.425	166	1663	0.39	ng/ul	98
14) Pentachlorophenol	16.776	266	144	0.41	ng/ul	94
15) Phenanthrene	17.165	178	3039	0.39	ng/ul	99
16) Anthracene	17.258	178	2535	0.42	ng/ul	99
19) Fluoranthene	19.190	202	3863	0.40	ng/ul	99
20) Pyrene	19.553	202	3912	0.40	ng/ul	98
21) Benzo(a)anthracene	21.306	228	3832	0.39	ng/ul	99
22) Chrysene	21.359	228	4583	0.41	ng/ul	99
24) Benzo(b)fluoranthene	22.904	252	4733	0.37	ng/ul	99
25) Benzo(k)fluoranthene	22.951	252	5036	0.37	ng/ul	99
26) Benzo(a)pyrene	23.489	252	4139	0.42	ng/ul#	96
27) Indeno(1,2,3-cd)pyrene	25.872	276	5611	0.40	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.892	278	4518	0.40	ng/ul	98
29) Benzo(g,h,i)perylene	26.570	276	4646	0.39	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051421\
Data File : BN014644.D
Acq On : 15 May 2021 12:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4014

Quant Time: May 15 12:39:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

