

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051520\
 Data File : BN010729.D
 Acq On : 15 May 2020 15:14
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 15 15:53:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 10:34:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.345	0.349	-1.2	118	0.00
3	n-Nitrosodimethylamine	0.199	0.197	1.0	117	-0.02
4 S	2-Fluorophenol	0.783	0.841	-7.4	120	0.00
5 S	Phenol-d6	0.957	1.045	-9.2	124	0.00
6	bis(2-Chloroethyl)ether	0.986	1.039	-5.4	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.278	0.278	0.0	115	0.00
9	Naphthalene	1.006	1.006	0.0	114	-0.01
10	Hexachlorobutadiene	0.230	0.226	1.7	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.652	0.654	-0.3	114	0.00
12	2-Methylnaphthalene	0.709	0.716	-1.0	115	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.145	0.139	4.1	126	0.00
15 S	2-Fluorobiphenyl	1.468	1.448	1.4	114	0.00
16	Acenaphthylene	1.634	1.554	4.9	118	0.00
17	Acenaphthene	1.104	1.099	0.5	116	-0.01
18	Fluorene	1.447	1.397	3.5	117	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.221	0.213	3.6	119	0.00
21	Hexachlorobenzene	0.263	0.249	5.3	116	-0.01
22	Pentachlorophenol	0.081	0.078	3.7	136	-0.01
23	Phenanthrene	1.073	1.048	2.3	124	0.00
24	Anthracene	0.902	0.863	4.3	127	-0.01
25 SURR	Fluoranthene-d10	1.085	1.057	2.6	130	0.00
26	Fluoranthene	1.219	1.181	3.1	128	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
28	Pyrene	1.434	1.424	0.7	130	0.00
29 S	Terphenyl-d14	0.949	0.991	-4.4	138	0.00
30	Benzo(a)anthracene	1.167	1.214	-4.0	150	0.00
31	Chrysene	1.321	1.358	-2.8	136	0.00
32	Bis(2-ethylhexyl)phthalate	0.511	0.528	-3.3	146	0.00
33	Indeno(1,2,3-cd)pyrene	1.175	1.469	-25.0#	177#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	137	0.00
35	Benzo(b)fluoranthene	1.322	1.252	5.3	137	0.00
36	Benzo(k)fluoranthene	1.424	1.406	1.3	140	0.00
37 C	Benzo(a)pyrene	1.144	1.137	0.6	140	0.00
38	Dibenzo(a,h)anthracene	0.978	1.038	-6.1	159#	0.00
39	Benzo(g,h,i)perylene	1.147	1.209	-5.4	152#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051520\
Data File : BN010729.D
Acq On : 15 May 2020 15:14
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 15 15:53:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 10:34:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		