

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051525\
 Data File : BN037026.D
 Acq On : 15 May 2025 10:31
 Operator : RC/JU
 Sample : PB167952BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB167952BL

Quant Time: May 15 11:25:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.618	152	2007	0.40	ng	0.00
7) Naphthalene-d8	10.394	136	5198	0.40	ng	#-0.01
13) Acenaphthene-d10	14.267	164	2997	0.40	ng	0.00
19) Phenanthrene-d10	17.009	188	6152	0.40	ng	0.00
29) Chrysene-d12	21.207	240	4881	0.40	ng	# 0.00
35) Perylene-d12	23.407	264	4331	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.213	112	1977	0.38	ng	0.00
5) Phenol-d6	6.795	99	2221	0.34	ng	0.00
8) Nitrobenzene-d5	8.771	82	1890	0.33	ng	0.00
11) 2-Methylnaphthalene-d10	11.996	152	2661	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	279	0.21	ng	0.00
15) 2-Fluorobiphenyl	12.888	172	4950	0.36	ng	0.00
27) Fluoranthene-d10	19.049	212	6063	0.36	ng	0.00
31) Terphenyl-d14	19.653	244	4251	0.41	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051525\
Data File : BN037026.D
Acq On : 15 May 2025 10:31
Operator : RC/JU
Sample : PB167952BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB167952BL

Quant Time: May 15 11:25:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 14 11:26:32 2025
Response via : Initial Calibration

