

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051619\
 Data File : BN005707.D
 Acq On : 16 May 2019 17:25
 Operator : JU/SJ
 Sample : K2690-23MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A5167MSD

Quant Time: May 16 18:36:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 07:15:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	661	0.40	ng/ul	0.00
2) Naphthalene-d8	10.37	136	2830	0.40	ng/ul	0.02
6) Acenaphthene-d10	14.22	164	2023	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.00	188	5080	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	5912	0.40	ng/ul	0.02
20) Perylene-d12	23.46	264	6378	0.40	ng/ul	0.04
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.98	152	1762	0.40	ng/ul	0.02
14) Fluoranthene-d10	19.03	212	6412	0.40	ng/ul	0.00
Target Compounds				Ovalue		
8) Acenaphthene	14.28	153	2406	0.334	ng/ul	97
11) Pentachlorophenol	16.70	266	334	0.254	ng/ul#	80
17) Pyrene	19.43	202	7643	0.350	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051619\
Data File : BN005707.D
Acq On : 16 May 2019 17:25
Operator : JU/SJ
Sample : K2690-23MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5167MSD

Quant Time: May 16 18:36:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

