

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051619\
 Data File : BN005724.D
 Acq On : 17 May 2019 04:22
 Operator : JU/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.479

Quant Time: May 17 06:47:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 17 04:47:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	3202	0.40	ng/ul	0.00
2) Naphthalene-d8	10.35	136	13665	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.23	164	8334	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.98	188	19908	0.40	ng/ul	0.00
16) Chrysene-d12	21.22	240	14095	0.40	ng/ul	0.00
20) Perylene-d12	23.43	264	13955	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.96	152	9267	0.43	ng/ul	0.00
14) Fluoranthene-d10	19.03	212	23196	0.37	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.40	128	14657	0.405	ng/ul	97
5) 2-Methylnaphthalene	12.03	142	10807	0.438	ng/ul	100
7) Acenaphthylene	13.95	152	17727	0.422	ng/ul	98
8) Acenaphthene	14.29	153	12302	0.414	ng/ul	97
9) Fluorene	15.29	166	14523	0.419	ng/ul	99
11) Pentachlorophenol	16.68	266	1915	0.371	ng/ul	92
12) Phenanthrene	17.03	178	23799	0.439	ng/ul	99
13) Anthracene	17.12	178	23492	0.434	ng/ul	98
15) Fluoranthene	19.06	202	27726	0.369	ng/ul	96
17) Pyrene	19.43	202	27528	0.529	ng/ul	96
18) Benzo(a)anthracene	21.20	228	22212	0.457	ng/ul	96
19) Chrysene	21.25	228	22265	0.373	ng/ul	95
21) Benzo(b)fluoranthene	22.77	252	19722	0.470	ng/ul	92
22) Benzo(k)fluoranthene	22.81	252	21438	0.408	ng/ul#	93
23) Benzo(a)pyrene	23.33	252	20015	0.427	ng/ul#	93
24) Indeno(1,2,3-cd)pyrene	25.63	276	22724	0.433	ng/ul	93
25) Dibenzo(a,h)anthracene	25.63	278	18679	0.431	ng/ul	95
26) Benzo(g,h,i)perylene	26.31	276	18835	0.421	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051619\
Data File : BN005724.D
Acq On : 17 May 2019 04:22
Operator : JU/SJ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.479

Quant Time: May 17 06:47:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 17 04:47:07 2019
Response via : Initial Calibration

