


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051622\
 Data File : BN019784.D
 Acq On : 16 May 2022 10:19
 Operator : CG/JU
 Sample : DFTPP247
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP247

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 17 01:41:05 2022



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.8	153816	PASS
68	69	0.00	2	2.0	2644	PASS
69	198	0.00	100	41.1	134923	PASS
70	69	0.00	2	0.6	767	PASS
127	198	10	80	52.3	171627	PASS
197	198	0.00	2	0.2	787	PASS
198	198	100	100	100.0	328363	PASS
199	198	5	9	6.3	20765	PASS
275	198	10	60	24.2	79363	PASS
365	198	1	100	2.8	9149	PASS
441	198	0.01	100	10.7	35112	PASS
442	198	50	100	71.8	235691	PASS
443	442	15	24	18.8	44427	PASS

m/z	Abundance
39.90	972.0
41.10	215.0
43.10	242.0
44.00	1560.0
48.90	437.0
57.00	208.0
73.00	450.0
83.90	330.0
206.90	1116.0
208.00	336.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
37.00	592.0
38.10	1735.0
39.00	12365.0
40.00	1338.0
41.00	631.0
43.00	264.0
44.00	1513.0
44.90	344.0
47.00	156.0
49.00	1519.0
50.00	42064.0
51.00	163136.0
52.00	8116.0
53.00	315.0
55.00	910.0
56.00	4370.0
57.00	10154.0
57.90	453.0
58.90	227.0
59.80	196.0
61.00	1747.0
62.00	2227.0
63.00	5471.0
63.90	845.0
65.00	2683.0
66.10	236.0
67.00	326.0
68.00	2816.0
69.00	139904.0
70.00	716.0
70.90	253.0
73.00	1440.0
74.00	14502.0
75.00	22600.0
76.00	7793.0
77.00	157056.0
78.00	11302.0
79.00	11041.0
80.00	8578.0
81.00	11723.0
82.00	2812.0
83.00	2878.0
83.90	621.0
85.00	1929.0
86.00	3586.0
87.00	1391.0
87.90	686.0
89.00	346.0
90.90	2114.0
92.00	3119.0
93.00	18296.0
94.00	1181.0
95.00	412.0
96.00	959.0
97.00	411.0
98.00	14159.0
99.00	13758.0
99.90	1274.0
101.00	6364.0
101.90	291.0
103.00	2215.0
104.00	3645.0
105.00	4150.0
106.00	1353.0
107.00	49656.0
108.00	7057.0
109.00	1373.0
110.00	87152.0
111.00	14389.0
112.00	1462.0
112.90	515.0
114.80	219.0
116.00	2691.0
117.00	42128.0
117.90	3440.0
118.90	416.0
119.90	507.0
120.90	247.0
121.90	3216.0
123.00	4529.0
124.00	2216.0
124.90	2300.0
126.10	1025.0
127.00	169792.0
128.00	13021.0
129.00	64032.0
130.00	5587.0
130.90	1009.0
132.00	563.0
132.80	388.0
133.90	1831.0
135.00	5380.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

135.90	2415.0
136.90	3478.0
137.90	688.0
138.90	339.0
140.00	878.0
141.00	8197.0
142.00	2726.0
142.90	1956.0
143.80	496.0
145.00	605.0
146.00	1495.0
147.00	4450.0
148.00	9940.0
149.00	2365.0
150.00	595.0
151.00	1148.0
151.90	520.0
153.00	2967.0
154.00	1781.0
155.00	4722.0
156.00	6873.0
157.00	1481.0
157.90	1563.0
158.90	1007.0
160.00	2520.0
161.00	3552.0
161.90	1098.0
162.90	450.0
164.00	352.0
164.90	3066.0
166.00	2579.0
167.00	15566.0
168.00	12229.0
169.00	1617.0
170.00	407.0
171.00	825.0
171.90	1308.0
173.00	1996.0
173.90	3991.0
175.00	6042.0
175.90	1720.0
177.00	2706.0
177.90	979.0
178.90	10891.0
180.00	9166.0
180.90	3923.0
182.00	781.0
182.90	367.0
183.90	1064.0
185.00	5622.0
186.00	43504.0
187.00	13188.0
188.00	1778.0
188.90	2839.0
190.00	462.0
191.00	1251.0
192.00	4141.0
193.00	4205.0
194.00	977.0
194.70	283.0
196.00	7882.0
197.90	311296.0
199.00	18560.0
199.90	1919.0
201.50	1252.0
202.90	2246.0
204.00	10929.0
205.00	17928.0
206.00	74960.0
207.00	11574.0
208.00	2978.0
209.00	1009.0
210.10	1127.0
211.00	3219.0
212.90	330.0
214.90	701.0
216.00	1446.0
217.00	19232.0
218.00	3058.0
218.90	275.0
221.00	11217.0
221.70	2798.0
223.00	3917.0
224.00	41752.0
225.00	10356.0
226.00	1361.0
227.00	16560.0
228.00	2519.0
228.90	3205.0
229.90	506.0
231.00	1271.0
232.00	257.0
232.90	343.0
233.90	1205.0
235.00	1103.0
235.80	751.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

237.00	2150.0
237.80	327.0
239.00	699.0
240.00	719.0
241.00	982.0
241.90	2020.0
243.00	2204.0
244.00	29976.0
245.00	4653.0
246.00	6014.0
247.00	1206.0
247.70	303.0
249.00	1103.0
250.00	223.0
250.80	243.0
252.00	194.0
252.90	698.0
255.00	162752.0
256.00	23392.0
256.90	1962.0
258.00	7787.0
258.90	1444.0
260.00	163.0
261.00	299.0
263.90	265.0
265.00	3209.0
265.90	548.0
270.80	381.0
272.00	287.0
272.90	4799.0
273.90	11261.0
275.00	68320.0
276.00	10099.0
277.00	5799.0
278.00	1234.0
279.00	244.0
280.80	207.0
281.90	272.0
282.90	736.0
284.00	482.0
285.00	913.0
286.10	264.0
288.90	198.0
289.70	179.0
290.90	153.0
292.00	221.0
293.00	1439.0
294.00	523.0
296.00	18528.0
297.00	2614.0
297.90	275.0
301.10	183.0
301.80	223.0
303.00	2028.0
304.00	573.0
307.90	353.0
308.80	250.0
309.90	296.0
313.90	953.0
314.90	2243.0
316.00	1238.0
316.90	181.0
321.00	625.0
322.00	188.0
322.90	5238.0
324.00	1026.0
326.20	156.0
326.90	887.0
327.90	383.0
332.00	370.0
333.00	541.0
334.00	3983.0
334.90	1022.0
336.00	160.0
340.90	568.0
342.00	233.0
345.90	1045.0
346.90	342.0
352.00	1652.0
352.90	1244.0
354.00	1902.0
354.90	388.0
365.00	7547.0
366.00	1108.0
371.00	430.0
372.00	2783.0
373.00	715.0
382.90	708.0
383.90	235.0
389.90	188.0
390.90	215.0
392.00	202.0
402.00	1077.0
402.90	1652.0
404.00	672.0
405.00	268.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

421.00	1313.0
422.00	1498.0
422.90	8917.0
424.00	3760.0
425.00	382.0
440.10	172.0
441.00	24808.0
442.00	168064.0
443.00	33136.0
444.00	3416.0
444.90	191.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
37.00	809.0
38.00	2073.0
39.00	13480.0
40.00	1445.0
41.00	585.0
43.00	305.0
44.00	1527.0
44.90	357.0
47.90	225.0
49.00	1745.0
50.00	46120.0
51.00	179008.0
52.00	9474.0
52.90	391.0
55.00	1172.0
55.90	4863.0
57.00	11895.0
57.90	608.0
58.90	170.0
59.90	156.0
61.00	2372.0
62.00	2225.0
63.00	6873.0
64.00	1009.0
65.00	3133.0
66.00	209.0
66.80	345.0
68.00	3018.0
69.00	158400.0
70.00	869.0
71.00	172.0
73.00	1714.0
74.00	15964.0
75.00	25768.0
76.00	8482.0
77.00	183296.0
78.00	12567.0
79.00	13202.0
79.90	9897.0
81.00	13143.0
82.00	3626.0
82.90	2954.0
84.00	506.0
85.00	2608.0
86.00	3927.0
87.00	1642.0
87.90	707.0
89.00	340.0
91.00	3114.0
91.90	3498.0
93.00	21224.0
93.90	1414.0
95.00	352.0
96.00	876.0
97.00	406.0
98.00	16680.0
99.00	14772.0
99.90	1272.0
101.00	7151.0
101.90	430.0
102.90	2338.0
104.00	4872.0
104.90	4220.0
106.00	1454.0
107.00	58416.0
108.00	8029.0
109.00	1699.0
110.00	104696.0
111.00	15444.0
112.00	2062.0
113.00	727.0
113.80	162.0
114.90	271.0
116.00	2990.0
117.00	47888.0
118.00	3679.0
119.10	534.0
120.00	601.0
120.90	268.0
122.00	3666.0
122.90	5878.0
123.90	2473.0
124.90	2621.0
127.00	201216.0
128.00	14614.0
129.00	76056.0
130.00	6880.0
131.00	1273.0
132.00	749.0
132.70	397.0
133.90	1895.0
135.00	5559.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

136.00	2875.0
137.00	3353.0
137.90	901.0
139.00	306.0
140.10	856.0
140.90	10005.0
141.90	3335.0
142.90	2571.0
144.00	510.0
144.90	432.0
145.90	1632.0
147.00	5469.0
147.90	11638.0
149.00	2606.0
149.90	563.0
151.00	1500.0
151.80	946.0
152.90	3105.0
153.90	2399.0
155.00	4912.0
156.00	8609.0
156.90	1261.0
157.90	1697.0
159.00	1516.0
160.00	3328.0
160.90	4688.0
161.90	1338.0
163.10	383.0
163.90	486.0
164.90	3457.0
166.00	3083.0
167.00	20008.0
168.00	11981.0
169.00	1880.0
169.90	582.0
170.90	679.0
171.90	1852.0
172.90	1937.0
174.00	4368.0
175.00	7319.0
176.00	2187.0
176.90	4066.0
178.00	1236.0
178.90	14318.0
179.90	9650.0
180.90	4722.0
182.00	763.0
182.70	235.0
183.20	275.0
183.90	1099.0
185.00	6913.0
186.00	55960.0
187.00	16243.0
188.00	2254.0
189.00	3551.0
189.90	606.0
191.00	1614.0
192.00	4591.0
193.00	4800.0
194.00	1019.0
194.80	666.0
196.00	11053.0
196.80	2360.0
198.00	387008.0
198.90	25872.0
200.00	2208.0
201.40	2155.0
203.00	3097.0
204.00	13557.0
205.00	22992.0
206.00	99296.0
207.00	13925.0
208.00	4064.0
209.00	1532.0
210.10	1633.0
211.00	3527.0
211.90	1016.0
212.90	298.0
213.90	183.0
214.90	1253.0
215.90	2091.0
216.90	25272.0
217.90	3277.0
218.90	299.0
221.00	13973.0
221.80	3281.0
223.00	4821.0
224.00	56048.0
225.00	14059.0
226.00	1661.0
227.00	23128.0
228.00	3249.0
229.00	4797.0
229.90	782.0
231.00	1797.0
232.00	369.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

232.90	430.0
233.90	1532.0
235.00	1635.0
235.90	1074.0
237.00	2038.0
237.90	356.0
239.00	969.0
240.00	825.0
240.90	1373.0
242.00	2336.0
243.00	2424.0
244.00	38416.0
245.00	5385.0
246.00	7175.0
246.90	1924.0
247.80	362.0
248.90	1157.0
250.00	280.0
250.80	305.0
251.80	300.0
253.00	965.0
254.10	1750.0
255.00	205632.0
256.00	29536.0
257.00	2645.0
258.00	12262.0
258.90	2083.0
259.90	471.0
261.00	223.0
264.10	367.0
264.90	4333.0
265.90	682.0
269.90	220.0
271.00	480.0
272.00	445.0
273.00	6379.0
274.00	15987.0
275.00	96344.0
276.00	11535.0
277.00	8133.0
278.00	1477.0
279.00	207.0
280.90	308.0
281.80	220.0
283.00	737.0
284.00	632.0
285.00	1187.0
285.80	152.0
288.90	450.0
290.00	221.0
290.90	187.0
291.90	409.0
293.00	1780.0
293.90	494.0
295.90	24048.0
297.00	3384.0
297.90	210.0
301.10	293.0
302.00	590.0
303.00	2824.0
304.00	765.0
305.00	159.0
308.00	388.0
308.90	205.0
310.00	334.0
313.00	229.0
314.00	1342.0
315.00	2471.0
316.00	1513.0
317.00	236.0
321.00	801.0
321.90	437.0
323.00	8387.0
324.10	1462.0
324.80	152.0
325.80	189.0
327.00	1460.0
328.00	745.0
331.90	658.0
332.90	742.0
334.00	4919.0
334.90	1185.0
335.80	196.0
341.00	1089.0
341.90	283.0
345.90	1713.0
346.90	354.0
352.00	2483.0
353.00	1910.0
354.00	2581.0
355.00	517.0
364.90	11129.0
365.90	1456.0
369.90	222.0
371.00	660.0
371.90	3997.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

373.00	1031.0
383.00	1011.0
383.90	370.0
384.80	189.0
389.90	539.0
390.90	399.0
391.90	179.0
400.80	194.0
401.90	1805.0
402.90	2217.0
403.90	687.0
404.90	192.0
421.00	2090.0
422.00	1964.0
422.90	13568.0
424.10	3538.0
424.90	622.0
441.00	43704.0
442.00	275904.0
443.00	54560.0
444.00	4696.0
444.90	265.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

m/z	Abundance
35.80	182.0
37.00	616.0
38.00	1608.0
39.00	8670.0
39.90	1292.0
41.00	297.0
43.10	360.0
44.00	1405.0
44.90	234.0
49.00	1216.0
50.00	30592.0
51.00	119304.0
52.00	6039.0
53.00	327.0
55.00	821.0
56.00	3545.0
57.00	7427.0
58.00	232.0
60.90	1494.0
62.00	1824.0
63.00	4783.0
64.00	523.0
65.00	2153.0
65.90	221.0
66.90	287.0
68.00	2099.0
69.00	106464.0
70.00	715.0
73.00	1182.0
74.00	11216.0
75.00	18608.0
76.00	6133.0
77.00	126648.0
78.00	8720.0
79.00	8229.0
79.90	6149.0
81.00	9201.0
82.00	2143.0
83.00	2142.0
83.90	455.0
84.90	1540.0
86.00	3012.0
87.00	1127.0
88.00	411.0
88.80	303.0
91.00	2048.0
92.00	2290.0
93.00	14852.0
93.90	1007.0
94.90	384.0
95.90	706.0
97.00	255.0
98.00	11582.0
99.00	9399.0
99.90	747.0
101.00	5275.0
102.00	314.0
102.90	1712.0
104.00	3113.0
105.00	2839.0
106.00	1083.0
107.00	38680.0
107.90	5734.0
109.00	1010.0
110.00	72928.0
111.00	10427.0
111.90	1316.0
112.90	338.0
114.00	172.0
114.80	155.0
116.10	2168.0
117.00	34264.0
117.90	2768.0
118.90	344.0
120.00	564.0
120.80	163.0
121.10	156.0
121.90	2700.0
122.90	4185.0
123.90	1859.0
125.00	1782.0
126.10	952.0
127.00	143872.0
128.00	10592.0
129.00	55816.0
130.00	4406.0
130.90	863.0
132.00	566.0
132.90	293.0
133.90	1733.0
135.00	4533.0
136.00	1936.0

Instrument :

BNA_N

ClientSampleId :

DFTPP247

137.00	2709.0
139.00	263.0
140.00	649.0
140.90	7244.0
142.00	2348.0
142.90	1599.0
144.00	590.0
145.00	478.0
145.90	1393.0
147.00	4177.0
148.00	8550.0
149.00	1668.0
149.90	612.0
151.00	1053.0
151.90	479.0
153.00	1980.0
153.90	1883.0
155.00	3743.0
156.00	6215.0
157.10	1319.0
157.90	1514.0
158.90	1051.0
160.00	2525.0
161.00	3077.0
162.00	1305.0
162.90	325.0
164.00	272.0
165.00	2407.0
165.90	2443.0
167.00	14653.0
168.00	8847.0
169.00	1084.0
170.10	453.0
171.00	504.0
172.00	1249.0
173.00	1533.0
174.00	3619.0
175.00	5431.0
175.90	1647.0
177.00	2905.0
177.90	773.0
179.00	11000.0
180.00	7531.0
180.90	4105.0
182.00	470.0
183.00	302.0
183.90	761.0
184.90	4976.0
186.00	38224.0
187.00	12141.0
187.90	1272.0
188.90	2457.0
190.00	357.0
191.00	1369.0
191.90	3424.0
192.90	3573.0
193.90	826.0
195.10	294.0
196.00	8345.0
197.90	286784.0
199.00	17864.0
200.00	1652.0
201.50	1350.0
203.00	2467.0
204.00	10746.0
205.00	17624.0
206.00	76976.0
207.00	10606.0
208.00	2473.0
209.00	1055.0
210.00	1118.0
211.00	2974.0
214.90	1106.0
216.00	1694.0
216.90	19096.0
217.90	2367.0
219.10	272.0
221.00	11148.0
221.90	2621.0
223.00	4340.0
224.00	43008.0
225.00	10380.0
226.00	1191.0
226.90	17608.0
228.00	2648.0
228.90	3747.0
229.90	473.0
230.90	1392.0
231.90	218.0
233.00	251.0
234.00	1099.0
235.00	1245.0
236.00	914.0
237.00	1707.0
237.90	340.0
238.90	820.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

239.90	797.0
240.90	1003.0
241.90	2004.0
243.00	2786.0
244.00	33616.0
245.00	4256.0
245.90	5926.0
247.00	1226.0
248.00	282.0
248.90	1259.0
249.90	179.0
250.90	299.0
251.90	311.0
253.00	881.0
254.10	1551.0
255.00	167296.0
256.00	24448.0
257.00	1975.0
258.00	9663.0
258.90	1569.0
259.90	203.0
260.90	377.0
263.80	310.0
265.00	3914.0
265.90	506.0
269.90	186.0
271.00	207.0
271.90	371.0
273.00	5049.0
274.00	12577.0
275.00	73424.0
276.00	9757.0
277.00	5857.0
277.90	1360.0
278.90	325.0
281.00	206.0
281.90	317.0
283.00	666.0
283.90	401.0
285.00	1303.0
285.90	168.0
288.90	301.0
289.90	183.0
290.90	172.0
292.00	297.0
292.90	1423.0
293.80	343.0
295.10	649.0
296.00	19000.0
297.00	2968.0
298.00	159.0
300.90	252.0
301.80	404.0
303.00	2232.0
304.00	898.0
304.90	213.0
307.80	227.0
308.90	165.0
310.00	264.0
312.90	163.0
314.00	1005.0
314.90	1910.0
316.00	1353.0
316.90	169.0
320.90	735.0
322.10	376.0
323.00	6361.0
324.00	1490.0
325.00	254.0
326.90	1356.0
328.00	516.0
332.00	558.0
332.90	537.0
334.00	4815.0
335.00	1025.0
335.90	152.0
339.00	157.0
340.10	156.0
340.90	682.0
341.80	178.0
346.00	1752.0
346.80	219.0
351.90	2129.0
353.00	1533.0
354.00	2264.0
354.90	601.0
358.90	171.0
364.90	8772.0
365.90	1303.0
369.90	190.0
370.90	636.0
372.00	3183.0
373.00	781.0
382.90	1155.0
384.00	371.0
384.80	156.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247

390.00	583.0
390.90	283.0
391.80	267.0
400.80	206.0
401.90	1446.0
403.00	2701.0
403.90	695.0
405.00	162.0
414.90	174.0
421.00	1903.0
421.90	1715.0
423.00	13071.0
423.90	3646.0
424.90	447.0
439.80	250.0
440.90	36824.0
442.00	263104.0
443.00	45584.0
443.90	4593.0
445.00	233.0

Instrument :
BNA_N
ClientSampleId :
DFTPP247