

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051623\
 Data File : BN025435.D
 Acq On : 16 May 2023 09:51
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: May 17 02:21:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 17 02:20:47 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.092	152	5476	0.400	ng	0.00
7) Naphthalene-d8	10.910	136	16057	0.400	ng	0.00
13) Acenaphthene-d10	14.715	164	10556	0.400	ng	0.00
19) Phenanthrene-d10	17.460	188	23984	0.400	ng	0.00
29) Chrysene-d12	21.652	240	21794	0.400	ng	0.00
35) Perylene-d12	24.188	264	21948	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.622	112	1163	0.099	ng	0.00
5) Phenol-d6	7.232	99	1625	0.103	ng	0.00
8) Nitrobenzene-d5	9.255	82	1058	0.096	ng	0.00
11) 2-Methylnaphthalene-d10	12.491	152	2154	0.096	ng	0.00
14) 2,4,6-Tribromophenol	16.219	330	279	0.078	ng	0.01
15) 2-Fluorobiphenyl	13.358	172	3712	0.096	ng	0.00
27) Fluoranthene-d10	19.485	212	5537	0.097	ng	0.00
31) Terphenyl-d14	20.085	244	3864	0.098	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.462	88	507	0.099	ng	96
3) n-Nitrosodimethylamine	3.780	42	23	0.004	ng	# 47
6) bis(2-Chloroethyl)ether	7.507	93	3173	0.169	ng	# 69
9) Naphthalene	10.963	128	3901	0.097	ng	# 94
10) Hexachlorobutadiene	11.252	225	794	0.101	ng	# 99
12) 2-Methylnaphthalene	12.563	142	2853	0.097	ng	97
16) Acenaphthylene	14.437	152	4617	0.095	ng	99
17) Acenaphthene	14.780	154	2855	0.094	ng	98
18) Fluorene	15.759	166	3788	0.095	ng	96
20) 4,6-Dinitro-2-methylph...	15.859	198	7	0.002	ng	# 1
21) 4-Bromophenyl-phenylether	16.653	248	1355	0.097	ng	94
22) Hexachlorobenzene	16.765	284	1248	0.097	ng	96
23) Atrazine	16.926	200	1133	0.097	ng	93
24) Pentachlorophenol	17.125	266	20	0.005	ng	# 19
25) Phenanthrene	17.497	178	6434	0.095	ng	99
26) Anthracene	17.596	178	6000	0.092	ng	99
28) Fluoranthene	19.519	202	7554	0.095	ng	100
30) Pyrene	19.878	202	7777	0.093	ng	99
32) Benzo(a)anthracene	21.634	228	7374	0.096	ng	98
33) Chrysene	21.697	228	7169	0.094	ng	99
34) Bis(2-ethylhexyl)phtha...	21.562	149	4606	0.103	ng	99
36) Indeno(1,2,3-cd)pyrene	26.857	276	8025	0.091	ng	100
37) Benzo(b)fluoranthene	23.410	252	7200	0.094	ng	# 85
38) Benzo(k)fluoranthene	23.460	252	7373	0.093	ng	# 81
39) Benzo(a)pyrene	24.079	252	6699	0.091	ng	# 80
40) Dibenzo(a,h)anthracene	26.877	278	6322	0.089	ng	# 87
41) Benzo(g,h,i)perylene	27.676	276	6904	0.092	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051623\
Data File : BN025435.D
Acq On : 16 May 2023 09:51
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: May 17 02:21:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 17 02:20:47 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.092 min Scan# 61
 Delta R.T. 0.001 min
 Lab File: BN025435.D
 Acq: 16 May 2023 09:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 5476
 Ion Ratio Lower Upper
 152 100
 150 149.4 123.7 185.5
 115 61.6 50.0 75.0



#2
 1,4-Dioxane
 Concen: 0.099 ng
 RT: 3.462 min Scan# 39
 Delta R.T. -0.007 min
 Lab File: BN025435.D
 Acq: 16 May 2023 09:51

Tgt Ion: 88 Resp: 507
 Ion Ratio Lower Upper
 88 100
 43 33.7 25.1 37.7
 58 72.8 55.5 83.3









#3
n-Nitrosodimethylamine
Concen: 0.004 ng
RT: 3.780 min Scan# 81
Delta R.T. -0.029 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 23
Ion Ratio Lower Upper
42 100
74 65.2 100.1 150.1#
44 452.2 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.099 ng
RT: 5.622 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion: 112 Resp: 1163
Ion Ratio Lower Upper
112 100
64 59.0 43.9 65.9
63 33.5 25.0 37.4









#5
Phenol-d6
Concen: 0.103 ng
RT: 7.232 min Scan# 561
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 1625
Ion Ratio Lower Upper
99 100
42 18.6 16.4 24.6
71 37.5 29.4 44.0



#6
bis(2-Chloroethyl)ether
Concen: 0.169 ng
RT: 7.507 min Scan# 599
Delta R.T. 0.008 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion: 93 Resp: 3173
Ion Ratio Lower Upper
93 100
63 33.8 48.5 72.7#
95 16.8 22.0 33.0#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.910 min Scan# 910
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:136 Resp: 16057
Ion Ratio Lower Upper
136 100
137 11.5 9.4 14.0
54 8.5 7.0 10.4
68 7.4 5.8 8.6



#8
Nitrobenzene-d5
Concen: 0.096 ng
RT: 9.255 min Scan# 799
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion: 82 Resp: 1058
Ion Ratio Lower Upper
82 100
128 49.9 35.7 53.5
54 58.2 38.6 57.8#









#9
Naphthalene
Concen: 0.097 ng
RT: 10.963 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 3901
Ion Ratio Lower Upper
128 100
129 14.8 9.4 14.2#
127 15.1 10.6 16.0



#10
Hexachlorobutadiene
Concen: 0.101 ng
RT: 11.252 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:225 Resp: 794
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.1 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 0.096 ng
RT: 12.491 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2154
Ion Ratio Lower Upper
152 100
151 20.5 16.5 24.7



#12
2-Methylnaphthalene
Concen: 0.097 ng
RT: 12.563 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:142 Resp: 2853
Ion Ratio Lower Upper
142 100
141 90.2 71.4 107.0
115 45.9 32.6 48.8





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.715 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:164 Resp: 10556

Ion Ratio Lower Upper

164 100

162 106.5 84.8 127.2

160 50.4 40.6 61.0



#14

2,4,6-Tribromophenol

Concen: 0.078 ng

RT: 16.219 min Scan# 1700

Delta R.T. 0.012 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:330 Resp: 279


Ion Ratio Lower Upper

330 100

332 99.6 75.5 113.3

141 62.4 35.3 52.9#









#15
2-Fluorobiphenyl
Concen: 0.096 ng
RT: 13.358 min Scan# 1439
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:172 Resp: 3712
Ion Ratio Lower Upper
172 100
171 35.1 27.0 40.6
170 24.3 18.2 27.2



#16
Acenaphthylene
Concen: 0.095 ng
RT: 14.437 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:152 Resp: 4617
Ion Ratio Lower Upper
152 100
151 19.8 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.094 ng
RT: 14.780 min Scan# 1572
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 2855
Ion Ratio Lower Upper
154 100
153 112.2 90.6 135.8
152 55.1 45.7 68.5



#18
Fluorene
Concen: 0.095 ng
RT: 15.759 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:166 Resp: 3788
Ion Ratio Lower Upper
166 100
165 99.7 76.7 115.1
167 14.6 10.5 15.7





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.460 min Scan# 1800

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 23984

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.4 8.7 13.1



#20

4,6-Dinitro-2-methylphenol

Concen: 0.002 ng

RT: 15.859 min Scan# 1671

Delta R.T. 0.013 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:198 Resp: 7


Ion Ratio Lower Upper

198 100

51 414.5 68.5 102.7#

105 370.9 57.6 86.4#









#21
4-Bromophenyl-phenylether
Concen: 0.097 ng
RT: 16.653 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 1355
Ion Ratio Lower Upper
248 100
250 96.4 79.0 118.4
141 66.3 46.5 69.7



#22
Hexachlorobenzene
Concen: 0.097 ng
RT: 16.765 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:284 Resp: 1248
Ion Ratio Lower Upper
284 100
142 41.9 35.2 52.8
249 36.5 27.1 40.7





#23

Atrazine

Concen: 0.097 ng

RT: 16.926 min Scan# 1757

Delta R.T. 0.012 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:200 Resp: 1133

Ion Ratio Lower Upper

200 100

173 32.4 22.4 33.6

215 50.4 36.7 55.1



#24

Pentachlorophenol

Concen: 0.005 ng

RT: 17.125 min Scan# 1773

Delta R.T. -0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:266 Resp: 20


Ion Ratio Lower Upper

266 100

264 0.0 49.5 74.3#

268 0.0 51.0 76.4#









#25
Phenanthrene
Concen: 0.095 ng
RT: 17.497 min Scan# 1803
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 6434
Ion Ratio Lower Upper
178 100
176 19.4 15.6 23.4
179 15.7 12.2 18.4



#26
Anthracene
Concen: 0.092 ng
RT: 17.596 min Scan# 1811
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:178 Resp: 6000
Ion Ratio Lower Upper
178 100
176 19.2 14.9 22.3
179 15.6 12.1 18.1









#27
Fluoranthene-d10
Concen: 0.097 ng
RT: 19.485 min Scan# 2195
Delta R.T. 0.000 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 5537
Ion Ratio Lower Upper
212 100
106 14.0 11.1 16.7
104 8.7 6.6 10.0



#28
Fluoranthene
Concen: 0.095 ng
RT: 19.519 min Scan# 2203
Delta R.T. 0.004 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:202 Resp: 7554
Ion Ratio Lower Upper
202 100
101 11.0 8.6 13.0
203 17.3 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.652 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:240 Resp: 21794

Ion Ratio Lower Upper

240 100

120 16.2 11.6 17.4

236 29.0 22.9 34.3



#30

Pyrene

Concen: 0.093 ng

RT: 19.878 min Scan# 2288

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:202 Resp: 7777


Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.4

203 18.2 14.2 21.4





#31

Terphenyl-d14

Concen: 0.098 ng

RT: 20.085 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:244 Resp: 3864

Ion Ratio Lower Upper

244 100

212 10.4 6.9 10.3#

122 13.2 7.4 11.2#



#32

Benzo(a)anthracene

Concen: 0.096 ng

RT: 21.634 min Scan# 2500

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:228 Resp: 7374


Ion Ratio Lower Upper

228 100

226 28.7 22.1 33.1

229 20.1 15.4 23.2





#33

Chrysene

Concen: 0.094 ng

RT: 21.697 min Scan# 2169

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 7169

Ion Ratio Lower Upper

228 100

226 29.6 23.9 35.9

229 20.6 15.8 23.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.103 ng

RT: 21.562 min Scan# 2492

Delta R.T. 0.000 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:149 Resp: 4606


Ion Ratio Lower Upper

149 100

167 28.1 22.2 33.4

279 3.5 2.5 3.7





#35
Perylene-d12

Concen: 0.400 ng

RT: 24.188 min Scan# 31

Delta R.T. -0.009 min

Lab File: BN025435.D


Acq: 16 May 2023 09:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1






Tgt Ion:264 Resp: 21948

Ion Ratio Lower Upper

264 100

260 26.7 21.0 31.6

265 32.0 25.5 38.3



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.091 ng

RT: 26.857 min Scan# 4109

Delta R.T. -0.015 min

Lab File: BN025435.D

Acq: 16 May 2023 09:51




Tgt Ion:276 Resp: 8025


Ion Ratio Lower Upper

276 100

138 26.7 21.0 31.6

277 24.7 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.094 ng
RT: 23.410 min Scan# 2930
Delta R.T. -0.006 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 7200
Ion Ratio Lower Upper
252 100
253 32.0 19.0 28.4#
125 16.2 9.4 14.2#



#38
Benzo(k)fluoranthene
Concen: 0.093 ng
RT: 23.460 min Scan# 2947
Delta R.T. -0.009 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51

Tgt Ion:252 Resp: 7373
Ion Ratio Lower Upper
252 100
253 32.8 19.1 28.7#
125 19.7 9.5 14.3#









#39
Benzo(a)pyrene
Concen: 0.091 ng
RT: 24.079 min Scan# 3157
Delta R.T. 0.006 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 6699
Ion Ratio Lower Upper
252 100
253 35.6 19.5 29.3#
125 18.6 10.2 15.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.089 ng
RT: 26.877 min Scan# 4116
Delta R.T. -0.015 min
Lab File: BN025435.D
Acq: 16 May 2023 09:51


Tgt Ion:278 Resp: 6322
Ion Ratio Lower Upper
278 100
139 23.5 15.0 22.4#
279 32.8 20.1 30.1#





#41
 Benzo(g,h,i)perylene
 Concen: 0.092 ng
 RT: 27.676 min Scan# 41
 Delta R.T. -0.017 min
 Lab File: BN025435.D
 Acq: 16 May 2023 09:51

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1



Tgt Ion:	276	Resp:	6904
Ion	Ratio	Lower	Upper
276	100		
277	27.5	19.6	29.4
138	27.1	19.0	28.6

