

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051624\
 Data File : BN031529.D
 Acq On : 15 May 2024 09:28
 Operator : MA/JU
 Sample : SSTD0.186
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.1216

Quant Time: May 15 12:16:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN051624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 12:15:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	11783	0.400	ng/ul	0.00
4) Naphthalene-d8	10.481	136	37659	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.336	164	20313	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.079	188	44643	0.400	ng/ul	0.00
17) Chrysene-d12	21.264	240	39998	0.400	ng/ul	0.00
23) Perylene-d12	23.500	264	35418	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.076	152	5613	0.084	ng/ul	0.00
18) Fluoranthene-d10	19.106	212	10454	0.072	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.530	128	10388	0.085	ng/ul	98
7) 2-Methylnaphthalene	12.147	142	6597	0.083	ng/ul	100
8) 1-Methylnaphthalene	12.367	142	7020	0.083	ng/ul	100
10) Acenaphthylene	14.054	152	9217	0.083	ng/ul	99
11) Acenaphthene	14.396	153	7092	0.083	ng/ul	99
12) Fluorene	15.386	166	7716	0.078	ng/ul	100
15) Phenanthrene	17.121	178	13319	0.083	ng/ul	99
16) Anthracene	17.210	178	10708	0.078	ng/ul	98
19) Fluoranthene	19.138	202	13588	0.070	ng/ul	99
20) Pyrene	19.501	202	13993	0.073	ng/ul	100
21) Benzo(a)anthracene	21.249	228	15590	0.094	ng/ul	99
22) Chrysene	21.299	228	15708	0.086	ng/ul	99
24) Benzo(b)fluoranthene	22.827	252	12537	0.073	ng/ul	86
25) Benzo(k)fluoranthene	22.874	252	13293	0.073	ng/ul#	85
26) Benzo(a)pyrene	23.400	252	12239	0.087	ng/ul#	80
27) Indeno(1,2,3-cd)pyrene	25.763	276	12669	0.079	ng/ul	95
28) Dibenzo(a,h)anthracene	25.784	278	9866	0.077	ng/ul#	90
29) Benzo(g,h,i)perylene	26.451	276	9538	0.071	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051624\
Data File : BN031529.D
Acq On : 15 May 2024 09:28
Operator : MA/JU
Sample : SST0.186
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1216

Quant Time: May 15 12:16:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN051624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 12:15:51 2024
Response via : Initial Calibration

