

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051721\
 Data File : BN014657.D
 Acq On : 17 May 2021 17:57
 Operator : CG/JU
 Sample : PB136460BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS460

Quant Time: May 18 01:22:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	454	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	1530	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1036	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	2555	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	2916	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	2636	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	309	0.66	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	844	0.32	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	2800	0.38	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.28	88	805	1.654	ng/ul#	85
5) Naphthalene	10.56	128	1421	0.320	ng/ul	98
7) 2-Methylnaphthalene	12.18	142	1000	0.313	ng/ul	100
8) 1-Methylnaphthalene	12.40	142	969	0.314	ng/ul	100
10) Acenaphthylene	14.09	152	1155	0.308	ng/ul	99
11) Acenaphthene	14.44	153	1133	0.318	ng/ul	96
12) Fluorene	15.43	166	1354	0.309	ng/ul	99
14) Pentachlorophenol	16.78	266	239	0.665	ng/ul	90
15) Phenanthrene	17.17	178	2554	0.321	ng/ul	99
16) Anthracene	17.26	178	2072	0.334	ng/ul	99
19) Fluoranthene	19.19	202	3312	0.366	ng/ul	98
20) Pyrene	19.55	202	3372	0.368	ng/ul	97
21) Benzo(a)anthracene	21.31	228	3171	0.344	ng/ul	99
22) Chrysene	21.36	228	4028	0.382	ng/ul	98
24) Benzo(b)fluoranthene	22.91	252	4063	0.351	ng/ul	96
25) Benzo(k)fluoranthene	22.95	252	4368	0.358	ng/ul	99
26) Benzo(a)pyrene	23.49	252	3366	0.380	ng/ul#	96
27) Indeno(1,2,3-cd)pyrene	25.87	276	4853	0.385	ng/ul#	97
28) Dibenzo(a,h)anthracene	25.90	278	3937	0.390	ng/ul	99
29) Benzo(g,h,i)perylene	26.57	276	4035	0.375	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051721\
Data File : BN014657.D
Acq On : 17 May 2021 17:57
Operator : CG/JU
Sample : PB136460BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS460

Quant Time: May 18 01:22:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

