

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051821\
 Data File : BN014667.D
 Acq On : 18 May 2021 17:12
 Operator : CG/JU
 Sample : PB136511BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS511

Quant Time: May 19 02:50:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	497	0.40	ng/ul	0.00
4) Naphthalene-d8	10.51	136	1622	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1141	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	2827	0.40	ng/ul	0.00
17) Chrysene-d12	21.33	240	3305	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	3001	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	299	0.59	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	939	0.33	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	3165	0.38	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.29	88	886	1.663	ng/ul#	85
5) Naphthalene	10.56	128	1527	0.324	ng/ul	96
7) 2-Methylnaphthalene	12.18	142	1067	0.315	ng/ul	97
8) 1-Methylnaphthalene	12.40	142	1034	0.316	ng/ul	97
10) Acenaphthylene	14.09	152	1290	0.313	ng/ul	100
11) Acenaphthene	14.44	153	1239	0.315	ng/ul	98
12) Fluorene	15.43	166	1497	0.310	ng/ul	96
14) Pentachlorophenol	16.78	266	288	0.724	ng/ul	87
15) Phenanthrene	17.17	178	2807	0.319	ng/ul	98
16) Anthracene	17.26	178	2229	0.325	ng/ul	99
19) Fluoranthene	19.19	202	3726	0.363	ng/ul	98
20) Pyrene	19.56	202	3734	0.359	ng/ul	99
21) Benzo(a)anthracene	21.31	228	3628	0.347	ng/ul	98
22) Chrysene	21.36	228	4534	0.380	ng/ul	99
24) Benzo(b)fluoranthene	22.91	252	4576	0.347	ng/ul	97
25) Benzo(k)fluoranthene	22.96	252	4945	0.356	ng/ul	99
26) Benzo(a)pyrene	23.50	252	3737	0.371	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	25.88	276	5413	0.377	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.90	278	4362	0.379	ng/ul	98
29) Benzo(a,h,i)perylene	26.58	276	4392	0.359	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051821\
Data File : BN014667.D
Acq On : 18 May 2021 17:12
Operator : CG/JU
Sample : PB136511BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS511

Quant Time: May 19 02:50:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

