

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051821\
 Data File : BN014670.D
 Acq On : 18 May 2021 19:36
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4020

Quant Time: May 19 02:51:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	562	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	1796	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.38	164	1242	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.13	188	3119	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	3761	0.40	ng/ul	0.00
23) Perylene-d12	23.59	264	3499	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	220	0.38	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	1262	0.41	ng/ul	0.00
18) Fluoranthene-d10	19.16	212	3876	0.41	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.29	88	215	0.357	ng/ul	96
5) Naphthalene	10.56	128	2049	0.393	ng/ul	96
7) 2-Methylnaphthalene	12.18	142	1478	0.394	ng/ul	97
8) 1-Methylnaphthalene	12.40	142	1452	0.401	ng/ul	99
10) Acenaphthylene	14.09	152	1871	0.417	ng/ul	99
11) Acenaphthene	14.44	153	1683	0.393	ng/ul	96
12) Fluorene	15.43	166	2021	0.385	ng/ul	99
14) Pentachlorophenol	16.78	266	199	0.453	ng/ul#	90
15) Phenanthrene	17.17	178	3675	0.379	ng/ul	100
16) Anthracene	17.26	178	3042	0.402	ng/ul	99
19) Fluoranthene	19.19	202	4654	0.398	ng/ul	99
20) Pyrene	19.55	202	4645	0.393	ng/ul	97
21) Benzo(a)anthracene	21.31	228	4440	0.374	ng/ul	99
22) Chrysene	21.36	228	5434	0.400	ng/ul	100
24) Benzo(b)fluoranthene	22.90	252	5802	0.377	ng/ul	98
25) Benzo(k)fluoranthene	22.95	252	5934	0.366	ng/ul	99
26) Benzo(a)pyrene	23.49	252	4811	0.409	ng/ul#	92
27) Indeno(1,2,3-cd)pyrene	25.88	276	6897	0.412	ng/ul#	67
28) Dibenzo(a,h)anthracene	25.89	278	5505	0.410	ng/ul	96
29) Benzo(a,h,i)perylene	26.57	276	5527	0.387	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN051821\
Data File : BN014670.D
Acq On : 18 May 2021 19:36
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleID :
SSTD0.4020

Quant Time: May 19 02:51:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

