


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051821\
 Data File : BN014661.D
 Acq On : 18 May 2021 09:34
 Operator : CG/JU
 Sample : DFTPP38
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP038

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 14 15:50:11 2021



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	17.3	101112	PASS
68	69	0.00	2	1.3	1646	PASS
69	198	0.00	100	21.7	127178	PASS
70	69	0.00	2	0.6	702	PASS
127	198	10	80	25.4	148594	PASS
197	198	0.00	2	0.9	5120	PASS
198	198	100	100	100.0	584896	PASS
199	198	5	9	7.4	43437	PASS
275	198	10	60	27.9	163477	PASS
365	198	1	100	3.0	17421	PASS
441	198	0.01	100	10.6	62272	PASS
442	198	50	100	70.9	414634	PASS
443	442	15	24	17.2	71333	PASS

m/z	Abundance
37.20	360.0
39.50	370.0
43.90	476.0
47.10	329.0
48.80	637.0
73.00	883.0
83.70	771.0
86.10	827.0
133.00	308.0
147.00	429.0
190.80	451.0
192.20	471.0
206.90	4405.0
208.20	957.0
209.20	799.0
280.90	672.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

m/z	Abundance
38.10	1232.0
39.00	7193.0
39.80	545.0
40.20	380.0
40.80	318.0
43.90	1035.0
47.60	306.0
49.00	1465.0
50.00	28928.0
51.00	112152.0
51.90	5421.0
53.10	545.0
54.80	732.0
55.90	3954.0
56.90	7575.0
57.90	635.0
59.80	400.0
61.00	1719.0
62.00	2811.0
63.10	6457.0
64.20	762.0
65.00	1766.0
68.00	3061.0
69.00	141824.0
70.00	1387.0
73.00	2350.0
74.10	14945.0
75.00	25520.0
76.00	7827.0
77.00	175488.0
78.10	12831.0
79.00	16303.0
80.00	10752.0
81.00	12967.0
81.90	3280.0
82.90	3959.0
84.00	2299.0
85.00	2263.0
86.00	4169.0
87.00	2566.0
87.90	1049.0
88.80	764.0
91.10	3199.0
92.10	4452.0
93.00	23688.0
93.90	2105.0
94.80	873.0
95.80	1039.0
97.10	897.0
98.00	24352.0
99.00	18400.0
99.90	1191.0
100.90	8057.0
102.20	1148.0
102.90	3246.0
104.00	6128.0
105.00	5969.0
106.30	2791.0
107.00	64240.0
108.00	7480.0
109.00	1952.0
109.90	134272.0
111.00	18960.0
111.90	2386.0
112.90	1915.0
114.30	361.0
115.30	677.0
116.00	2983.0
117.00	60584.0
118.10	3432.0
118.70	342.0
119.30	337.0
119.90	807.0
121.20	1076.0
122.00	2595.0
123.00	4084.0
124.10	2490.0
124.80	2125.0
125.10	2012.0
127.00	158528.0
128.00	12416.0
129.00	68240.0
129.90	5434.0
130.80	503.0
131.90	1489.0
134.00	2795.0
135.00	4820.0
135.90	2591.0
136.90	3202.0
138.10	631.0
139.10	377.0
139.70	667.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

141.00	10167.0
141.90	3358.0
142.90	1459.0
144.90	439.0
145.90	2460.0
146.90	6195.0
147.90	10979.0
149.00	3110.0
149.80	968.0
151.00	1628.0
151.70	627.0
152.20	872.0
152.90	2987.0
153.90	2195.0
155.00	4951.0
156.10	7748.0
156.90	1776.0
158.00	3255.0
159.10	1819.0
160.00	3331.0
160.90	3666.0
161.90	1301.0
162.80	464.0
165.00	2945.0
166.00	3401.0
167.00	30936.0
168.00	12985.0
169.10	1899.0
169.80	1019.0
170.80	908.0
172.10	1716.0
173.00	2403.0
174.00	5394.0
175.00	6911.0
176.00	3744.0
176.90	4179.0
177.90	1627.0
179.00	12388.0
179.90	12584.0
180.90	5647.0
182.00	615.0
183.00	587.0
184.00	1444.0
185.00	10153.0
186.00	63064.0
187.00	18304.0
188.00	3071.0
189.00	4250.0
189.80	689.0
191.10	3045.0
191.90	5580.0
193.10	4976.0
193.80	832.0
194.30	1352.0
196.00	8159.0
197.10	6083.0
198.00	546112.0
198.90	40320.0
199.90	2796.0
200.90	1370.0
201.60	1850.0
202.90	3229.0
204.10	12531.0
205.00	26576.0
206.00	119616.0
207.00	14734.0
207.90	4506.0
208.90	2308.0
210.10	2238.0
211.10	5744.0
213.10	1081.0
214.90	2462.0
216.00	2870.0
216.90	23120.0
218.00	3830.0
219.20	1503.0
220.00	785.0
221.00	14017.0
221.90	4689.0
222.90	5496.0
224.00	66936.0
225.00	19328.0
225.90	2641.0
227.00	31000.0
228.00	5359.0
229.00	5230.0
230.10	591.0
231.10	2275.0
232.00	444.0
233.20	737.0
233.90	2249.0
234.90	2294.0
236.00	1883.0
237.10	3754.0
239.00	750.0
240.00	1438.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

241.00	1598.0
242.00	2527.0
243.10	3743.0
244.00	56584.0
245.10	6754.0
246.00	12506.0
247.00	2566.0
247.80	543.0
249.00	969.0
250.70	308.0
252.00	1181.0
252.80	1670.0
255.00	290176.0
256.00	43512.0
257.00	4800.0
258.00	16528.0
258.90	3545.0
260.00	445.0
261.00	389.0
263.90	466.0
264.90	5421.0
266.00	2008.0
266.90	418.0
267.70	379.0
270.90	802.0
272.10	579.0
273.00	9265.0
274.00	25176.0
275.00	136128.0
276.00	20336.0
277.00	12294.0
278.10	1669.0
278.80	658.0
281.20	514.0
282.00	1455.0
283.00	1582.0
284.00	1066.0
284.80	1497.0
285.30	1399.0
286.00	335.0
288.80	440.0
290.20	332.0
291.00	612.0
291.90	815.0
293.00	3218.0
294.10	1634.0
294.80	996.0
295.90	37528.0
297.00	5569.0
298.20	591.0
300.90	799.0
301.40	302.0
302.10	1446.0
303.00	4039.0
304.10	1389.0
307.90	1483.0
309.70	447.0
313.00	1386.0
313.90	1454.0
314.90	4082.0
316.00	1738.0
321.00	1702.0
322.00	1007.0
323.00	13199.0
324.00	3807.0
324.90	315.0
326.00	622.0
327.10	1980.0
328.10	1585.0
332.00	1190.0
333.00	1257.0
334.00	8422.0
335.20	2148.0
338.90	532.0
341.20	2553.0
342.90	544.0
345.90	1839.0
347.00	1093.0
351.90	3747.0
352.90	2319.0
354.00	5710.0
355.00	840.0
365.00	12207.0
366.00	3703.0
370.90	1145.0
372.00	5840.0
373.10	1821.0
382.80	1254.0
383.80	1002.0
384.20	818.0
390.10	1055.0
391.20	488.0
392.00	553.0
393.10	392.0
401.00	777.0
401.90	2825.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

403.00	4590.0
404.10	2762.0
405.10	1326.0
415.30	562.0
421.00	2776.0
422.00	2775.0
423.10	20856.0
424.00	9955.0
425.00	992.0
438.90	339.0
439.90	850.0
441.00	49136.0
442.00	297728.0
443.00	61824.0
444.10	5534.0
444.90	599.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

m/z	Abundance
37.20	451.0
38.00	1874.0
39.00	5968.0
40.10	763.0
44.00	393.0
48.60	695.0
49.20	1184.0
50.00	31736.0
51.00	122848.0
52.00	5025.0
55.00	1573.0
55.90	2682.0
57.00	9177.0
58.10	469.0
61.00	2195.0
62.00	3521.0
63.00	6894.0
63.90	781.0
65.00	2772.0
69.00	148608.0
70.20	689.0
73.20	1911.0
74.00	18072.0
75.00	27648.0
76.00	10888.0
77.00	192832.0
78.00	13492.0
79.00	18208.0
80.00	8453.0
80.90	11133.0
81.90	4248.0
83.00	2696.0
83.70	781.0
84.90	2411.0
86.00	6568.0
86.90	1687.0
88.00	544.0
91.00	4063.0
92.00	3909.0
93.00	28880.0
93.90	2253.0
96.20	2463.0
96.90	454.0
98.00	24384.0
98.90	16359.0
100.00	1969.0
100.90	8521.0
102.90	3464.0
103.90	2782.0
104.90	4726.0
107.00	66048.0
108.00	9267.0
109.10	3428.0
110.00	144960.0
111.00	16920.0
112.00	1946.0
113.10	1235.0
115.90	2763.0
117.00	63584.0
118.00	5739.0
118.80	875.0
119.30	350.0
119.90	689.0
121.10	310.0
121.90	5648.0
122.90	7830.0
124.10	2353.0
124.90	3864.0
125.90	718.0
127.00	174272.0
127.90	15535.0
129.00	72144.0
129.90	8104.0
131.00	1061.0
131.80	480.0
134.00	3280.0
134.90	8825.0
136.00	2701.0
137.00	3151.0
137.80	866.0
139.00	964.0
140.10	1194.0
141.00	8554.0
142.00	3457.0
143.00	2491.0
144.00	535.0
145.10	1422.0
146.00	1659.0
147.00	5615.0
148.00	14469.0
149.00	4480.0
150.80	903.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

151.40	829.0
152.20	1416.0
153.00	3031.0
154.00	2212.0
155.00	6244.0
156.00	8116.0
156.80	1443.0
157.10	1627.0
157.80	1971.0
158.90	1843.0
160.00	4386.0
161.00	6200.0
161.80	1800.0
163.00	641.0
163.90	595.0
164.20	662.0
165.00	3307.0
166.10	3936.0
167.00	24776.0
168.00	18800.0
169.00	2690.0
169.90	1125.0
171.00	884.0
171.90	2187.0
172.90	2686.0
174.00	4314.0
175.00	9358.0
175.90	1951.0
177.00	5801.0
177.70	1306.0
179.00	17008.0
180.00	13437.0
181.00	5159.0
182.10	554.0
182.80	653.0
183.90	2191.0
185.00	12966.0
186.00	81584.0
187.00	23904.0
188.00	4523.0
189.00	4554.0
191.00	1968.0
192.10	4824.0
193.00	6967.0
194.10	2700.0
195.20	1380.0
196.00	13222.0
197.10	5168.0
198.00	695808.0
199.00	53240.0
200.00	3168.0
201.40	3058.0
203.10	5041.0
204.00	20504.0
205.00	33376.0
206.00	140288.0
207.00	24696.0
208.00	5087.0
209.00	1954.0
210.00	1976.0
211.10	5461.0
211.90	1388.0
212.80	659.0
214.30	578.0
215.00	4434.0
216.00	3106.0
217.00	31696.0
218.10	4520.0
219.10	582.0
220.30	1071.0
221.10	21600.0
222.00	5703.0
223.00	9953.0
224.00	86728.0
225.10	20232.0
226.00	3199.0
227.00	37536.0
228.00	5778.0
229.00	6369.0
229.90	1214.0
231.10	3085.0
233.20	1389.0
233.90	2053.0
235.00	2057.0
236.10	2086.0
237.00	6129.0
237.70	319.0
239.00	1546.0
240.00	1324.0
241.00	2951.0
242.00	4038.0
243.10	4932.0
244.00	68136.0
245.00	11113.0
246.00	16824.0
246.80	3091.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

247.90	324.0
248.90	3682.0
250.70	660.0
251.30	762.0
252.20	711.0
253.10	1838.0
255.00	386624.0
256.00	49664.0
257.20	4666.0
258.00	18264.0
259.00	2635.0
260.00	958.0
260.90	1401.0
263.90	736.0
265.00	8202.0
266.20	730.0
268.10	349.0
269.90	566.0
270.90	816.0
272.00	1315.0
273.00	11944.0
274.00	43448.0
275.00	189632.0
276.00	28392.0
277.00	18216.0
278.00	3184.0
281.00	1113.0
283.10	2339.0
284.00	709.0
285.10	2615.0
289.00	641.0
289.90	647.0
291.20	813.0
292.10	1152.0
293.00	4877.0
294.20	916.0
295.00	1474.0
296.00	43200.0
297.10	6011.0
297.90	773.0
301.10	1481.0
301.80	957.0
303.00	5790.0
304.10	2480.0
308.00	686.0
309.00	507.0
310.20	688.0
312.80	593.0
313.90	3585.0
314.90	3892.0
316.00	2647.0
317.00	828.0
321.00	1702.0
321.90	802.0
323.00	19384.0
324.00	4480.0
325.80	302.0
327.00	2033.0
328.00	1815.0
332.90	1552.0
334.00	13464.0
335.00	4616.0
336.00	1190.0
340.30	358.0
340.90	1268.0
342.10	567.0
343.00	405.0
345.90	4078.0
346.80	1324.0
352.10	5558.0
353.10	3749.0
354.00	4022.0
355.20	1614.0
365.00	20528.0
365.90	2633.0
367.00	307.0
370.00	630.0
371.00	1508.0
372.00	8503.0
373.00	2963.0
383.00	1800.0
384.20	1485.0
384.80	319.0
390.00	1026.0
391.10	1291.0
391.70	375.0
392.20	1009.0
395.90	308.0
401.10	602.0
401.90	3553.0
403.00	6488.0
403.90	2583.0
404.80	546.0
414.90	760.0
421.10	3111.0
421.90	4277.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

423.00	33776.0
423.90	10243.0
425.00	1962.0
439.00	844.0
441.00	64880.0
442.00	514624.0
443.00	73112.0
444.10	8793.0
445.10	650.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

m/z	Abundance
38.00	826.0
38.90	3788.0
40.20	561.0
43.90	617.0
48.80	468.0
50.10	18064.0
51.00	68336.0
52.00	2899.0
55.10	909.0
55.80	2285.0
57.00	6003.0
58.90	541.0
60.90	594.0
62.00	1036.0
63.10	3652.0
64.00	643.0
64.80	1767.0
68.00	1878.0
69.00	91104.0
69.90	719.0
73.10	1181.0
74.00	13665.0
74.90	16648.0
76.10	5286.0
77.00	121616.0
78.10	7992.0
78.90	10317.0
79.90	6944.0
80.90	9429.0
82.10	2284.0
82.90	1977.0
84.00	833.0
85.00	2221.0
86.10	2055.0
87.10	1098.0
88.10	504.0
90.90	3319.0
92.00	1767.0
93.00	15007.0
94.10	829.0
96.00	1007.0
97.10	1572.0
97.90	16258.0
99.00	12118.0
100.00	615.0
101.00	5208.0
103.00	1440.0
104.00	3609.0
105.00	5924.0
106.20	1917.0
107.00	41224.0
107.90	7649.0
108.90	1806.0
110.00	98408.0
111.00	12597.0
111.90	2036.0
116.20	4847.0
117.00	47176.0
118.00	2030.0
118.90	953.0
120.10	450.0
121.90	4250.0
123.00	6453.0
123.90	1932.0
124.80	2009.0
125.40	759.0
126.00	1103.0
127.00	112984.0
127.90	9738.0
129.00	48208.0
129.90	5078.0
131.00	731.0
131.80	409.0
133.10	517.0
134.00	1929.0
135.00	4822.0
136.00	1442.0
136.90	2576.0
140.20	957.0
140.90	8518.0
142.00	2791.0
142.80	1115.0
144.00	793.0
145.00	625.0
145.90	3025.0
147.10	3675.0
148.00	11048.0
149.20	1640.0
150.20	942.0
150.80	1069.0
153.00	3788.0
154.00	2015.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

154.90	4217.0
155.90	7191.0
157.20	1890.0
157.90	2425.0
158.70	1117.0
160.00	2686.0
161.00	5064.0
161.70	1206.0
162.10	945.0
163.00	446.0
164.00	992.0
164.80	1446.0
165.20	1136.0
166.10	2646.0
167.00	21232.0
168.00	10027.0
169.10	1620.0
170.00	1580.0
171.00	798.0
171.60	911.0
172.10	2238.0
172.80	2051.0
174.00	3189.0
175.00	8058.0
176.00	1947.0
176.90	2996.0
178.00	1345.0
179.00	12153.0
180.00	9609.0
181.00	3501.0
182.70	518.0
183.90	1219.0
184.90	7661.0
186.00	57816.0
187.00	20008.0
188.10	1603.0
188.90	2994.0
190.20	650.0
191.10	1774.0
192.00	3547.0
193.00	3970.0
194.00	866.0
195.00	1377.0
195.90	9553.0
197.00	4110.0
197.90	512768.0
199.00	36752.0
199.90	2568.0
200.90	378.0
201.30	1518.0
203.10	2535.0
204.00	12196.0
205.00	33264.0
206.00	120744.0
207.10	16640.0
208.00	2793.0
209.00	1927.0
209.90	1290.0
210.90	3864.0
212.00	679.0
214.20	345.0
215.10	1970.0
216.00	1925.0
216.90	26176.0
218.00	3078.0
219.00	484.0
220.00	735.0
220.90	11916.0
221.80	4514.0
223.10	5544.0
224.00	59648.0
225.10	13342.0
226.10	2852.0
227.00	30240.0
228.10	3517.0
228.90	6012.0
229.90	411.0
231.00	1654.0
233.20	815.0
233.80	1220.0
234.80	1935.0
235.80	1150.0
236.00	1139.0
236.90	2087.0
238.80	680.0
239.20	682.0
239.90	961.0
240.90	1101.0
242.00	2819.0
242.90	2997.0
244.00	49808.0
245.00	9208.0
246.00	10757.0
247.10	2789.0
247.80	627.0
249.00	2022.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

250.00	690.0
250.90	782.0
251.70	982.0
252.80	1782.0
255.00	287104.0
256.00	39896.0
257.00	4142.0
257.90	19560.0
259.00	3750.0
260.90	529.0
264.90	6073.0
265.80	1381.0
267.20	643.0
268.10	740.0
270.00	347.0
271.10	1356.0
272.00	849.0
273.00	9724.0
274.00	28520.0
275.00	164672.0
276.10	17032.0
277.00	13269.0
278.00	3215.0
280.90	520.0
283.00	1239.0
284.10	332.0
285.10	3220.0
288.90	462.0
289.90	493.0
291.30	383.0
291.90	982.0
292.90	2213.0
293.80	773.0
294.60	1015.0
296.00	44624.0
297.00	6619.0
298.00	514.0
300.80	504.0
302.20	554.0
303.10	5888.0
304.10	1637.0
305.10	567.0
308.00	984.0
309.80	980.0
311.00	337.0
312.80	441.0
314.00	2214.0
315.00	4891.0
316.00	3508.0
316.80	461.0
317.20	426.0
319.80	418.0
320.90	1155.0
322.10	986.0
323.10	13265.0
324.00	2992.0
326.90	2664.0
328.00	1716.0
329.20	308.0
332.00	1420.0
333.00	1156.0
333.90	7684.0
335.00	3973.0
335.90	447.0
339.90	467.0
340.90	1739.0
342.00	722.0
346.00	3222.0
346.90	1931.0
351.90	2612.0
352.90	2192.0
354.10	5970.0
355.00	1234.0
361.00	594.0
365.00	19528.0
365.80	2451.0
366.10	2147.0
370.10	525.0
371.10	1173.0
371.90	9954.0
373.10	2540.0
382.90	2288.0
384.10	1100.0
389.90	1291.0
391.10	498.0
392.10	627.0
400.80	676.0
402.00	3350.0
403.00	3785.0
404.00	1947.0
420.90	4139.0
422.00	5084.0
423.00	31912.0
424.00	7597.0
425.10	982.0
437.90	489.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038

439.00	922.0
439.70	458.0
441.00	72800.0
442.00	431552.0
442.90	79064.0
443.90	6234.0

Instrument :
BNA_N
ClientSampleId :
DFTPP038