

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN051824\  
 Data File : BN031557.D  
 Acq On : 17 May 2024 09:07  
 Operator : MA/JU  
 Sample : PB160909BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

**Instrument :**  
 BNA\_N  
**ClientSampleId :**  
 SBLK909

Quant Time: May 17 11:42:27 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-SIM-BN051624.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed May 15 12:57:23 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	15156	0.400	ng/ul	0.00
4) Naphthalene-d8	10.481	136	47841	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.332	164	24586	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.075	188	54120	0.400	ng/ul	0.00
17) Chrysene-d12	21.258	240	36730	0.400	ng/ul	0.00
23) Perylene-d12	23.491	264	35152	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	10622	0.582	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.070	152	19563	0.289	ng/ul	0.00
18) Fluoranthene-d10	19.101	212	35345	0.347	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN051824\  
 Data File : BN031557.D  
 Acq On : 17 May 2024 09:07  
 Operator : MA/JU  
 Sample : PB160909BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK909

Quant Time: May 17 11:42:27 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-SIM-BN051624.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed May 15 12:57:23 2024  
 Response via : Initial Calibration

